

ATTACHMENT 3

DEC 20 2001

K012323

510(K) SUMMARY OF SAFETY & EFFECTIVENESS

Official Contact David J. Vanella
Manager, Regulatory Affairs/Product Assurance
Respironics, Inc.
1001 Murry Ridge Lane
Murrysville, PA 15668

Classification Reference 21 CFR 868.5905

Product Code MNS – Non-Continuous ventilator

Common/Usual Name Ventilator, continuous, non-life supporting

Proprietary Name Respironics BiPAP Synchrony Ventilatory Support System

Predicate Device(s) Respironics BiPAP Synchrony HC (K992530)

Reason for submission Modified design

Substantial Equivalence

The modified device has the following similarities to the previously cleared predicate device:

- Same intended use.
- Same operating principle.
- Same technology.
- Same manufacturing process.

Design verification tests were performed on the Respironics BiPAP Synchrony Ventilatory Support System as a result of the risk analysis and product requirements. All tests were verified to meet the required acceptance criteria. Respironics has determined that the modifications have no impact on the safety and effectiveness of the device. In summary, the device described in this submission is substantially equivalent to the predicate device.

The modified device complies with the applicable standards referenced in the Guidance for FDA Reviewers and Industry "Guidance for the Content of Pre-market Submissions for Software Contained in Medical Devices", May 1998.

Intended Use

The Respironics BiPAP Synchrony Ventilatory Support System is intended to provide non-invasive ventilation in adult patients (>30 kg) for the treatment of respiratory insufficiency (a condition in which the patient can continue without ventilation for some period, such as overnight) or obstructive sleep apnea. The Synchrony may be used in the hospital or home.

The Synchrony is intended for use with nasal masks and full-face masks as recommended by Respironics.

Device Description

The Respironics BiPAP Synchrony Ventilatory Support System is a microprocessor controlled blower based Bi-level positive pressure system that delivers two different positive pressure levels (IPAP/EPAP). The dual pressure levels provide a more natural means of delivering pressure support therapy to the patient resulting in improved patient comfort. Respironics is modifying the existing device with the addition of 3 alarms, expanding the EPAP setting to ²⁵ ~~20~~ cmH₂O and adding a remote control accessory for use in a clinical setting. The BiPAP Synchrony Ventilatory Support System is intended for use with a patient circuit that is used to connect the device to the patient interface device (mask). A typical patient circuit consists of a six-foot disposable or reusable smooth lumen 22mm tubing, a method of venting exhaled gases, and a patient interface device.



DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

DEC 20 2001

Ms. Zita Yurko
Respironics, Inc.
1001 Murry Ridge Lane
Murrysville, PA 15668-8550

Re: K012323
Synchrony Ventilatory Support System
Regulation Number: 868.5895
Regulation Name: Continuous Ventilator
Regulatory Class: Class II (two)
Product Code: MNS
Dated: November 23, 2001
Received: December 3, 2001

Dear Ms. Yurko:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.


Page 2 - Ms. Zita Yurko

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801 and additionally 21 CFR Part 809.10 for in vitro diagnostic devices), please contact the Office of Compliance at (301) 594-4646. Additionally, for questions on the promotion and advertising of your device, please contact the Office of Compliance at (301) 594-4639. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). Other general information on your responsibilities under the Act may be obtained from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address <http://www.fda.gov/cdrh/dsma/dsmamain.html>

Sincerely yours,



Bram D. Zuckerman, M.D.
Acting Director
Division of Cardiovascular
and Respiratory Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

510(k) Number (if known): K012323

Device Name: Respironics BiPAP Synchrony Ventilatory Support System

Intended Use/Indications for Use

The Synchrony is intended to provide non-invasive ventilation in adult patients (>30 kg) for the treatment of respiratory insufficiency (a condition in which the patient can continue without ventilation for some period, such as overnight) or obstructive sleep apnea. The Synchrony may be used in the hospital or home.

The Synchrony is intended for use with nasal masks and full-face masks as recommended by Respironics.

Environment of Use/Patient Population

For use in the home or hospital/institutional environment on adult patients.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription Use X
(Per 21 CFR 801.109)

OR

Over-The-Counter Use
(Optional Format 1-2-96)


Division of Cardiovascular & Respiratory Devices
510(k) Number K012323