

K955262

SECTION 2 - SUMMARY AND CERTIFICATION

510(k) SUMMARY

Monaghan Medical Corporation
5 Latour Ave., Suite 1600
P.O. Box 2805
Plattsburgh, NY 12901-0299
Phone: 518-561-7330 Fax: 518-561-5660
Contact: James A. Cochie
QA/RA Manager

Date Prepared: October 31, 1995

Device Proprietary Name: TruZone Peak Flow Meter
Device Classification name: Peak Flow Meter for
Spirometry, 73 BZH

Legally Marketed Device Identity: TruZone Peak Flow
Meter, 510(k) #926307

Device Description:

The TruZone Peak Flow Meter is a simple mechanical device that responds to and indicates the Peak Expiratory Flow Rate (PEFR) generated by the user during a forced exhalation maneuver.

The inlet of the device is placed into the user's mouth after which the maximum rate at exhalation is attempted from nearly full lungs. As this forced exhalation commences an internal piston extends a spring as reaction to an internal pressure spike. An indicator is pushed by the moving piston and remains at the maximum displacement position of the piston after the maneuver is terminated. This position corresponds to the PEFR and is quantified in liters per minute by reading a scale next to the indicator. The piston displacement function is logarithmic thereby providing greater resolution of scale at lower PEFR readings.

The TruZone Peak Flow Meter is identical in form and function to the TruZone Peak Flow Meter as described in its 510(k) #926307 except for the addition of a separate package of ColorZone tapes as an accessory to the TruZone Peak Flow Meter. The essence of the modification to this 510(k) is adding the aforementioned accessory package for sale to physicians.