



**Critical  
Device  
Corporation**

499 Nibus Street • Brea • CA • 92821

FEB 10 1997

K964812

## 510(k) Summary

---

1. **Submitter:** **Critical Device Corporation**  
499 Nibus Street  
Brea, CA 92821  
Tel: 714-257-2197 or 800-321-6462  
Fax: 800-299-4849
2. **Contact:** Dan Hyun, President  
Critical Device Corporation
3. **Date prepared:** November 22, 1996
4. **Device trade name:** NIMA™ Needleless Injectionsite Master Adapter and IV Sets  
**Common name:** Intravenous (IV) Medication Connector and IV Sets
5. **Predicate device:** NIMA™ Needleless Injectionsite Master Adapter and IV Sets
6. **Description:** The NIMA™ adapter is a two-way valve that permits easy needleless intermittent and continuous access in IV therapy. It can be connected to peripheral or central venous catheters or Y-sites. The normally closed NIMA™ adapter valve is opened by inserting a standard male luer taper, such as on an extension set, IV tubing or syringe to the female end of the NIMA™ adapter.  

This device does not require the use of hypodermic needles for the infusion of IV solutions. Thus preventing needle stick injuries. The NIMA™ adapter is packaged individually and as attached as part of extension sets
7. **Intended Use:**
  - a. For use as a needleless alternative to IV set injection ports. The NIMA™ adapter will replace the conventional Y-site on a primary IV line as a continuous or intermittent connection.
  - b. For use as part of a program to reduce needle stick injuries and the associated transmission of blood borne pathogens such as HIV and HBV.
  - c. For use as a replacement to the injection cap (heparin cap) on an I. V. catheter for intermittent injections.

- d. For use for injection, as a gravity flow connector, and as an access port for withdrawal of fluids.
  - e. For use with standard luer taper connections.
  - f. For single patient use.
8. Technological comparison to predicate device:  
The technological characteristics are equivalent to the currently approved NIMA™ Needleless Injectionsite Master Adapter and IV Sets. Alternate materials and use duration extended from 24 to 96 hours are approved in this application.
9. Nonclinical test summary:  
New materials meet all performance specifications established for the originally approved device. Tripartite biocompatibility testing indicates that the new materials are safe and biocompatible.
10. Conclusion:  
The NIMA™ modification(s) have demonstrated safety and effectiveness, and is substantially equivalent to the legally marketed predicate device.