



# SRS INTERNATIONAL CORPORATION

Suite 1000 • 1625 K Street, NW • Washington, DC 20006-1604  
Telephone (202) 223-0157/0298 • Telecopy (202) 835-8970

OCT 31 1997

## PREMARKET NOTIFICATION [510(K)] SUMMARY

### 1. Name/Address/Phone/Fax, Contact Person, Date:

Submitter's Name: SRS International Corporation  
Address: 1625 K St., NW, Suite 1000 Washington, DC 20006  
Telephone: 202-223-1057  
Fax: 202-835-8970  
Contact Person: Michael G. Farrow, Ph.D.  
Date of Summary: February 8, 1997

### 2. Name of Device, Proprietary Name, Common Name, Classification

Name of Device: Transcom Colon Hydrotherapy  
Proprietary Name: Transcom Colon Hydrotherapy  
Common Name: Colonic Irrigation System  
Classification: Irrigator, Colonic (Gastro/Urology) 78 KPL

### 3. Legally Marketed Predicate Devices of Substantial Equivalence

Specialty Health Models A&B  
Clearwater PPC-101  
Doltolo Research Corp. Toxygen

### 4. Description of Device

This device is *not* a kit. This device is an instrument for hydrotherapy of the colon. It introduces filtered water at a comfortable temperature into the large intestine. It is an automatic system for filling and emptying water into and out of the colon and thus cleansing the colon of its contents when medically indicated such as before radiological or endoscopic examination. It is hygienic, painless and odorless. The instrument has an automatic pressure safety system; automatic disinfection; a system for collection of samples for analysis; remote and frontal panel control; a counter for the number of liters of water used; a three-way faucet; flow regulator; thermostatic mixer and an automatic temperature safety feature. It contains disposable components for use only once in the therapy

### 5. Intended Use of Device

The indication for use of this device must be restricted to colon cleansing when medically indicated, such as before radiological or endoscopic examination.

## 6. Summary of Technological Characteristics Compared to Predicates

Summaries between Transcom and three predicates manufactured by Specialty Health, Clearwater, and Dotolo Research, appear below.

Differences in technological characteristics between Transcom and the predicate devices include the following.

### Between Transcom and Specialty Health

1. Specialty Health uses a gravity flow system, Transcom a syringe.

### Between Transcom and Clearwater

1. A sink and faucet are needed for the Clearwater unit; Transcom is plumbed.
2. Clearwater has no speculum collection; Transcom does.
3. Clearwater's unit is portable; Transcom is wall mounted
4. Clearwater uses a pump; Transcom uses a syringe

### Between Transcom and Doltolo

1. Dotolo uses a gravity flow system, Transcom a syringe

Details in the technological characteristics for Transcom and the predicates are itemized on the following pages.

Parameter	TRANSKOM	SPECIALTY HEALTH Hydro-San	CLEARWATER PPC-101	DOTOLO RESEARCH TOXYGEN
Water Source	Household/Commercial	Household/Commercial	Household/Commercial	Household/Commercial
Water Flow (Pressure)	Yes	Yes	Yes	Yes
Water Flow Control Valve	Yes	Yes	Yes	Yes
Drainage System (Gravity)	Yes	Yes	Yes	Yes
Closed System Water Drainage	Yes Yes	Yes Yes	Yes Yes	Yes Yes
Mixing Valve Grohemlx Powers Hydroguard	Yes N/A	Yes	Yes Yes	Yes
Fittings Brass Stainless	Yes Yes	Yes Yes	Yes Yes	Yes Yes
Cabinet Composition Stainless Steel Sheet Metal (Painted)	Yes N/A	N/A Yes	Yes N/A	N/A Yes
Sink & Faucet	N/A	N/A	Yes	N/A
Cabinet Design	Wall-Mount	Wall-Mount	Portable (on-wheels)	Wall-Mount
Weight	132.3 lbs.	Approx 40 lbs.	70 lbs.	Approx 40 lbs.
Physical Dimensions	1020 x 720 x 270 CM	20" w x 7" d x 20" h	19" w x 16" d x 41" h	20" w x 7" d x 20" h
Pumps	None	None	Sanitation System	None

Parameter	Mode of Operation	Major Separate System Components	Intended Usage	Monitoring Systems	Water Temperature	Pressure	View Tube Assembly (Fluid Discharge Path)	Fluid Pathway	Specimen Collection
TRANSCOM	Continuous Water Flow Gravity Flow Drainage Manual Operation	Mixing Valve Pressure Regulator Water Manifold System Relief Valve System Monitoring System (Temp. Pressure) Viewing Assembly Sanitation System Lighting Assembly Drainage Assembly None	Colonic Irrigator	Yes	Yes	Yes	Yes	Disinfected, and disposable	Yes
SPECIALTY HEALTH Hydro-San	Continuous Water Flow Gravity Flow Drainage Manual Operation	Mixing Valve Pressure Regulator Water Manifold System Relief Valve System Monitoring System (Temp. Pressure) Viewing Assembly Sanitation System Lighting Assembly Drainage Assembly Sink & Faucet	Colonic Irrigator	Yes	Yes	Yes	Yes	Disinfected, and disposable	Yes
CLEARWATER PPC-101	Continuous Water Flow Gravity Flow Drainage Manual Operation	Mixing Valve Pressure Regulator Water Manifold System Relief Valve System Monitoring System (Temp. Pressure) Viewing Assembly Sanitation System Lighting Assembly Drainage Assembly None	Colonic Irrigator	Yes	Yes	Yes	Yes	Disinfected, and disposable	No
DOTOLO RESEARCH TOXYGEN	Continuous Water Flow Gravity Flow Drainage Manual Operation	Mixing Valve Pressure Regulator Water Manifold System Relief Valve System Monitoring System (Temp. Pressure) Viewing Assembly Sanitation System Lighting Assembly Drainage Assembly None	Colonic Irrigator	Yes	Yes	Yes	Yes	Disinfected, and disposable	Yes

Parameter	TRANSCOM	SPECIALTY HEALTH Hydro-San	CLEARWATER PPC-101	DOTOLO RESEARCH TOXYGEN
Regulators (Water Pressure)	Norren Yes	Norren Yes	Norren Yes	Norren Yes
Relief Valve (Water Pressure)	Norren Yes	Norren Yes	Norren Yes	Norren Yes
Over Pressure Relief System	Yes	Yes	Yes	Yes
Systems Check Valves	Yes	Yes	Yes	Yes
Sanitation System	Yes	Yes	Yes	Yes
Pump	Syringe	Gravity Flow System	Yes	Gravity Flow System
Gauges Water Temperature Water Pressure	Yes Yes	Yes Yes	Yes Yes	Yes Yes
Electrical Requirements	110/120 VAC 50/60 Hz service, power outlet to be grounded and polarized and GFI	110/120 VAC 50/60 Hz service, power outlet to be grounded and polarized and GFI	110/120 VAC 50/60 Hz service, power outlet to be grounded and polarized and GFI	110/120 VAC 50/60 Hz service, power outlet to be grounded and polarized and GFI
View Tube Assembly Back Lighting	Yes	Yes	Yes	Yes

**7. Non-Clinical Performance Data**

Substantial equivalence is not based on an assessment of non-clinical performance data.

**8. Clinical Performance Data**

Substantial equivalence is not based on an assessment of clinical performance data.

**9. Conclusions from Non-Clinical and Clinical**

Not applicable



Food and Drug Administration  
9200 Corporate Boulevard  
Rockville MD 20850

OCT 31 1997

TRANSCQM Transcendencias Comerciales, S.L.  
c/o Michael G. Farrow, Ph.D.  
SRS International Corporation  
1625 K Street, N.W.  
Washington, D.C. 20006-1604

Re: K970482  
TRANSCOM Colon Hydrotherapy Model HC-2000  
Dated: September 10, 1997  
Received: September 10, 1997  
Regulatory class: II  
21 CFR §876.5220/Product code: 78 KPL

Dear Dr. Farrow:

We have reviewed your Section 510(k) notification of intent to market the device referenced above and we have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (Premarket Approval), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 895. A substantially equivalent determination assumes compliance with the Current Good Manufacturing Practice requirement, as set forth in the Quality System Regulation (QS) for Medical Devices: General regulation (21 CFR Part 820) and that, through periodic QS inspections, the Food and Drug Administration (FDA) will verify such assumptions. Failure to comply with the GMP regulation may result in regulatory action. In addition, FDA may publish further announcements concerning your device in the Federal Register. Please note: this response to your premarket notification submission does not affect any obligation you might have under sections 531 through 542 of the Act for devices under the Electronic Product Radiation Control provisions, or other Federal laws or regulations.

This letter will allow you to begin marketing your device as described in your 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801 and additionally 809.10 for in vitro diagnostic devices), please contact the Office of Compliance at (301) 594-4613. Additionally, for questions on the promotion and advertising of your device, please contact the Office of Compliance at (301) 594-4639. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR 807.97). Other general information on your responsibilities under the Act may be obtained from the Division of Small Manufacturers Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address <http://www.fda.gov/cdrh/dsmamain.html>.

Sincerely yours,

Lillian Yin, Ph.D.  
Director, Division of Reproductive,  
Abdominal, Ear, Nose and Throat,  
and Radiological Devices  
Office of Device Evaluation  
Center for Devices and  
Radiological Health

Enclosure

510(k) Number (if known): K 970482

Device Name: TRANSCOM COLON HYDROTHERAPY MODEL HC-2000

Indications For Use:

The indication for use of this device must be restricted to colon cleansing when medically indicated, such as before radiological or endoscopic examination.

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Robert D. Rathling  
(Division Sign-Off)  
Division of Reproductive, Abdominal, ENT,  
and Radiological Devices  
510(k) Number K970482

Prescription Use   
(Per 21 CFR 801.109)

OR

Over-The-Counter Use

(Optional Format 1-2-96)