



Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20856

OCT 8 1997

Dideco, S.P.A.
c/o Mr. Barry Sall
Senior Regulatory Consultant
Pařexel International Corporation
1601 Trapelo Road
Waltham, Massachusetts 02154

Re: K971877
Lilliput 2 Infant Hollow Fiber Oxygenator
Regulatory Class: III (Three)
Product Code: 74 DTZ
Dated: August 25, 1997
Received: August 26, 1997

Dear Mr. Sall:

We have reviewed your Section 510(k) notification of intent to market the device referenced above and we have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.


If your device is classified (see above) into either class II (Special Controls) or class III (Premarket Approval), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 895. A substantially equivalent determination assumes compliance with the Current Good Manufacturing Practice requirements, as set forth in the Quality System Regulation (QS) for Medical Devices: General regulation (21 CFR Part 820) and that, through periodic QS inspections, the Food and Drug Administration (FDA) will verify such assumptions. Failure to comply with the GMP regulation may result in regulatory action. In addition, FDA may publish further announcements concerning your device in the Federal Register. Please note: this response to your premarket notification submission does not affect any obligation you might have under sections 531 through 542 of the Act for devices under the Electronic Product Radiation Control provisions, or other Federal laws or regulations.

Page 2 - Mr. Barry Sall

This letter will allow you to begin marketing your device as described in your 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801 and additionally 809.10 for in vitro diagnostic devices), please contact the Office of Compliance at (301) 594-4648. Additionally, for questions on the promotion and advertising of your device, please contact the Office of Compliance at (301) 594-4639. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR 807.97). Other general information on your responsibilities under the Act may be obtained from the Division of Small Manufacturers Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its internet address "<http://www.fda.gov/cdrh/dsmamain.html>."

Sincerely yours,



Thomas J. Callahan, Ph.D.
Director
Division of Cardiovascular, Respiratory,
and Neurological Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

870.4350 - DTZ H - CPB Oxygenator

510(k) Number (if known): K 971877

Device Name: Dideco D902 Lilliput 2 Membrane Oxygenator

Indications For Use:

The Dideco D902 Lilliput 2 Membrane Oxygenator is intended for use in infants less than 20 Kg (44 Lb.) who undergo undergo cardiopulmonary bypass surgery requiring extracorporeal circulation. It provides oxygenation and carbon dioxide removal from venous blood. The integrated heat exchanger provides blood temperature control and allows the use of hypothermia or aids in the maintenance of normothermia during surgery. The venous reservoir is intended to collect blood during normal operation, to always assure the proper oxygenation capability of the device. The D 902 LILLIPUT 2 should not be used longer than 6 hours. Contact with blood for longer periods is not advised.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Beta B. Bempsee
(Division Sign-Off)
Division of Cardiovascular, Respiratory,
and Neurological Devices

510(k) Number K 971877

Prescription Use
(Per 21 CFR 801.109)

OR

Over-The-Counter Use

(Optional format 1-2-96)

005A