

3.0 Summary of Safety and Effectiveness Information

SPONSOR: Synthes (USA)
1690 Russell Road
Paoli, PA 19301
(610) 647-9700
Contact: Thomas M. Maguire

DEVICE NAME: Synthes Proximal Humeral Nail

CLASSIFICATION: Class II, Section 888.3020 – Intramedullary fixation rod.

PREDICATE DEVICE: Synthes (USA) Unreamed Humeral Nail; Acumed Polarus Proximal Humeral Fixation Rod

DEVICE DESCRIPTION: The Synthes Proximal Humeral Nail is an intramedullary rod that features a distal taper design. It is 150 mm in length and has holes in both the proximal and distal sections that accept locking screws.

INTENDED USE: The Synthes Proximal Humeral Nail is intended for use in fractures of the proximal humerus.

MATERIAL: Ti-6Al-7Nb



OCT - 6 2000

Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

Mr. Thomas M. Maguire
Project Leader, Regulatory Affairs
Synthes (USA)
P.O. Box 1766
1690 Russell Road
Paoli, Pennsylvania 19301

Re: K002729
Trade Name: Synthes Proximal Humeral Nail
Regulatory Class: II
Product Code: JDS
Dated: August 31, 2000
Received: September 1, 2000

Dear Mr. Maguire:


We have reviewed your Section 510(k) notification of intent to market the device referenced above and we have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act). You may, therefore, market the device, subject to the general control provisions of the Act. The general control provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (Premarket Approval), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 895. A substantially equivalent determination assumes compliance with the current Good Manufacturing Practice requirement, as set forth in the Quality System Regulation (QS) for Medical Devices: General regulation (21 CFR Part 820) and that, through periodic (QS) inspections, the Food and Drug Administration (FDA) will verify such assumptions. Failure to comply with the GMP regulation may result in regulatory action. In addition, FDA may publish further announcements concerning your device in the Federal Register. Please note: this response to your premarket notification submission does not affect any obligation you might have under sections 531 through 542 of the Act for devices under the Electronic Product Radiation Control provisions, or other Federal laws or regulations.

This letter will allow you to begin marketing your device as described in your 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801 and additionally 809.10 for in vitro diagnostic devices), please contact the Office of Compliance at (301) 594-4659. Additionally, for questions on the promotion and advertising of your device, please contact the Office of Compliance at (301) 594-4639. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR 807.97). Other general information on your responsibilities under the Act may be obtained from the Division of Small Manufacturers Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address "<http://www.fda.gov/cdrh/dsmamain.html>".

Sincerely yours,

for 

Celia M. Witten, Ph.D., M.D.
Director

Division of General, Restorative and
Neurological Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure



2.0 Indications for Use Statement

Page 1 of 1

510(k) Number (if known): _____

Device Name: Synthes (USA) Proximal Humeral Nail

Indications/Contraindications: The Synthes Proximal Humeral Nail is indicated for use in fractures of the proximal humerus.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Mark N. Milburn

(Division Sign-Off)
Division of General Restorative Devices

510(k) Number K002729

Prescription Use _____
(Per 21 CFR 801.109)

OR

Over-The-Counter Use_