

3/4/99

K981812

510k Summary of Safety and Effectiveness

This summary of 510(k) safety and effectiveness information is being submitted in accordance with the requirements of SMDA 1990 and 21 CFR 807.92.

The assigned 510(k) number is: K981812

Applicant Information:

Date Prepared: May 18, 1998
Name: Columbia Bioscience, Inc.
Address: 8775 M Centre Park Drive, #559
Columbia, MD 21045

Contact Person: Norman Jenkins
Phone Number: 410-995-1278
Fax Number: 410-995-0508

Device Information:

Trade Name:  EBV VCA IgG ELISA Kit
Common Name: EBV Viral Capsid Antigen EIA Test
Classification Name: Epstein Barr Virus Serological Reagent

Equivalent Device Description:

Wampole VCA IgG ELISA.

Wampole VCA IgG ELISA kit contains instructions and materials for the qualitative and semi-quantitative detection of IgG antibodies to EBV-VCA IgG in human serum by indirect ELISA

The  EBV-VCA IgG ELISA Kit is an enzyme-linked immunosorbent assay (ELISA) for the detection of IgG to Epstein Barr viral capsid antigen in human serum.

Intended Use: For the qualitative and semi-quantitative determination of IgG antibodies to Epstein-Barr Virus (recombinant) Viral Capsid Antigen (EBV-VCA IgG) in human serum by indirect enzyme immunoassay. The Is-EBV-VCA IgG Test Kit may be used in combination with other Epstein-Barr serologies (Viral Capsid Antigen (VCA) IgM, Epstein-Barr Nuclear Antigen-1 (EBNA-1) IgG and IgM, Early Antigen-Diffuse (EA-D) IgG and IgM and heterophile antibody as an aid in the diagnosis of infectious mononucleosis (IM). The evaluation of paired sera, to determine a significant increase in VCA IgG antibody titer, can also aid in the diagnosis of acute infection. These reagents can be used either manually or in conjunction with the MAGO® Plus Automated Processor.

Principle of Procedure:

Recombinant VCA antigen is bound to microwells. Diluted patient sera, Cut-Off Calibrator and controls are placed in the microwells and incubated. Anti-VCA IgG antibodies, if present, will bind to the antigen forming antigen-antibody complexes. Residual sample is eliminated by aspirating and washing. Conjugate (horseradish peroxidase-labeled anti-human IgG) is added and will bind to these complexes. Unbound conjugate is removed by aspiration and washing. Substrate is then added and incubated. In the presence of bound enzyme the substrate is converted to an end product. The absorbance of this end product can be read spectrophotometrically at 450

nm (reference 600-630 nm) and is directly proportional to the concentration of IgG antibodies to VCA present in the sample.

The Is-EBV-VCA IgG ELISA kit and the Wampole VCA IgG ELISA are substantially equivalent in that

1. Both are *in vitro* immunologic methods.
2. Both are intended for use in the detection of IgG antibody to EBV-VCA in human serum
3. Both are based on the formation of a complex between VCA antigens and antibody
4. Both use antigen coated microtiter plates.
5. Both are qualitative/semi-quantitative assays.
6. Both use goat anti-human IgG conjugated to horseradish peroxidase.
7. Both use TMB as the enzyme substrate.

A detailed comparison between the proposed device and the predicate device is shown in Table 1

Conclusions: The Diamedix Is-EBV-VCA IgG is substantially equivalent to the Wampole VCA ELISA for the detection of IgG antibodies to EBV-VCA in human serum to aid in the diagnosis of infectious mononucleosis. The device is as safe, as effective, and performs as well as the legally marketed device described.

Table 1

1883678

	PROPOSED DEVICE Diamedix Is-EBV-VCA IgG ELISA Kit	PREDICATE DEVICE Wampole EBV-VCA IgG ELISA
Intended Use	For the qualitative and semi-quantitative determination of IgG antibodies to Epstein-Barr Virus Viral Capsid Antigen (EBV-VCA IgG) in human serum by indirect enzyme immunoassay. The Is-EBV-VCA IgG Test Kit may be used in combination with other Epstein-Barr serologies (Viral Capsid Antigen (VCA) IgM, Epstein-Barr Nuclear Antigen-1 (EBNA-1) IgM, Early Antigen-Diffuse (EA-D) IgG and IgM and heterophile antibody as an aid in the diagnosis of infectious mononucleosis (IM). The evaluation of paired sera, to determine a significant increase in VCA IgG antibody titer, can also aid in the diagnosis of acute infection. These reagents can be used either manually or in conjunction with the MAGO® Plus Automated Processor.	The Wampole Laboratories (Wampole) Epstein-Barr Virus-Viral Capsid IgG Enzyme-Linked Immunosorbent Assay (ELISA) is intended for the qualitative and semi-quantitative determination of IgG antibody in human serum to Epstein-Barr virus in human sera. A single serum specimen may be used to indicate previous infection or immune status with the Epstein-Barr virus. Paired sera, acute and convalescent, may be used to demonstrate seroconversion or a significant rise in antibody level as an aid in the diagnosis of a recent or current infection, or reactivation, and as an aid in the diagnosis of Infectious Mononucleosis (IM).
Methodology	Enzyme immunoassay (EIA)	Enzyme Linked Immunosorbent Assay (ELISA)
Specifications	For in vitro diagnostic use. For use with fresh or frozen human serum. Avoid lipemic, hemolyzed, contaminated, or icteric sera. Assay performed on 1:21 dilution of serum at 18-30°C. Store at 2-8°C.	For in vitro diagnostic use. For use with fresh or frozen human serum. Assay performed on 1:21 dilution of serum at 21-25°C. Store at 2-8°C.
Design	Is-EBV-VCA IgG Test Kit. 96 determinations. Undiluted Calibrator, Positive, and Negative controls.	VCA IgG ELISA. 96 determinations. Undiluted Calibrator, High positive, Low positive, and Negative controls.
Principles of Operation	Purified, recombinant VCA antigen is bound to microwells (solid phase). Diluted human serum is added to the microwell which binds human anti-VCA IgG, if present. Solid phase is washed and exposed to anti-human IgG conjugate. Solid phase is washed and exposed to enzyme substrate to develop color. Strong acid is added to stop reaction. The color is read at 450/600 nm on an EIA reader.	Diluted patient serum is incubated with purified, VCA antigen bound to the solid surface of a microtiter well. If IgG antibodies against EBV-VCA are present in the serum, antigen-antibody complexes are formed. These complexes bind with HRP-labeled anti-human IgG which react with the addition of chromogen, resulting in a color development. The absorbance is measured at 450/600 nm.
Performance Characteristics	Relative Sensitivity: 100% Relative Sensitivity (Current Infection): 84.8% Relative Specificity (Seronegative): 93.8% Agreement: 95.7% Intra-assay Precision (Positive samples, all sites) Overall Manual- 1.56-9.30 MAGO Plus- 2.48-12.03 Interassay Precision (Positive samples, all sites) Overall Manual- 4.28-10.50 MAGO Plus- 6.99-9.21 No Cross-reactivity	Relative Sensitivity: 100% Relative Specificity: 100% Agreement: 100% Inter-Site Precision (Positive samples, all sites) Overall: 1.9-9.6% No Cross-reactivity
Enzyme Used	Horesradish Peroxidase	Horesradish Peroxidase
Substrate	TMB	TMB
Specimen	Serum	Serum
Calculation of Results	Sample Absorbance/Cut-off Absorbance = Index Value	Sample Absorbance/Cut-off Absorbance = ISR
Interpretation	<0.90 Negative for VCA IgG 0.90-1.09 Equivocal for VCA IgG ≥ 1.10 Positive for VCA IgG	<0.90 Negative for VCA IgG 0.91-1.09 Equivocal for VCA IgG ≥ 1.10 Positive for VCA IgG
Materials	96 microwells in 12x8 strips, Wash concentrate, Sample Diluent, Conjugate, Calibrator, Controls, Substrate, Stop Solution	96 microwells in 12x8 strips, Wash Buffer, Serum Diluent, HRP Conjugate, Calibrator, Controls, Chromogen, Stop Solution

Performance Characteristics

A. Clinical Sensitivity and Specificity Using Characterized Sera

Frozen retrospective sera from one hundred and seventy-five patients were characterized using commercially available kits for VCA IgG, VCA IgM, EBNA IgG and heterophile antibodies. Based on the results of this testing, the patient sera were characterized as follows :

- 102 sera were characterized as convalescent (past infection). These were positive for VCA IgG and/or EBNA IgG antibodies and negative for VCA IgM and heterophile antibody.
- 34 sera were characterized as seronegative. These were negative for VCA IgG, VCA IgM, EBNA IgG and heterophile antibody.
- 39 sera were characterized as having a current (recent) infection. These were positive for VCA IgM and/or heterophile antibody and were negative for EBNA IgG.

All 175 sera were then tested by an independent clinical commercial laboratory using the Is-EBV-VCA IgG Test Kit. The results obtained are shown in Table 2:

	<i>EBV Serological Status</i>		
	<i>Convalescent</i>	<i>Current Infection</i>	<i>Seronegative</i>
Is-EBV-VCA IgG POSITIVE	99	28	2
NEGATIVE	0	5	30
*EQUIVOCAL	3	6	2

- Of the 102 past infection sera tested, 99 were positive for anti-VCA IgG, none were negative, and 3 were equivocal.
- Of the thirty-seven current (recent) infection samples tested, twenty-eight were positive, five were negative, and 6 were equivocal.
- Of the thirty-four seronegative sera tested, thirty were negative, two were positive, and two were equivocal.
- The overall agreement of the Is-VCA IgG Test Kit compared to EBV serological status was $157/164 = 95.7\%$.

B. Precision

To determine the precision of the Is-EBV-VCA IgG Test Kit, four positive and two negative sera were assayed ten times each in three different runs at three different sites. The three sites included: the manufacturer, a research & development laboratory, and a clinical commercial laboratory. The intra- and interassay precision obtained at each site is shown in Tables 3, 4 and 5.

TABLE 3 : Site #1 - Intra-Assay and Interassay Precision

SERUM	INTRA-ASSAY RUN 1		INTRA-ASSAY RUN 2		INTRA-ASSAY RUN 3		INTERASSAY			
	MEAN INDEX	CV%	MEAN INDEX	CV%	MEAN INDEX	CV%	MEAN INDEX	CV%		
A (POS)	1.74	3.19	1.98	6.84	1.64	7.16	1.79	10.13		
B (POS)	2.05	5.34	2.42	4.11	1.98	6.56	2.15	10.50		
C (POS)	1.42	4.56	1.55	9.30	1.48	9.05	1.48	8.63		
D (POS)	3.12	4.38	3.31	4.49	3.07	4.17	3.17	5.40		
E (NEG)	0.54	16.22	0.58	11.87	0.54	11.17	0.55	13.30		
F (NEG)	0.21	16.31	0.25	21.86	0.21	12.12	0.22	19.06		
							CAL	1.04	13.48	n = 9
							PC	1.75	15.91	n = 3
							NC	0.12	0.00	n = 3

TABLE 4 : Site #2- Intra-Assay and Interassay Precision

SERUM	INTRA-ASSAY RUN 1		INTRA-ASSAY RUN 2		INTRA-ASSAY RUN 3		INTERASSAY			
	MEAN INDEX	CV%	MEAN INDEX	CV%	MEAN INDEX	CV%	MEAN INDEX	CV%		
A (POS)	1.59	4.45	1.52	3.20	1.67	3.23	1.60	5.31		
B (POS)	1.96	3.41	1.87	2.14	2.04	3.45	1.96	4.71		
C (POS)	1.54	3.86	1.33	3.43	1.52	1.56	1.46	7.10		
D (POS)	3.27	2.40	2.90	1.86	3.23	2.74	3.13	5.91		
E (NEG)	0.55	5.11	0.55	2.29	0.64	3.76	0.58	8.22		
F (NEG)	0.231	3.36	0.25	4.86	0.28	9.96	0.25	10.66		
							CAL	1.00	1.80	n = 18
							PC	1.47	4.24	n = 12
							NC	0.16	6.13	n = 12

TABLE 5 : Site #3 - Intra-assay and Interassay Precision

SERUM	INTRA-ASSAY RUN 1		INTRA-ASSAY RUN 2		INTRA-ASSAY RUN 3		INTERASSAY			
	MEAN INDEX	CV%	MEAN INDEX	CV%	MEAN INDEX	CV%	MEAN INDEX	CV%		
A (POS)	1.59	3.51	1.56	5.21	1.52	3.00	1.55	4.37		
B (POS)	1.91	4.17	1.89	4.58	1.85	6.07	1.88	5.04		
C (POS)	1.44	4.60	1.36	3.53	1.37	3.26	1.39	4.61		
D (POS)	3.05	4.38	3.02	3.43	2.91	3.88	2.99	4.28		
E (NEG)	0.55	6.77	0.60	7.37	0.54	5.08	0.56	8.25		
F (NEG)	0.28	8.78	0.36	64.52	0.27	15.22	0.30	46.46		
							CAL	1.00	3.87	n = 9
							PC	1.33	16.45	n = 3
							NC	0.20	28.00	n = 3


C. Specificity with Potentially Cross-Reactive Sera

Sixteen sera, non-reactive (negative) for IgG antibodies to VCA in the Is-EBV-VCA IgG Test Kit, were tested by EIA for IgG antibody to varicella zoster, cytomegalovirus and herpes simplex virus. 15/15 anti-VZV IgG positive sera were non-reactive for anti-VCA IgG; 3/3 anti-CMV IgG positive sera were non-reactive for anti-VCA IgG and 3/3 anti-HSV positive sera were non-reactive for anti-VCA IgG. This suggests that no specific cross-reactivity should be expected with the Is-EBV-VCA IgG Test Kit from these analytes.

D. Correlation of Manual and MAGO Plus Results

The Is-EBV-VCA IgG Test Kit has been developed for automated as well as manual use. To demonstrate the equivalence of the manual and MAGO Plus procedures, the results of 196 serum samples tested by both methods were plotted. A scattergram and regression line of the results obtained with 95% confidence intervals is shown in Figure 3. The data indicate good correlation with a Pearson Correlation Coefficient of 0.986.

FIGURE 3 : Manual and MAGO Plus Result Correlation



D. MAGO Plus Precision

The precision of the assay when performed on the MAGO Plus Automated EIA Processor was determined by assaying six sera ten times each in three different runs. Table 6 shows the intra-and interassay precision obtained using the MAGO Plus.

TABLE 6 : Site #2- Intra-Assay and Interassay Precision - MAGO Plus

SERUM	INTRA-ASSAY RUN 1		INTRA-ASSAY RUN 2		INTRA-ASSAY RUN 3		INTERASSAY			
	MEAN INDEX	CV%	MEAN INDEX	CV%	MEAN INDEX	CV%	MEAN INDEX	CV%		
A (POS)	1.68	3.76	1.52	4.16	1.67	7.50	1.62	6.99	n = 9	
B (POS)	2.15	5.48	1.89	4.63	2.07	4.58	2.04	7.23	n = 3	
C (POS)	1.63	2.96	1.44	5.86	1.49	12.03	1.52	9.21	n = 3	
D (POS)	3.57	3.75	3.06	3.51	3.18	2.48	3.27	7.50		
E (NEG)	0.60	0.00	0.52	8.11	0.51	11.13	0.54	10.46		
F (NEG)	0.30	15.71	0.25	21.08	0.21	15.06	0.25	22.56		
							CAL	1.00	5.54	n = 9
							PC	1.43	4.03	n = 3
							NC	0.20	0.00	n = 3



MAR - 4 1999

Diamedix Corp.
c/o Norman Jenkins
President
Columbia Bioscience, Inc.
8775 M Centre Park Drive, #559
Columbia, MD 21045

Food and Drug Administration
2098 Gaither Road
Rockville MD 20850

Re: K981812
Trade Name: EBV-VCA IgG ELISA Test System
Regulatory Class: I
Product Code: LSE
Dated: December 8, 1998
Received: December 9, 1998

Dear Mr. Jenkins:

We have reviewed your Section 510(k) notification of intent to market the device referenced above and we have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (Premarket Approval), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 895. A substantially equivalent determination assumes compliance with the Current Good Manufacturing Practice requirements, as set forth in the Quality System Regulation (QS) for Medical Devices: General regulation (21 CFR Part 820) and that, through periodic QS inspections, the Food and Drug Administration (FDA) will verify such assumptions. Failure to comply with the GMP regulation may result in regulatory action. In addition, FDA may publish further announcements concerning your device in the Federal Register. Please note: this response to your premarket notification submission does not affect any obligation you might have under sections 531 through 542 of the Act for devices under the Electronic Product Radiation Control provisions, or other Federal laws or regulations.


Page 2

Under the Clinical Laboratory Improvement Amendments of 1988 (CLIA-88), this device may require a CLIA complexity categorization. To determine if it does, you should contact the Centers for Disease Control and Prevention (CDC) at (770)488-7655.

This letter will allow you to begin marketing your device as described in your 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801 and additionally 809.10 for in vitro diagnostic devices), please contact the Office of Compliance at (301) 594-4588. Additionally, for questions on the promotion and advertising of your device, please contact the Office of Compliance at (301) 594-4639. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR 807.97). Other general information on your responsibilities under the Act may be obtained from the Division of Small Manufacturers Assistance at its toll free number (800) 638-2041 or at (301) 443-6597 or at its internet address "<http://www.fda.gov/cdrh/dsmamain.html>"

Sincerely yours,



Steven I. Gutman, M.D., M.B.A.
Director
Division of Clinical Laboratory Devices
Office of Device Evaluation
Center for Devices and Radiological Health

Enclosure

510(k) Number: K981812

Device Name:  EBV- VCA IgG ELISA

Indications For Use: For the qualitative and semi-quantitative determination of IgG antibodies to Epstein-Barr Virus (recombinant) Viral Capsid Antigen (EBV-VCA IgG) in human serum by indirect enzyme immunoassay. The Is-EBV-VCA IgG Test Kit may be used in combination with other Epstein-Barr serologies (Viral Capsid Antigen (VCA) IgM , Epstein-Barr Nuclear Antigen-1 (EBNA-1) IgG and IgM, Early Antigen-Diffuse (EA-D) IgG and IgM and heterophile antibody as an aid in the diagnosis of infectious mononucleosis (IM). The evaluation of paired sera, to determine a significant increase in VCA IgG antibody titer, can also aid in the diagnosis of acute infection. These reagents can be used either manually or in conjunction with the MAGO® Plus Automated Processor.

PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE
IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription Use
(Per 21 CFR 801.109)

OR

Over-The Counter Use
(Optional Format 1-2-96)



(Division Sign/Off)
Division of Clinical Laboratory Devices
510(k) Number K981812