

4/20/99

**510(k) Summary of Safety and Effectiveness**

This 510(k) summary of safety and effectiveness information is being submitted in accordance with the requirements of SMDA 1990 and 21 CFR 807.92.

The assigned 510(k) number is:     K981995    

**Applicant Information:**

Date Prepared:      October 22, 1998  
Name:                 Heartport, Inc.  
Address:             200 Chesapeake Drive  
                          Redwood City, CA 94063

Contact Person:     Marianne C. Drennan, Regulatory Affairs Manager  
Phone Number:      (650) 482-4405  
Fax Number:         (650) 482-4346

**Device Information:**

Classification        Class II  
Trade Name:         Heartport® QuickDraw™ venous cannula  
Classification Name: Cardiopulmonary Bypass Cannula  
                          21 CFR §870.4210

**Equivalent Devices:**

The subject device is substantially equivalent in intended use, method of operation and/or materials to the currently marketed Heartport Endovenous Drainage™ Cannula, Baxter Healthcare Corporation- Cardiovascular Group [Research Medical, Inc. (RMI)] RMI Venous Drainage Cannula, and the Heartport Endoarterial Return™ Cannula

**Intended Use:**

The Heartport QuickDraw venous cannula is indicated for use in patients undergoing cardiopulmonary bypass. The Heartport QuickDraw venous cannula serves to drain non-oxygenated blood from the venae cavae and/or right atrium during cardiopulmonary bypass.

**Non-Clinical Test Results:**

Performance testing demonstrated that the subject device meets established specifications. The materials used in the Heartport QuickDraw venous cannula have proven biocompatibility.

**Summary:**

Based on the intended use, product performance and biocompatibility information provided in this notification, the subject device has been shown to be substantially equivalent to currently marketed predicate devices.



Food and Drug Administration  
9200 Corporate Boulevard  
Rockville MD 20850

APR 20 1999

Ms. Marianne C. Drennan  
Regulatory Affairs Manager  
HeartPort, Inc.  
700 Bay Road  
Redwood City, CA 94063

Re: K981995  
HeartPort™ QuickDraw™ Venous Cannula  
Regulatory Class: II (Two)  
Product Code: 74 DWF  
Dated: January 27, 1999  
Received: January 28, 1999

Dear Ms. Drennan:

We have reviewed your Section 510(k) notification of intent to market the device referenced above and we have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.


If your device is classified (see above) into either class II (Special Controls) or class III (Premarket Approval), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 895. A substantially equivalent determination assumes compliance with the Current Good Manufacturing Practice requirements, as set forth in the Quality System Regulation (QS) for Medical Devices: General regulation (21 CFR Part 820) and that, through periodic QS inspections, the Food and Drug Administration (FDA) will verify such assumptions. Failure to comply with the GMP regulation may result in regulatory action. In addition, FDA may publish further announcements concerning your device in the Federal Register. Please note: this response to your premarket notification submission does not affect any obligation you might have under sections 531 through 542 of the Act for devices under the Electronic Product Radiation Control provisions, or other Federal laws or regulations.

Page 2 - Ms. Marianne C. Drennan

This letter will allow you to begin marketing your device as described in your 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801 and additionally 809.10 for in vitro diagnostic devices), please contact the Office of Compliance at (301) 594-4648. Additionally, for questions on the promotion and advertising of your device, please contact the Office of Compliance at (301) 594-4639. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR 807.97). Other general information on your responsibilities under the Act may be obtained from the Division of Small Manufacturers Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its internet address "<http://www.fda.gov/cdrh/dsma/dsmamain.html>".

Sincerely yours,



Thomas J. Callahan, Ph.D.  
Director  
Division of Cardiovascular, Respiratory  
and Neurological Devices  
Office of Device Evaluation  
Center for Devices and  
Radiological Health

Enclosure

**Indications for Use**

510(k) Number (if known): K 98 1995

Device Name: Heartport® QuickDraw™ venous cannula

**Indications for Use:**

The QuickDraw venous cannula is indicated for use in patients undergoing cardiopulmonary bypass. The QuickDraw venous cannula serves to drain non-oxygenated blood from the venae cavae and/or right atrium during cardiopulmonary bypass.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

(Division Sign-Off)  
Division of Cardiovascular, Respiratory,  
and Neurological Devices

510(k) Number K 98 1995

Prescription Use              
(Per 21 CFR 801.109)

OR

Over- The Counter Use             
(Optional Format 1-2-96)