

K98 3902

JAN 7 1999

Attachment 1
510(k) Summary of Safety and
Effectiveness

1.0 SUBMITTER INFORMATION:

- 1.1 Submitter: Field Effects, a division of Intermagnetics General Corporation
300 Vesper Executive Park
Tyngsborough, MA 01879
PH: **978** 649-8590
FX: **978** 649-8520
Email: GHURST@IGC.com
- 1.2 Contact: Greg Hurst, Ph.D.
Mgr., Imaging Science and Applications
- 1.3 Date: October 30, 1998
- 1.4 Regulatory Counsel:
Jonathan S. Kahan, Attorney at Law
Hogan & Hartson
555 Thirteenth St. NW
Washington, DC 20004-1109
PH: **202** 637-5794
FX: **202** 637-5910

2.0 DEVICE NAME:

- 2.1 Classification Panel: Radiology
- 2.2 Classification Number: 892.1000 Magnetic Resonance Diagnostic Device
- 2.3 Product Nomenclature: System, Nuclear Magnetic Resonance Imaging
- 2.4 Product Code(s): 90LNH
- 2.5 Trade/Proprietary Name: Field Effects MRI
- 2.6 Predicate Device(s): IMiG-MRI, K963953
Hitachi MRP-5000 with Version 5.0 Operating System
Software, K943799

3.0 DEVICE DESCRIPTION:

3.1 FUNCTION

Identical to the IMiG-MRI 510(k) K963953. The original IMiG-MRI system was renamed as TREX MRI in 510(k) submission K982157. As a result of changing business arrangements, the TREX MRI is now renamed as the Field Effects MRI System. In addition, following successful FDA determination of substantial equivalence of this 510(k) submission for the Image Processing software, Field Effects will cease commercialization efforts and support for the Field Effects MRI System. Service Support will be available through SMIS (refer to Item 9, Address of Manufacturing Facilities for SMIS contact information). With the exception of the device modifications identified in this 510(k) submission, and the renaming of the product, the Field Effects MRI System is identical to the original IMiG-MRI system (and to the TREX.MRI).

The Field Effects MRI 0.15T Elliptical MRI Magnetic Resonance Diagnostic Device is being enhanced by image processing software to increase the clinical utility of the Field Effects MRI System in the stationary configuration.

3.2 Safety Parameter Summary

Maximum static magnetic field	0.15 Tesla
Maximum rate of magnetic field change	18.4 Tesla/sec
Maximum RF power deposition	0.05W/kg
Acoustic noise levels	114dB peak; 95dB A-weighted RMS

3.3 Performance Parameter Summary

Identical to the Predicate Device.

3.4 General Safety and Effectiveness Summary

Safe and effective use of the machine is assured by the associated labeling. This labeling includes: advertising brochures, Site Planning Guide, and Instructions for Use (comprised of Clinical Users Guide, User Safety Guide, User Training Guide, User Applications Guide, and User QA & Maintenance Guide).

4.0 DEVICE INTENDED USE:

The Field Effects MRI system produces cross-sectional images:

- Anatomical Region: General body anatomy, including head, spine, torso, and extremities
- Nucleus excited: ^1H nuclei (Proton)
- Diagnostic uses: Images correspond to the distribution of ^1H nuclei exhibiting nuclear magnetic resonance, with image intensity dependent upon NMR parameters, including spin-lattice relaxation time (T1)
spin-spin relaxation time (T2)
density of nuclei (ρ)
flow velocity
chemical shift (δ)
- Clinical use: Images may be interpreted by a trained physician to yield information that can be useful in the determination of a diagnosis

5.0 DEVICE TECHNOLOGICAL CHARACTERISTICS:

Identical to the Predicate Device.



JAN 7 1999

Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

Ian L. Pykett, Ph.D., F.Inst.P.
Vice President, Technology Development
Intermagnetics General Corporation
450 Old Niskayuna Road
P.O. Box 461
Latham, NY 12110-0461

Re: K983902
Field Effects MRI Magnetic Resonance
Diagnostic Device
Dated: October 30, 1998
Received: November 3, 1998
Regulatory class: II
21 CFR 892.1000/Procode: 90 LNH

Dear Dr. Pykett:


We have reviewed your Section 510(k) notification of intent to market the device referenced above and we have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (Premarket Approval), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 895. A substantially equivalent determination assumes compliance with the Current Good Manufacturing Practice requirements, as set forth in the Quality System Regulation (QS) for Medical Devices: General regulation (21 CFR Part 820) and that, through periodic QS inspections, the Food and Drug Administration (FDA) will verify such assumptions. Failure to comply with the GMP regulation may result in regulatory action. In addition, FDA may publish further announcements concerning your device in the Federal Register. Please note: this response to your premarket notification submission does not affect any obligation you might have under sections 531 through 542 of the Act for devices under the Electronic Product Radiation Control provisions, or other Federal laws or regulations.

This letter will allow you to begin marketing your device as described in your 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801 and additionally 809.10 for in vitro diagnostic devices), please contact the Office of Compliance at (301) 594-4613. Additionally, for questions on the promotion and advertising of your device, please contact the Office of Compliance at (301) 594-4639. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR 807.97). Other general information on your responsibilities under the Act may be obtained from the Division of Small Manufacturers Assistance at its toll-free number (800) 638-2041 or (301) 443-6597, or at its internet address "<http://www.fda.gov/cdrh/dsma/dsmamain.html>".

Sincerely yours,



Capt. Daniel G. Schultz, M.D.
Acting Director, Division of Reproductive,
Abdominal, Ear, Nose and Throat,
and Radiological Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

510(k) Number (if known): K983902

Device Name: Image Processing Software for Field Effects MRI

Indications for Use:

The Field Effects MRI system produces cross-sectional images:

- Anatomical Region: General body anatomy, including head, spine, torso, and extremities
- Nucleus excited: ¹H nuclei (Proton)
- Diagnostic uses: Images correspond to the distribution of ¹H nuclei exhibiting nuclear magnetic resonance, with image intensity dependent upon NMR parameters, including
 - spin-lattice relaxation time (T1)
 - spin-spin relaxation time (T2)
 - density of nuclei (ρ)
 - flow velocity
 - chemical shift (δ)
- Clinical use: Images may be interpreted by a trained physician to yield information that can be useful in the determination of a diagnosis.
- RF Coils:
 - Head - quadrature
 - Cervical Spine - quadrature
 - Lumbar Spine, Thoracic Spine, and Abdomen - quadrature
 - Knee - linear
- Image Acquisition
 - Spin Echo
 - Gradient Echo (GE, RGE)
 - Fast Spin Echo (FSE)
 - Fast Dual Echo (FDE)
 - 3D Fast Gradient Echo (3DGE, 3DRGE)
 - Inversion Recovery with Fast Spin Echo (FSE STIR, FSE FLAIR)

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

[Signature] 12/21/98

(Division Sign-Off)
Division of Reproductive, Abdominal, ENT, and Radiological Devices Over-the-Counter Use _____

Prescription Use
(Per 21 CFR 801-109)

510(k) Number K983902

(Optional Format 1-2-96)