

K992452

OCT 18 1999

## 510(k) Summary

**Submitter's Name/Address:**  
 American Bio Medica Corporation  
 122 Smith Road  
 Kinderhook, N.Y. 12106

**Contact Person:**  
 Henry Wells  
 Vice President of Product  
 Development  
 Phone: 518-758-8158  
 Fax: 518-758-8171

**Date of Preparation of this Summary:**

September 24, 1999

**Device Trade or Proprietary Name:**

'RapidOne'-Amphetamine Test

**Device Common/Usual Name or  
 Classification Name:**

Amphetamine test system

**Classification Number/Class**

21 C.F.R. § 862.3650 (Class II)

This 510(k) Summary is being submitted in accordance with the requirements of 21 CFR 807.92.

The assigned 510(k) is:

**Predicate Device:** Forefront Diagnostics, Inc. 'InstaCheck' Drug Screen -  
 Amphetamine Test (510(k) No.K971960.)

**Test Description:**

The assay employed in the 'RapidOne'- Amphetamine Test is based on the same principle of highly specific reaction between antigens and antibodies.

This assay is a one-step, immunoassay in which a specially labeled drug (drug conjugate) competes with drug which may be present in the sample for the limited number of binding sites on the antibody. The test device consists of a membrane strip onto which anti-amphetamine monoclonal antibody has been immobilized. A colloidal gold-BSA-amphetamine complex is dried on a reagent pad. In the absence of any drug in the urine sample, the colloidal gold-complex moves with the urine by capillary action to contact the immobilized drug antibody. An antibody-antigen reaction occurs forming a visible line in the 'test' area. The formation of a visible line in the test area occurs when the test is negative.

When drug is present in the urine sample, the drug or metabolite will compete with the immobilized drug conjugate in the test area for the limited antibody sites on the colloidal gold-antibody complex. If sufficient amount of drug is present, it will fill all of the available binding sites, thus preventing attachment of the labeled antibody to the drug



conjugate. An absence of a color band (line) in the test area is indicative of a positive result.

A control band (line), comprised of a different antibody/antigen reaction, is present on the membrane strip. The control line is not influenced by the presence or absence of drug in the urine, and therefore, should be present in all reactions.

A negative urine will produce two colored bands, and a positive sample will produce only one band.

#### Intended Use:

'RapidOne'- Amphetamine Test is used for the qualitative detection of d-amphetamine in human urine. This immunoassay is a simplified qualitative screening method that provides only a preliminary result for use in determining the need for additional or confirmatory testing, i.e., gas chromatography/mass spectrometry (GC/MS).

#### Performance Characteristics:

'RapidOne'-Amphetamine Test will detect 1000 ng/ml of d-amphetamine in urine.

'RapidOne'-Amphetamine Test was compared to EMIT II. Ninety (90) samples were selected for evaluation, fifty (50) of which were found to be drug-free and forty (40) tested as positive by Syva EMIT-II. The forty positive specimens were confirmed and quantified by GC/MS. Both immunoassays correctly identified all of the specimens which contained no drug as negative and determined the 40 drug-containing specimens, ranging in concentration of 750 ng/ml to 58,000 ng/ml, to be positive.

Reproducibility was evaluated using control urines containing concentrations above and below the stated cut-off. Negative urines were also used. Each sample, at each concentration of analyte, was tested 4 times, twice daily, for 5 days.

Conc. (ng/ml)	#	Result
0	40	40/40 neg
500	40	5/40 pos
750	40	30/40 pos
1000	40	40/40 pos
1250	40	40/40 pos

#### Conclusion:

'RapidOne'-Amphetamine Test is substantially equivalent to 'Instacheck' - Drug Screen - Amphetamine Test



DEPARTMENT OF HEALTH & HUMAN SERVICES

Food and Drug Administration  
2098 Gaither Road  
Rockville MD 20850

OCT 18 1999

American Bio Medica  
c/o Mr. John B. Dubeck  
KELLER AND HECKMAN LLP  
1001 G Street, N.W.  
Suite 500 West  
Washington, D.C. 20001

Re: K992452  
Trade Name: 'Rapid One' – Amphetamine Test  
Regulatory Class: II  
Product Code: DKZ, DOD  
Dated: July 22, 1999  
Received: July 22, 1999

Dear Mr. Dubeck:

We have reviewed your Section 510(k) notification of intent to market the device referenced above and we have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (Pre-market Approval), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 895. A substantially equivalent determination assumes compliance with the Current Good Manufacturing Practice requirements, as set forth in the Quality System Regulation (QS) for Medical Devices: General regulation (21 CFR Part 820) and that, through periodic QS inspections, the Food and Drug Administration (FDA) will verify such assumptions. Failure to comply with the GMP regulation may result in regulatory action. In addition, FDA may publish further announcements concerning your device in the Federal Register. Please note: this response to your premarket notification submission does not affect any obligation you might have under sections 531 through 542 of the Act for devices under the Electronic Product Radiation Control provisions, or other Federal laws or regulations.

Page 2

Under the Clinical Laboratory Improvement Amendments of 1988 (CLIA-88), this device may require a CLIA complexity categorization. To determine if it does, you should contact the Centers for Disease Control and Prevention (CDC) at (770) 488-7655.

This letter will allow you to begin marketing your device as described in your 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801 and additionally 809.10 for in vitro diagnostic devices), please contact the Office of Compliance at (301) 594-4588. Additionally, for questions on the promotion and advertising of your device, please contact the Office of Compliance at (301) 594-4639. Also, please note the regulation entitled, "Misbranding by reference to premarket notification"(21 CFR 807.97). Other general information on your responsibilities under the Act may be obtained from the Division of Small Manufacturers Assistance at its toll-free number (800) 638-2041 or (301) 443-6597, or at its internet address "<http://www.fda.gov/cdrh/dsma/dsmamain.html>".

Sincerely yours,



Steven I. Gutman, M.D, M.B.A.  
Director  
Division of Clinical  
Laboratory Devices  
Office of Device Evaluation  
Center for Devices and  
Radiological Health

Enclosure

510(k) Number (if known): K992452

Device Name: 'RapidOne' - Amphetamine Test

Indications For Use:

'RapidOne' - Amphetamine Test is a one-step, lateral flow immunoassay for the detection of amphetamine in urine. 'RapidOne' - Amphetamine Test is intended for use in the qualitative detection of amphetamine in human urine at 1000 ng/ml.

'RapidOne' - Amphetamine Test is intended for professional use. It is not intended for over the counter sale to non-professionals. The assay is easy to perform, but should not be used without proper supervision. This immunoassay is a simplified qualitative screening method that provides only a preliminary result for use in determining the need for additional or confirmatory testing, i.e., gas-chromatography/mass spectrometry (GC/MS).

'RapidOne' - Amphetamine Test provides only a preliminary analytical test result. A more specific alternate chemical method must be used in order to obtain a more confirmed analytical result. Gas chromatography/mass spectrometry (GC/MS) is the preferred confirmatory method. Clinical and professional judgment should be applied to any drug of abuse test result, particularly when preliminary positive results are used.

Sean Cooper  
(Division Sign-Off)  
Division of Clinical Laboratory Devices  
510(k) Number K 992452

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription Use   
(Per 21 CFR 801.109)

OR

Over-The-Counter Use

(Optional Format 1-2-96)