

K101913

1073

### 510(k) SUMMARY

MAR 11 2011

As required by section 807.92(c)

Submitter	<b>ANTHOGYR SAS</b> 2237, avenue Andre Lasquin SALLANCHES, FRANCE 74700 Registration Number:8020776
Contacts	Eric GENEVE <a href="mailto:e.geneve@anthogyr.com">e.geneve@anthogyr.com</a> Phone (33) (0)4 50 58 02 37 Fax (33) (0)4 50 93 78 60 Regulatory: Dr Isabelle DRUBAIX (PhD) IDEE CONSULTING <a href="mailto:idrubaix@nordnet.fr">idrubaix@nordnet.fr</a>
Preparation date	June 16, 2010
Trade Name	ANTHOGYR DENTAL IMPLANT SYSTEMS : AXIOM
Classification Name	ENDOSSEOUS DENTAL IMPLANT
Class	II
Product Code	DZE
CFR section	21CFR 872.3640
Device panel	DENTAL
Legally marketed predicate devices	PRIMACONNEX INTERNAL CONNECTION IMPLANT SYSTEM (K051614)

#### **Description**

AXIOM dental implant system is designed to enhance the functional and aesthetic integration of implant supported restorations. AXIOM implants are made of grade V titanium and present a single connection (2.7 mm) common to all implant forms in the range. AXIOM implants are available in 4 diameters (3.4, 4.0, 4.6 and 5.2 mm) and various heights from 8 to 14 mm.

AXIOM dental implant system includes all required prosthetic components and surgical instrumentation.

#### **Intended Use**

ANTHOGYR Dental Implant System implant AXIOM is intended for use in partially or fully edentulous mandibles and maxillae, in support of single or multiple-unit restorations including; cement retained, screw retained, or overdenture restorations, and terminal or intermediate abutment support for fixed bridgework.

K101913

2073

**Performance data**

ANTHOGYR Endosseous dental implant system AXIOM conforms to Class II Special Controls Guidance Document: Root-form Endosseous Dental Implants and Endosseous Dental Abutments issued on May 12, 2004. Non clinical performance data presented within the submission included: Fatigue testing, Implant to abutment compatibility, corrosion testing and modified surfaces information.

**Clinical data**

None included

**Substantial equivalence**

ANTHOGYR Endosseous dental implant system AXIOM is substantially equivalent to its predicate device in terms of intended use, material, design, mechanical properties and function.

	<b>AXIOM</b>	<b>PRIMACONNEX (K051614)</b>
21CFR / Product's code / Panel	21CFR 872.3640 / DZE / DENTAL PANEL	
Indications for use statement	Both Dental Implant Systems are intended for use in partially or fully edentulous mandibles and maxillae, in support of single or multiple-unit restorations including; cement retained, screw retained, or overdenture restorations, and terminal or intermediate abutment support for fixed bridgework.	
Device	Root form endosseous dental implants and abutments	
Range of products	Both Dental Implant Systems include implants, abutments, all required prosthetic components and surgical instrumentation	
Design and models	Internal lobe connection	
	Designed to allow clinical flexibility in a single implant system - Time-saving technologies - Intuitive user-friendly design	
	Implants made of Titanium	
	Unique prosthetic platform	3 prosthetic platforms
	Implant $\phi$ : 3.4 -5.2 mm	Implant $\phi$ : 3.5 -5.0 mm
Surface treatment	Sand blasting roughened surface using calcium phosphate resorbable blast media	
	Bi Calcium Phosphate (BCP)	Resorbable Blast Media (RBM)
Sterilization	Single-use - Provided sterile - Gamma radiation	
Labeling	Both Dental Implant Systems share same intended use, contra-indications, warnings precautions and potential adverse events	
Performance	Both Dental Implant Systems conform to Class II Special Controls Guidance Document:: Root-form Endosseous Dental Implants and Endosseous Dental Abutments - Guidance for Industry and FDA Staff Document issued on May 12, 2004	

Performance data included within this submission demonstrates safety, effectiveness and substantial equivalence.  
Revised January 20, 2011

K101913

373

1. The table under Tab 4, on page 44/45 indicates that your OSSFIT has abutments. This table is incorrect because this implant is a one piece implant. Please submit a revised table.

Not applicable – OSSFIT has been deleted from K101913



Food and Drug Administration  
10903 New Hampshire Avenue  
Document Control Room -WO66-G609  
Silver Spring, MD 20993-0002

Mr. Eric Geneve  
CEO  
ANTHOGYR SAS  
2237 avenue Andre Lasquin  
SALLANCHES, FRANCE 74700

MAR 11 2011

Re: K101913

Trade/Device Name: Anthogyr Dental Implant System: Axiom  
Regulation Number: 21 CFR 872.3640  
Regulation Name: Endosseous Dental Implant  
Regulatory Class: II  
Product Code: DZE  
Dated: March 1, 2011  
Received: March 7, 2011

Dear Mr. Geneve:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Page 2- Mr. Geneve


Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to

<http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,



Anthony D. Watson, B.S., M.S., M.B.A.

Director

Division of Anesthesiology, General Hospital,

Infection Control and Dental Devices

Office of Device Evaluation

Center for Devices and

Radiological Health

Enclosure

K101913

1 of 1

**INDICATIONS FOR USE**

**510(k) Number (if known):** \_\_\_\_\_

**Device Name:** ANTHOGR DENTAL IMPLANT SYSTEM: AXIOM

**Indications for Use:**

ANTHOGR Dental Implant System AXIOM is intended for use in partially or fully edentulous mandibles and maxillae, in support of single or multiple-unit restorations including; cement retained, screw retained, or overdenture restorations, and terminal or intermediate abutment support for fixed bridgework.

\_\_\_\_\_  
Prescription Use  AND/OR Over-The-Counter Use \_\_\_\_\_  
(Part 21 CFR 801 Subpart D) (21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE OF  
NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Susan Perro

(Division Sign-Off)

Division of Anesthesiology, General Hospital

Infection Control, Dental Devices

510(k) Number: K101913