



K102388  
NOV - 4 2010

---

**Section 5 — 510(k) Summary**

Vivid i/q

---

**510(k) Summary**

In accordance with 21 CFR 807.92 the following summary of information is provided:

- Date:** 18th August 2010
- Submitter:** GE Healthcare [GE Medical Systems Ultrasound and Primary Care Diagnostics, LLC]  
9900 Innovation Dr  
Wauwatosa, WI 53226
- Primary Contact Person:** Bryan Behn  
Regulatory Affairs Manager  
GE Healthcare, [GE Medical Systems Ultrasound and Primary Care Diagnostics, LLC]  
T:(414)721-4214  
F:(414)918-8275  
GE Healthcare
- Secondary Contact Person:** Jim Turner  
Regulatory Affairs Manager America's Service  
GE Healthcare, [GE Medical Systems Ultrasound and Primary Care Diagnostics, LLC]  
T:(262) 544-3359  
F:(414)908-9225
- Device: Trade Name:** Vivid i and Vivid q Diagnostic Ultrasound System
- Common/Usual Name:** Vivid i, Vivid q
- Classification Names:** Class II
- Product Code:** Ultrasonic Pulsed Doppler Imaging System, 21 CFR 892.1550, 90-IYN  
Ultrasonic Pulsed Echo Imaging System, 21 CFR 892.1560, 90-IYO  
Diagnostic Ultrasonic Transducer, 21 CFR 892.1570, 90-ITX
- Predicate Device(s):** Vivid i and Vivid q Diagnostic Ultrasound Systems K092140 and Vivid S5 and Vivid S6 Diagnostic Ultrasound Systems K092079, currently in commercial distribution.
- Device Description:** The Vivid-i and Vivid-q are compact and portable diagnostic ultrasound systems with integrated keyboard, fold-up LCD type display and interchangeable electronic-array transducers. They have an overall size approximately 36 cm wide, 31.5 cm deep and 6 cm high in transport configuration and provide digital acquisition, processing and display capability. The user interface includes a computer keyboard, an intuitive layout of specialized controls, color GUI display and Doppler audio.



**Intended Use:** The current modifications do not change the indications for use. As previously reported and cleared, the Vivid i/q ultrasound systems are intended for use by, or under the direction of, a qualified physician for ultrasound imaging and analysis in Abdominal/GYN; Urology; Fetal/OB; Small Organ (breast, testes, thyroid); Pediatric; Neonatal & Adult Cephalic; Cardiac (adult & pediatric); Peripheral Vascular; Musculo-skeletal (conventional & superficial); Transesophageal; Intraoperative (abdominal, thoracic and PV); Transvaginal and Transrectal, Intra-cardiac and intra-luminal applications.



Technology: The modified Vivid i/q employs the same fundamental scientific technology as its predicate devices.

Comparison with the predicate device shows the modified Vivid i/q is of a comparable type and substantially equivalent to the current Vivid i/q. It has the same overall characteristics, key safety and effectiveness features, physical design, construction, and materials, and has the same intended uses and operating modes as the predicate device.

Determination of  
Substantial Equivalence:

Summary of Non-Clinical Tests:

The device has been evaluated for acoustic output, biocompatibility, cleaning and disinfection effectiveness as well as thermal, electrical, electromagnetic, and mechanical safety, and has been found to conform with applicable medical device safety standards. The modified Vivid i/q and its applications comply with voluntary standards as detailed in Section 9, 11 and 17 of this premarket submission. The following quality assurance measures were applied to the development of the system:

- Risk Analysis
- Requirements Reviews
- Design Reviews
- Testing on unit level (Module verification)
- Integration testing (System verification)
- Final Acceptance Testing (Validation)
- Performance testing (Verification)
- Safety testing (Verification)

Transducer materials and other patient contact materials are biocompatible.

Summary of Clinical Tests:

The subject of this premarket submission, the modified Vivid i/q, did not require clinical studies to support substantial equivalence.

Conclusion: GE Healthcare considers the modified Vivid i/q to be as safe, and effective as the predicate device(s). The performance of the modified Vivid i/q is substantially equivalent to the predicate device(s).

Intended uses and other key features are consistent with traditional clinical practice, FDA guidelines, and established methods of patient examination. The design and development process of the manufacturer conforms with 21 CFR 820, and ISO13485 quality systems. The device conforms to applicable medical device safety standards and compliance is verified through independent evaluation with ongoing factory surveillance. Diagnostic ultrasound has accumulated a long history of safe and effective performance. Therefore, it is the opinion of GE Healthcare that the Vivid i/q Diagnostic Ultrasound is substantially equivalent with respect to safety and effectiveness to devices currently cleared for market.



Food and Drug Administration  
10903 New Hampshire Avenue  
Silver Spring, MD 20993

NOV - 4 2010

Mr. Bryan Behn  
Regulatory Affairs Manager  
GE Healthcare (GE Medical Systems Ultrasound and Primary Care Diagnostics, LLC)  
9900 Innovation Dr.  
WAUWATOSA WI 53226

Re: K102388

Trade/Device Name: Vivid i and Vivid q Diagnostic Ultrasound System  
Regulation Number: 21 CFR 892.1550  
Regulation Name: Ultrasonic pulsed doppler imaging system  
Regulatory Class: II  
Product Code: IYN, IYO, and ITX  
Dated: October 4, 2010  
Received: October 5, 2010

Dear Mr. Behn:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and we have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

This determination of substantial equivalence applies to the following transducers intended for use with the Vivid i and Vivid q Diagnostic Ultrasound System, as described in your premarket notification:

Transducer Model Number

9L-RS

E8C-RS

6Tc-RS

6T-RS

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 895. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to begin marketing your device as described in your premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus permits your device to proceed to market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to <http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

If you have any questions regarding the content of this letter, please contact Jana Delfino at (301) 796-6503.

Sincerely yours,



David G. Brown, Ph.D.  
Acting Director  
Division of Radiological Devices  
Office of *In Vitro* Diagnostic Device  
Evaluation and Safety  
Center for Devices and Radiological Health

Enclosure(s)



510(k) Number (if known):

Device Name: Vivid i/q

K102388

Indications for Use:

NOV - 4 2010

The current modifications do not change the indications for use. As previously reported and cleared, the Vivid i/q ultrasound systems are intended for use by, or under the direction of, a qualified physician for ultrasound imaging and analysis in Abdominal/GYN; Urology; Fetal/OB; Small Organ (breast, testes, thyroid); Pediatric; Neonatal & Adult Cephalic; Cardiac (adult & pediatric); Peripheral Vascular; Musculo-skeletal (conventional & superficial); Transesophageal; Intraoperative (abdominal, thoracic and PV); Transvaginal and Transrectal, Intra-cardiac and intra-luminal applications.

Prescription Use   X   AND/OR Over-The-Counter Use       

(Part 21 CFR 801 Subpart D)

(Part 21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of In Vitro Diagnostic Devices (OIVD)

(Division Sign-Off)

Division of Radiological Devices

Office of *In Vitro* Diagnostic Device Evaluation and Safety

510(k) Number   K102388



**Diagnostic Ultrasound Indications for Use Form**  
**Vivid i and Vivid q Diagnostic Ultrasound System**


Intended Use: Ultrasound imaging, measurement and analysis of the human body as follows:

Clinical Application Anatomy/Region of Interest	Mode of Operation										
	B	M	PW Doppler	CW Doppler	Color Doppler	Color M Doppler	Power Doppler	Combined Modes*	Harmonic Imaging	Coded Pulse <sup>t</sup>	Other
Ophthalmic											
Fetal/Obstetrics	P	P	P	P	P	P	P	P	P	P	
Abdominal <sup>[1]</sup>	P	P	P	P	P	P	P	P	P	P	
Pediatric	P	P	P	P	P	P	P	P	P	P	
Small Organ (specify) <sup>[2]</sup>	P	P	P		P	P	P	P	P	P	
Neonatal Cephalic	P	P	P	P	P	P	P	P	P	P	
Adult Cephalic	P	P	P	P	P	P	P	P	P	P	
Cardiac <sup>[3]</sup>	P	P	P	P	P	P	P	P	P	P	
Peripheral Vascular	P	P	P	P	P	P	P	P	P	P	
Musculo-skeletal Conventional	P	P	P		P	P	P	P	P	P	
Musculo-skeletal Superficial	P	P	P		P	P	P	P	P	P	
Other <sup>[4]</sup>	P	P	P		P	P	P	P	P	P	
Exam Type, Means of Access											
Transesophageal	P	P	P	P	P	P		P	P	N	
Transrectal											
Transvaginal	P	P	P		P	N	P	P	P	N	
Transurethral											
Intraoperative (specify) <sup>[5]</sup>	P	P	P	P	P		P	P	P		
Intraoperative Neurological											
Intracardiac and Intraluminal	P	P	P	P	P	P		P	P		
Laparoscopic											

N = new indication; P = previously cleared by FDA; E = added under Appendix E

Notes:

- [1] Abdominal includes GYN/Pelvic and Renal.
- [2] Small organ includes breast, testes, thyroid.
- [3] Cardiac is Adult and Pediatric.
- [4] Other use includes Urology.
- [5] Intraoperative includes abdominal, thoracic (cardiac), and vascular (PV).
- [\*] Combined modes are B/M, B/PWD, B/CWD, B/Color/PWD, B/Amplitude/PWD.
- [t] Coded Pulse includes Coded Octave Imaging (COI), and Coded Phase Inversion (CPI).

  
(Division Sign-Off)  
Division of Radiological Devices  
Office of In Vitro Diagnostic Device Evaluation and Safety  
510K K102388

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of In Vitro Diagnostic Devices (OIVD)  
Prescription User (Per 21 CFR 801.109)



**Diagnostic Ultrasound Indications for Use Form**

**Vivid i/q with 9L-RS Transducer**

Intended Use: Ultrasound imaging, measurement and analysis of the human body as follows:

Clinical Application Anatomy/Region of Interest	Mode of Operation										
	B	M	PW Doppler	CW Doppler	Color Doppler	Color M Doppler	Power Doppler	Combined Modes*	Harmonic Imaging	Coded Pulse†	Other
Ophthalmic											
Fetal / Obstetrics											
Abdominal <sup>[1]</sup>	P	P	P		P	P	P	P	P	P	
Pediatric	P	P	P		P	P	P	P	P	P	
Small Organ <sup>[2]</sup>	P	P	P		P	P	P	P	P	P	
Neonatal Cephalic											
Adult Cephalic											
Cardiac											
Peripheral Vascular	P	P	P		P	P	P	P	P	P	
Musculo-skeletal Conventional	P	P	P		P	P	P	P	P	P	
Musculo-skeletal Superficial	P	P	P		P	P	P	P	P	P	
Other (specify)											
Exam Type, Means of Access											
Transesophageal											
Transrectal											
Transvaginal											
Transurethral											
Intraoperative (specify)											
Intraoperative Neurological											
Intravascular											
Laparoscopic											

N = new indication

P = previously cleared by FDA; (Transducer previously cleared on Vivid S5/S6 (K092079));

E = added under Appendix E

Notes: [1] Abdominal includes GYN/Pelvic and Renal.

[2] Small organ includes breast, testes, thyroid.

[\*] Combined modes are B/M, B/PWD, B/Color/PWD, B/Power/PWD, B/Color M.

[†] Coded Pulse is for digitally encoded harmonics.

  
 (Division Sign-Off)  
 Division of Radiological Devices  
 Office of In Vitro Diagnostic Device Evaluation and Safety  
 510K K102388

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of In Vitro Diagnostic Devices (OIVD)

Prescription User (Per 21 CFR 801.109)



**Diagnostic Ultrasound Indications for Use Form**  
**Vivid i/q with E8C-RS Transducer**

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application Anatomy/Region of Interest	Mode of Operation										
	B	M	PW Doppler	CW Doppler	Color Doppler	Color M Doppler	Power Doppler	Combined Modes*	Harmonic Imaging	Coded Pulse <sup>T</sup>	Other
Ophthalmic											
Fetal / Obstetrics	P	P	P		P*	N	P	P	P	P*	
Abdominal <sup>[1]</sup>	P	P	P		P*	N	P	P	P	P*	
Pediatric											
Small Organ (specify)											
Neonatal Cephalic											
Adult Cephalic											
Cardiac											
Peripheral Vascular											
Musculo-skeletal Conventional											
Musculo-skeletal Superficial											
Other <sup>[4]</sup>	P	P	P		P	P*	P	P	P	P*	
Exam Type, Means of Access											
Transesophageal											
Transrectal											
Transvaginal	P	P	P		P	P*	P	P	P	P*	
Transurethral											
Intraoperative (specify)											
Intraoperative Neurological											
Intravascular											
Laparoscopic											

N = new indication

P = previously cleared by FDA on Vivid i/q (K033139) P\* = previously cleared by FDA on Vivid S5/S6 (K092079);


E = added under Appendix E

Notes: [1] Abdominal includes GYN/Pelvic and Renal.

[4] Other use includes Urology.

[\*] Combined modes are B/M, B/PWD, B/Color/PWD, B/Power/PWD, B/Color M.

[T] Coded Pulse is for digitally encoded harmonics.

  
 (Division Sign-Off)  
 Division of Radiological Devices  
 Office of In Vitro Diagnostic Device Evaluation and :  
 510K K102388

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of In Vitro Diagnostic Devices (OIVD)

Prescription User (Per 21 CFR 801.109)



**Diagnostic Ultrasound Indications for Use Form**

**Vivid i/q with 6Tc-RS Transducer**

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application Anatomy/Region of Interest	Mode of Operation										
	B	M	PW Doppler	CW Doppler	Color Doppler	Color M Doppler	Power Doppler	Combined Modes*	Harmonic Imaging	Coded Pulse†	Other
Ophthalmic											
Fetal / Obstetrics											
Abdominal											
Pediatric											
Small Organ (specify)											
Neonatal Cephalic											
Adult Cephalic											
Cardiac <sup>[3]</sup>	P	P	P	P	P	P	P	P	P	P*	
Peripheral Vascular											
Musculo-skeletal Conventional											
Musculo-skeletal Superficial											
Other											
Exam Type, Means of Access											
Transesophageal	P	P	P	P	P	P	P	P	P	P*	
Transrectal											
Transvaginal											
Transurethral											
Intraoperative (specify)											
Intraoperative Neurological											
Intravascular											
Laparoscopic											

N = new indication;

P = previously cleared by FDA, P\* = previously cleared by FDA on Vivid S5/S6 (K092079)


E = added under Appendix E

Notes:

[3] Cardiac is Adult and Pediatric.

[\*] Combined modes are B/M, B/PWD, B/Color/PWD, B/Power/PWD, B/Color M.

[†] Coded Pulse is for digitally encoded harmonics.

  
 (Division Sign-Off)  
 Division of Radiological Devices  
 Office of In Vitro Diagnostic Device Evaluation and Safety  
 510K K102388

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of In Vitro Diagnostic Devices (OIVD)

Prescription User (Per 21 CFR 801.109)



**Diagnostic Ultrasound Indications for Use Form**

**Vivid i/q -with 6T-RS Transducer**

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application <i>Anatomy/Region of Interest</i>	Mode of Operation										
	B	M	PW Doppler	CW Doppler	Color Doppler	Color M Doppler	Power Doppler	Combined Modes*	Harmonic Imaging	Coded Pulse	Other
Ophthalmic											
Fetal / Obstetrics											
Abdominal											
Pediatric											
Small Organ (specify)											
Neonatal Cephalic											
Adult Cephalic											
Cardiac <sup>[3]</sup>	P	P	P	P	P	P	P	P	P	P*	
Peripheral Vascular											
Musculo-skeletal Conventional											
Musculo-skeletal Superficial											
Other <sup>[4]</sup>											
<i>Exam Type, Means of Access</i>											
Transesophageal	P	P	P	P	P	P	P	P	P	P*	
Transrectal											
Transvaginal											
Transurethral											
Intraoperative (specify) <sup>[5]</sup>											
Intraoperative Neurological											
Intravascular											
Laparoscopic											

N = new indication;

P = previously cleared by FDA, P\* = previously cleared by FDA on Vivid S5/S6 (K092079);

E = added under Appendix E

Notes:

[3] Cardiac is Adult and Pediatric.

[\*] Combined modes are B/M, B/PWD, B/Color/PWD, B/Power/PWD, B/Color M.

[†] Coded Pulse is for digitally encoded harmonics.

  
 (Division Sign-Off)  
 Division of Radiological Devices  
 Office of In Vitro Diagnostic Device Evaluation and Safety  
 10K K102388

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of In Vitro Diagnostic Devices (OIVD)

Prescription User (Per 21 CFR 801.109)