

DEC 23 2010

**TSRH® Spinal System
510(k) Summary****October 14, 2010**

- I. Company:** Medtronic Sofamor Danek USA
1800 Pyramid Place
Memphis, Tennessee 38132
Telephone: (901) 396-3133
Fax: (901) 346-9738
- Contact:** Lila Joe
Sr. Regulatory Affairs Specialist
- II. Proposed Proprietary Trade Name:** TSRH® Spinal System
- III. Classification Name(s):** Spinal Interlaminar Fixation Orthosis, Spinal Intervertebral Body Fixation Orthosis, and Pedicle Screw Spinal System (per 21CFR Section 888.3050, 888.3060, and/or 888.3070, respectively);
Product Code(s): KWQ, KWP, MNI, MNH, NKB
- IV. Description:**
The TSRH® Spinal System is intended to help provide immobilization and stabilization of spinal segments as an adjunct to fusion of the thoracic, lumbar, and/or sacral spine.

The TSRH® Spinal System consists of a variety of shapes and sizes of rods, hooks, screws, cross connectors, staples, plates and connecting components as well as implant components from other Medtronic spinal systems, which can be rigidly locked into a variety of configurations, with each construct being tailor-made for the individual case.

Certain implant components from other Medtronic spinal systems can be used with the TSRH® Spinal System. These components include GDLH® rods, GDLH® rod/bolt connectors, GDLH® Variable Angle T-Bolts, GDLH® set screws and locking screws, DYNALOK® PLUS™ bolts, CD HORIZON® Low Profile MULTI-SPAN® CROSSLINK® Plates, VANTAGE™ Anterior Fixation System screws, and CD HORIZON® rods, set screws and locking screws.

The hooks are intended for posterior use only and the staples are for anterior use only. The TSRH-3D® and TSRH-3Dx™ connectors and TSRH-3D® and TSRH-3Dx™ screws are intended for posterior use only. ALL CROSSLINK® Plates are for

posterior use and the **CROSSLINK® Axial and Offset Plates** may be used anteriorly as well.

The **TSRH® Spinal System** implant components are fabricated from medical grade stainless steel, medical grade titanium or titanium alloy, and/or medical grade cobalt-chromium-molybdenum alloy. Medical grade titanium, titanium alloy, and/or cobalt-chromium-molybdenum alloy may be used together. Certain **TSRH® Spinal System** components may be coated with hydroxyapatite. **Never use titanium, titanium alloy, and/or cobalt-chromium-molybdenum alloy with stainless steel in the same construct.**

The purpose of this submission is to add **3Dx Chromaloy Plus (cobalt-chromium-molybdenum alloy) Pre-Cut Contoured Rods** to the **TSRH® Spinal System**.

V. Indications for Use:

When used as a pedicle screw fixation system of the non-cervical posterior spine in skeletally mature patients, the **TSRH® Spinal System** is indicated for one or more of the following: (1) degenerative disc disease (defined as back pain of discogenic origin with degeneration of the disc confirmed by patient history and radiographic studies), (2) degenerative spondylolisthesis with objective evidence of neurologic impairment, (3) fracture, (4) dislocation, (5) scoliosis, (6) kyphosis, (7) spinal tumor, and/or (8) failed previous fusion (pseudarthrosis).

In addition, when used as a pedicle screw fixation system, the **TSRH® Spinal System** is indicated for skeletally mature patients: (1) having severe spondylolisthesis (Grades 3 and 4) of the fifth lumbar-first sacral (L5-S1) vertebral joint; (2) who are receiving fusions using autogenous bone graft only; (3) who are having the device fixed or attached to the lumbar and sacral spine (L3 and below); and (4) who are having the device removed after the development of a solid fusion mass.

When used as a posterior, non-cervical, non-pedicle screw fixation system, the **TSRH® Spinal System** is intended for the following indications: (1) degenerative disc disease (as defined by back pain of discogenic origin with degeneration of the disc confirmed by patient history and radiographic studies), (2) spondylolisthesis, (3) fracture, (4) spinal deformities (i.e., scoliosis, kyphosis, and/or lordosis), (5) spinal stenosis, (6) pseudarthrosis, (7) tumor resection, and/or (8) unsuccessful previous attempts at spinal fusion.

When used as a unilateral supplemental fixation device in the antero-lateral thoracic/lumbar region, the TSRH® L-Plate and VANTAGE™ screws are intended for the following indications: spondylolisthesis; trauma (i.e., fracture or dislocation); spinal stenosis; curvatures (i.e., scoliosis, kyphosis and/or lordosis); tumor; pseudarthrosis; and/or failed previous fusion.

For anterior use only the TSRH® Spinal System has the additional indication of: spondylolysis.

VI. Summary of the Technological Characteristics

The purpose of this Special 510(k) submission is to include pre-cut, contoured 5.5mm diameter rods in lengths from 30mm to 120mm manufactured from ASTM F1537 Wrought Cobalt-28Chromium-6Molybdenum Alloy that is single annealed.

The legally marketed predicates are the TSRH® Titanium Alloy 5.5mm diameter pre-cut, contoured rods cleared by the Agency in the K010720 (S.E. 04/11/2001), the TSRH® Titanium Alloy 6.35mm diameter pre-cut, contoured rods cleared by the Agency in K022778 (S.E. 09/18/2002) and the CD HORIZON® Cobalt Chrome 5.5mm Rod cleared by the Agency into the TSRH® Spinal System in K093058 (S.E. 10/28/2009).

The table below lists the differences between the predicate devices and the subject device.

Predicate Feature TSRH® Spinal System (K010720, K022778)	Predicate Feature CD HORIZON Spinal System (K093058)	Subject TSRH® Spinal System Pre-Cut Contoured Rods
Material Titanium Alloy (ASTM F136)	Wrought Cobalt - 28Chromium - 6Molybdenum, Double Annealed (ASTM F1537)	Wrought Cobalt - 28Chromium - 6Molybdenum, Single Annealed (ASTM F1537)
Length 30mm - 100 mm	500mm	30mm - 120mm
Diameter 5.5mm, 6.35mm	5.5mm	5.5mm
Shape Contoured	Straight	Contoured
Fundamental Scientific Technology (Rod and Screw System)	Identical	Identical

VII. Identification of Legally Marketed Devices

Documentation was provided demonstrating that the TSRH® Spinal System is substantially equivalent to other commercially available fixation systems including the TSRH® Spinal System in K010720 (S.E. 04/11/2001), K022778 (S.E. 09/18/2002) and K093058 (S.E. 10/28/2009).

VI. Discussion of the Non-Clinical Testing**Subject Device: TSRH® 3Dx Chromaloy Plus Pre-Cut Contoured Rods**

Medtronic believes that documentation provided demonstrates that the TSRH® 3Dx Chromaloy Plus Pre-Cut Contoured Rods do not introduce new issues of safety or effectiveness.

The test performed per ASTM F2193-02 (2007), "Standard Specification and Test Methods for Components Used in the Surgical Fixation of the Spinal Skeletal System" (approved Nov. 2007) was Four Point Bending Fatigue.

The test performed per ASTM F1798-97 (2008), "Standard Guide for Evaluating the Static and Fatigue Properties of Interconnection Mechanisms and Subassemblies Used in Spinal Arthrodesis Implants." (approved 2008) was axial grip.



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Room W-066-0609
Silver Spring, MD 20993-0002

Medtronic Sofamor Danek USA
% Ms. Lila Joe
Senior Regulatory Affairs Specialist
1800 Pyramid Place
Memphis, Tennessee 38132

DEC 23 2010

Re: K103049

Trade/Device Name: TSRH[®] Spinal System
Regulation Number: 21 CFR 888.3070
Regulation Name: Pedicle screw spinal system
Regulatory Class: Class III
Product Code: NKB, KWP, KWQ, MNH, MNI
Dated: December 01, 2010
Received: December 02, 2010

Dear Ms. Joe:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you; however, that device labeling must be truthful and not misleading.


If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to <http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,



Mark N. Melkerson
Director
Division of Surgical, Orthopedic
And Restorative Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

510(k) Number (if known): K103049

DEC 23 2010

Device Name: TSRH® Spinal System

Indications for Use:

When used as a pedicle screw fixation system of the non-cervical posterior spine in skeletally mature patients, the TSRH® Spinal System is indicated for one or more of the following: (1) degenerative disc disease (defined as back pain of discogenic origin with degeneration of the disc confirmed by patient history and radiographic studies), (2) degenerative spondylolisthesis with objective evidence of neurologic impairment, (3) fracture, (4) dislocation, (5) scoliosis, (6) kyphosis, (7) spinal tumor, and/or (8) failed previous fusion (pseudarthrosis).

In addition, when used as a pedicle screw fixation system, the TSRH® Spinal System is indicated for skeletally mature patients: (1) having severe spondylolisthesis (Grades 3 and 4) of the fifth lumbar-first sacral (L5-S1) vertebral joint; (2) who are receiving fusions using autogenous bone graft only; (3) who are having the device fixed or attached to the lumbar and sacral spine (L3 and below); and (4) who are having the device removed after the development of a solid fusion mass.

When used as a posterior, non-cervical, non-pedicle screw fixation system, the TSRH® Spinal System is intended for the following indications: (1) degenerative disc disease (as defined by back pain of discogenic origin with degeneration of the disc confirmed by patient history and radiographic studies), (2) spondylolisthesis, (3) fracture, (4) spinal deformities (i.e., scoliosis, kyphosis, and/or lordosis), (5) spinal stenosis, (6) pseudarthrosis, (7) tumor resection, and/or (8) unsuccessful previous attempts at spinal fusion.


When used as a unilateral supplemental fixation device in the antero-lateral thoracic/lumbar region, the TSRH® L-Plate and VANTAGE™ screws are intended for the following indications: spondylolisthesis; trauma (i.e., fracture or dislocation); spinal stenosis; curvatures (i.e., scoliosis, kyphosis and/or lordosis); tumor; pseudarthrosis; and/or failed previous fusion.

For anterior use only the TSRH® Spinal System has the additional indication of: spondylolysis.

Prescription Use X OR Over-The-Counter Use
Per 21 CFR 801.109

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)



(Division Sign-Off)
Division of Surgical, Orthopedic,
and Restorative Devices

510(k) Number K103049