



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Room W-066-0609
Silver Spring, MD 20993-0002

David Chadwick, Ph.D.
Regulatory Scientist
Cook Incorporated
750 Daniels Way PO Box 489
Bloomington, IN 47404

JAN 14 2011

Re: P100028
Formula™ Balloon-Expandable Renal Stent System (see page 5 for model numbers)
Filed: July 19, 2010
Amended: October 19, 2010
Procode: NIN

Dear Dr. Chadwick:

The Center for Devices and Radiological Health (CDRH) of the Food and Drug Administration (FDA) has completed its review of your premarket approval application (PMA) for the Formula™ Balloon-Expandable Renal Stent System. This device is indicated for use in patients with atherosclerotic disease of the renal arteries following sub-optimal percutaneous transluminal renal angioplasty (PTRA) of a de novo or restenotic lesion (≤ 18 mm in length) located within 10 mm of the renal ostium and with a reference vessel diameter of 4.0 – 7.0 mm. Suboptimal PTRA is defined as $\geq 50\%$ residual stenosis, ≥ 20 mmHg systolic or ≥ 10 mmHg mean translesional pressure gradient, or flow-limiting dissection. We are pleased to inform you that the PMA is approved. You may begin commercial distribution of the device in accordance with the conditions of approval described below.

The sale and distribution of this device are restricted to prescription use in accordance with 21 CFR 801.109 and under section 515(d)(1)(B)(ii) of the Federal Food, Drug, and Cosmetic Act (the act). The device is further restricted under section 515(d)(1)(B)(ii) of the act insofar as the labeling must specify the specific training or experience practitioners need in order to use the device. FDA has determined that these restrictions on sale and distribution are necessary to provide reasonable assurance of the safety and effectiveness of the device. Your device is therefore a restricted device subject to the requirements in sections 502(q) and (r) of the act, in addition to the many other FDA

subject to the requirements in sections 502(q) and (r) of the act, in addition to the many other FDA requirements governing the manufacture, distribution, and marketing of devices.

Expiration dating for this device has been established and approved at one year for the 414 stent systems and three years for the 418 stent systems. This is to advise you that the protocol you used to establish this expiration dating is considered an approved protocol for the purpose of extending the expiration dating as provided by 21 CFR 814.39(a)(7).

Continued approval of this PMA is contingent upon the submission of periodic reports, required under 21 CFR 814.84, at intervals of one year (unless otherwise specified) from the date of approval of the original PMA. Two copies of this report, identified as "Annual Report" and bearing the applicable PMA reference number, should be submitted to the address below. The Annual Report should indicate the beginning and ending date of the period covered by the report and should include the information required by 21 CFR 814.84.

In addition to the above, and in order to provide continued reasonable assurance of the safety and effectiveness of the device, the Annual Report must include, separately for each model number (if applicable), the number of devices sold and distributed during the reporting period, including those distributed to distributors. The distribution data will serve as a denominator and provide necessary context for FDA to ascertain the frequency and prevalence of adverse events, as FDA evaluates the continued safety and effectiveness of the device.

In addition to the Annual Report requirements, you have agreed to provide the following data in post-approval study reports (PAS). Two copies, identified as "PMA Post-Approval Study Report" and bearing the applicable PMA reference number, should be submitted to the address below.

You have agreed to follow the existing 100-patient REFORM study cohort for a total of three years to assess the long-term safety and effectiveness of the device. The 1-year and 18-month visits will be phone call contact, and the 2-year and 3-year visits will be office (clinical) visits. Clinical visits will include assessment of blood pressure, blood chemistry, and antihypertensive medication usage. Adverse events will continue to be recorded and monitored through the 3-year follow-up. Upon completion of the study, you will submit a PMA Supplement to update the product labeling with the study results. Post-approval study reports should be submitted at intervals of 6 months for the first two years and yearly thereafter.

Be advised that the failure to conduct any such study in compliance with the good clinical laboratory practices in 21 CFR part 58 (if a non-clinical study subject to part 58) or the institutional review board regulations in 21 CFR part 56 and the informed consent regulations in 21 CFR part 50 (if a clinical study involving human subjects) may be grounds for FDA withdrawal of approval of the PMA.

Before making any change affecting the safety or effectiveness of the device, you must submit a PMA supplement or an alternate submission (30-day notice) in accordance with 21 CFR 814.39. All PMA supplements and alternate submissions (30-day notice) must comply with the applicable

requirements in 21 CFR 814.39. For more information, please refer to the FDA guidance document entitled, "Modifications to Devices Subject to Premarket Approval (PMA) - The PMA Supplement Decision-Making Process"

(www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/ucm089274.htm).

You are reminded that many FDA requirements govern the manufacture, distribution, and marketing of devices. For example, in accordance with the Medical Device Reporting (MDR) regulation, 21 CFR 803.50 and 21 CFR 803.52, you are required to report adverse events for this device. Manufacturers of medical devices, including in vitro diagnostic devices, are required to report to FDA no later than 30 calendar days after the day they receive or otherwise becomes aware of information, from any source, that reasonably suggests that one of their marketed devices:

1. May have caused or contributed to a death or serious injury; or
2. Has malfunctioned and such device or similar device marketed by the manufacturer would be likely to cause or contribute to a death or serious injury if the malfunction were to recur.

Additional information on MDR, including how, when, and where to report, is available at www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm.

In accordance with the recall requirements specified in 21 CFR 806.10, you are required to submit a written report to FDA of any correction or removal of this device initiated by you to: (1) reduce a risk to health posed by the device; or (2) remedy a violation of the act caused by the device which may present a risk to health, with certain exceptions specified in 21 CFR 806.10(a)(2). Additional information on recalls is available at www.fda.gov/Safety/Recalls/IndustryGuidance/default.htm.

CDRH does not evaluate information related to contract liability warranties. We remind you; however, that device labeling must be truthful and not misleading. CDRH will notify the public of its decision to approve your PMA by making available, among other information, a summary of the safety and effectiveness data upon which the approval is based. The information can be found on the FDA CDRH Internet HomePage located at www.fda.gov/MedicalDevices/ProductsandMedicalProcedures/DeviceApprovalsandClearances/PMAApprovals/default.htm. Written requests for this information can also be made to the Food and Drug Administration, Dockets Management Branch, (HFA-305), 5630 Fishers Lane, Rm. 1061, Rockville, MD 20852. The written request should include the PMA number or docket number. Within 30 days from the date that this information is placed on the Internet, any interested person may seek review of this decision by submitting a petition for review under section 515(g) of the act and requesting either a hearing or review by an independent advisory committee. FDA may, for good cause, extend this 30-day filing period.

Failure to comply with any post-approval requirement constitutes a ground for withdrawal of approval of a PMA. The introduction or delivery for introduction into interstate commerce of a device that is not in compliance with its conditions of approval is a violation of law.

You are reminded that, as soon as possible and before commercial distribution of your device, you must submit an amendment to this PMA submission with copies of all approved labeling in final printed form. Final printed labeling that is identical to the labeling approved in draft form will not routinely be reviewed by FDA staff when accompanied by a cover letter stating that the final printed labeling is identical to the labeling approved in draft form. If the final printed labeling is not identical, any changes from the final draft labeling should be highlighted and explained in the amendment.

All required documents should be submitted in triplicate, unless otherwise specified, to the address below and should reference the above PMA number to facilitate processing. One of those three copies may be an electronic copy (eCopy), in an electronic format that FDA can process, review and archive (general information:

<http://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/HowtoMarketYourDevice/Pre-marketSubmissions/ucm134508.htm>; clinical and statistical data:

<http://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/HowtoMarketYourDevice/Pre-marketSubmissions/ucm136377.htm>)

U.S. Food and Drug Administration
Center for Devices and Radiological Health
PMA Document Mail Center – WO66-G609
10903 New Hampshire Avenue
Silver Spring, MD 20993-0002

If you have any questions concerning this approval order, please contact Todd Courtney at 301-796-6371.

Sincerely yours,



 Bram Zuckerman, M.D.
Director
Division of Cardiovascular Devices
Office of Device Evaluation
Center for Devices and Radiological Health

Model	Model
FOR418-18-80-4-12	FOR418-18-135-7-12
FOR418-18-80-4-16	FOR418-18-135-7-16
FOR418-18-80-4-20	FOR418-18-135-7-20
FOR418-18-80-5-12	FORX414-14-80-5-12
FOR418-18-80-5-16	FORX414-14-80-5-16
FOR418-18-80-5-20	FORX414-14-80-5-20
FOR418-18-80-6-12	FORX414-14-80-6-12
FOR418-18-80-6-16	FORX414-14-80-6-16
FOR418-18-80-6-20	FORX414-14-80-6-20
FOR418-18-80-7-12	FORX414-14-80-7-12
FOR418-18-80-7-16	FORX414-14-80-7-16
FOR418-18-80-7-20	FORX414-14-80-7-20
FOR418-18-135-4-12	FORX414-14-135-5-12
FOR418-18-135-4-16	FORX414-14-135-5-16
FOR418-18-135-4-20	FORX414-14-135-5-20
FOR418-18-135-5-12	FORX414-14-135-6-12
FOR418-18-135-5-16	FORX414-14-135-6-16
FOR418-18-135-5-20	FORX414-14-135-6-20
FOR418-18-135-6-12	FORX414-14-135-7-12
FOR418-18-135-6-16	FORX414-14-135-7-16
FOR418-18-135-6-20	FORX414-14-135-7-20