

510(k) Summary

K110028
MAR 17 2011

Submitted by: Masimo Corporation
40 Parker, Irvine, CA 92618
Phone: (949) 297-7000; FAX: (949) 297-7592

Contact Shelly Harris, Manager of Regulatory Affairs

Date Summary Prepared January 3, 2010.

Establishment Registration No. 2031172

Trade Name Masimo Radical 7 Pulse CO-Oximeter and Accessories

Common Name Pulse Oximeter and Sensor

Regulation Number/Name/ Class 21 CFR 870.2700/ Oximeter/ Class II

Product Code DQA, BZQ, DPZ, JKS

Substantially Equivalent Device Masimo Rainbow SET[®] Radical 7R Pulse CO-Oximeters and Accessories, 510(k) Number – K100428

Description of the Device

The Radical 7 Pulse CO-Oximeter and accessories (Radical 7) include the Masimo Rainbow SET technology. The Radical 7 Pulse CO-Oximeter provides noninvasive monitoring of arterial oxygen saturation (SpO₂), pulse rate (PR), carboxyhemoglobin saturation (SpCO), methemoglobin saturation (SpMet), total hemoglobin concentration (g/dl SpHb), and/or respiration rate (RRa). Other information displayed by the Radical 7 Pulse CO-Oximeter includes: Low Signal IQ (Low SIQ), Perfusion Index (PI), Pleth Variability Index (PVI), Total Arterial Oxygen Content (SpOC), Hematocrit (SpHct), Signal Identification Quality (SIQa), Respiration Indicator (RI), alarm status, alarm silence, battery life, sensor status, trends, and pleth waveform. The Radical 7 Pulse CO-Oximeter also has a touchscreen and output interfaces, which include: SatShare connection to multi-parameter monitors, Nurse Call analog output, and RS-232 serial output, and wireless radio.

Intended Use/Indications for Use

The Masimo Radical 7 Pulse CO-Oximeter and accessories are indicated for the continuous non-invasive monitoring of functional oxygen saturation of arterial hemoglobin (SpO₂), pulse rate, carboxyhemoglobin saturation (SpCO), methemoglobin saturation (SpMet), total hemoglobin concentration (SpHb), and/or respiratory rate (RRa). The Masimo Radical 7 Pulse CO-Oximeter and accessories are indicated for use with adult, pediatric, and neonatal patients during both no motion and motion conditions, and for patients who are well or poorly perfused in hospitals, hospital-type facilities, mobile, and home environments. In addition, the Masimo Radical 7 Pulse CO-Oximeter and accessories are indicated to provide the continuous non-invasive monitoring data obtained from the Masimo Radical 7 Pulse CO-Oximeter and accessories of functional oxygen saturation of arterial hemoglobin (SpO₂) and pulse rate to multi-parameter devices for the display of those devices.

Principles of Operation

SpO₂ and Pulse Rate: Pulse oximetry is governed by the principles that blood contents differ in their absorption of red and infrared light and blood pulsation.

SpCO, SpMet, and SpHb: The Radical 7 includes the Masimo Rainbow SET technology board, which calculates functional oxygen saturation (SpO₂), fractional concentration of carboxyhemoglobin (SpCO), fractional concentration of methemoglobin (SpMet), total hemoglobin concentration (SpHb) and pulse rate.

Respiratory or Respiration Rate (RRa) General Description: The Masimo Rainbow SET technology also provides respiratory or respiration rate measurements, based on vibratory signals from respiratory sounds.

510(k) Summary

Comparison to Predicate Device

The Radical 7 in this filing is the same as the predicate in the K100428 filing, with the addition of wireless radio for remote monitoring, a Bluetooth radio module, and update to the user interface controls. Below are the specifications for the Radical 7.

FEATURES	SPECIFICATIONS
Display Range	SpO ₂ : 0-100%; PR: 25-240 bpm; SpCO: 0-99%; SpMet: 0-99.9%; SpHb: 0-25 g/dL; RRa: 0-70 breaths per minute; SpOC: 0-35 ml/dl, SpHct: 0-75%; PI: 0.02-20%; PVI: 0-100%
Accuracy: SpO ₂	No Motion (adults/pediatrics/infants): 60-80 ± 3%; 70-100 ± 2%, ± 3%(neonates) Motion (adults/pediatrics/infants/neonates): 70-100 ± 3% Low Perfusion (adults/pediatrics/infants/neonates): 70-100 ± 2%
Accuracy: PR	No Motion/ Low Perfusion(adults/pediatrics/ infants/neonates): 25-240 ± 3 bpm Motion (adults/pediatrics/infants/neonates): 25-240 ± 5 bpm
Accuracy: SpCO	1-40 + 3%, adults/pediatrics/infants
Accuracy: SpMet	1-15 + 1%, adults/pediatrics/infants/neonates
Accuracy: SpHb	8-17 +1 g/dl (arterial or venous), adults/pediatrics
Accuracy: RRa	4-70 + 1 breath per minute, adults
Resolution	SpO ₂ : 1%; PR: 1 bpm; SpCO: 1%; SpMet: 0.1%; SpHb: 0.1 g/dl; RRa: 1 breath per minute
Measurements	Low Signal IQ, PI, SpOC, SpHct, PVI, SIQa, RI
Power	AC Input Range: 100-240 VAC, 47-63 Hz; Rechargeable batteries
Environmental	Op temp: 32 to 122°F; Storage temp: -40 to 158°F; RH: 10-95% non-condensing
SpO ₂ Averaging Mode/ Sensitivity	Averaging: 2, 4, 6, 8, 10, 12 and 16 seconds; FastSat Sensitivity: APOD, Normal, Maximum
Alarm	Volume (pulse/tone), out of limit (high/low), sensor condition, system failure, low battery
Display/ Indicators	SpO ₂ , PR, SpCO, SpMet, SpHb, SpHbv, RRa, SpOC, SpHct, PI, PVI, Pleth waveform, SIQ, SIQa, RI, sensitivity, sensor status/time/messages, alarm status, battery status
Output Interface	Trends, Satshare to Multiparameter monitors (SpO ₂ only), RS-232 (PC/printer), Vuelink, Spacelabs Flexport, RadNet, PSN, wireless radio
Compliance	EMC: EN 60601-1-2, Class B; Electrical Safety: IEC 60601-1/UL 60601-1, Class 1 (AC Power), internally powered, patient cable-Type BF applied part, Satshare cable-type CF applied part; 802.11 a/b/g
Operation Mode	Continuous

Clinical Summary

Clinical Studies: SpO₂, SpCO and SpMet accuracy was determined by testing healthy adult volunteers in the range of 60-100% SpO₂, 0-40% SpCO, and 0-15% SpMet against a laboratory CO-Oximeter. SpHb accuracy has been validated on healthy adult volunteers and on surgical patients with light to dark skin pigmentation in the range of 8-17 g/dl SpHb against a laboratory CO-oximeter. The variation in these studies equals one standard deviation which encompasses 68% of the population. SpO₂ and SpMet accuracy was determined on 16 neonatal NICU patients ranging in age from 7-135 days old and weighing between 0.5-4.25 kg. Seventy-nine (79) data samples were collected over a range of 70-100% SaO₂ and 0.5-2.5% MetHb with a resultant accuracy of 2.9% SpO₂ and 0.9% SpMet.

Clinical Results: No device-related adverse events. The clinical studies were performed in accordance with ISO 9919:2005. The studies resulted in accuracies (rms) of equal to or less than the respective accuracies as stated above in the Radical 7 specifications.

Non-Clinical Summary

The Radical 7 complies with the voluntary standards as detailed in this submission. Laboratory testing for environmental and safety was conducted to verify that the Radical 7 met all design specifications and was substantially equivalent to the predicate device. This device was not tested for biocompatibility because like the predicate device, it does not contain any patient contacting materials.

510(k) Summary

Conclusions

The information in this 510(k) submission demonstrates that the Masimo Radical 7 Pulse CO-Oximeter is substantially equivalent to the predicate device, with respect to safety, effectiveness, and performance.



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Room -WO66-G609
Silver Spring, MD 20993-0002

Ms. Shelly Harris
Manager of Regulatory Affairs
Masimo Corporation
40 Parker
Irvine, California 92618

MAR 17 2011

Re: K110028

Trade/Device Name: Radical 7 Pulse CO-Oximeter and Accessories
Regulation Number: 21 CFR 870.2700
Regulation Name: Oximeter
Regulatory Class: II
Product Code: DQA, BZQ, DPZ, JKS
Dated: February 14, 2011
Received: February 15, 2011

Dear Ms. Harris:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to

<http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,



Anthony D. Watson, B.S., M.S., M.B.A.
Director
Division of Anesthesiology, General Hospital,
Infection Control and Dental Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

Indications for Use

510(k) Number (if known): _____

Device Name: Radical 7 Pulse CO-Oximeter and Accessories

Indications For Use:

The Masimo Radical 7 Pulse CO-Oximeter and accessories are indicated for the continuous non-invasive monitoring of functional oxygen saturation of arterial hemoglobin (SpO₂), pulse rate, carboxyhemoglobin saturation (SpCO), methemoglobin saturation (SpMet), total hemoglobin concentration (SpHb), and/or respiratory rate (RRa). The Masimo Radical 7 Pulse CO-Oximeter and accessories are indicated for use with adult, pediatric, and neonatal patients during both no motion and motion conditions, and for patients who are well or poorly perfused in hospitals, hospital-type facilities, mobile, and home environments. In addition, the Masimo Radical 7 Pulse CO-Oximeter and accessories are indicated to provide the continuous noninvasive monitoring data obtained from the Masimo Radical 7 Pulse CO-Oximeter and accessories of functional oxygen saturation of arterial hemoglobin (SpO₂) and pulse rate to multi-parameter devices for the display of those devices.


Prescription Use X
(Per 21 CFR 801.109 Subpart D)

AND/OR

Over-The-Counter Use _____
(Per 21 CFR 801.109 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)


(Division Sign-Off)

Division of Anesthesiology, General Hospital
Infection Control, Dental Devices

510(k) Number: K1100281