

Endoscopy
Smith & Nephew, Inc.
150 Minuteman Road
Andover, MA 01810

978 749 1000
978 749 1599 Fax
www.smith-nephew.com

 We are **smith&nephew**

K110545
p12

JUN 24 2011

SECTION IV

510(k) SUMMARY OF SAFETY AND EFFECTIVENESS INFORMATION

As required by the Safe Medical Devices Act of 1990 and codified in 21 CFR 807.92 upon which the substantial equivalence is based.

Smith & Nephew Next Generation Fully Threaded PEEK Suture Anchor

Date Prepared: June 13, 2011

A. Submitter's Name

Smith & Nephew, Inc., Endoscopy Division
150 Minuteman Road
Andover, MA 01810

B. Company Contact

Elizabeth Lavelle
Senior Regulatory Affairs Specialist
Phone: (508) 261-3607
Fax: (508) 261-3620

C. Device Name

Trade Name: Smith & Nephew Next Generation Fully Threaded PEEK Suture Anchor
Common Name: Suture Anchor
Classification Name: Fastener, fixation, non-degradable, soft tissue

D. Predicate Devices

The Smith & Nephew Next Generation Fully Threaded PEEK Suture Anchor is substantially equivalent in intended use and fundamental scientific technology to the following legally marketed suture anchors: Smith & Nephew TWINFIX FT PK (K072785), Smith & Nephew TWINFIX AB 5.0 (K011299) and Arthrex SwiveLock (K101823).

K110545
p2/2

E. Description of Device

The Next Generation Fully Threaded PEEK Suture Anchor is manufactured from PEEK (polyetheretherketone) and is offered in diameters of 4.5mm and 5.5mm sizes. The screw-in anchor is pre-assembled onto a stainless steel inserter and pre-loaded with up to three strands of suture.

F. Intended Use

The Smith & Nephew Next Generation Fully Threaded PEEK Suture Anchor is intended for use for the reattachment of soft tissue to bone for the following indications:

Shoulder:

- Bankart lesion repairs
- Slap lesion repairs
- Capsular shift or capsulolabral reconstructions
- Acromioclavicular separation repairs
- Deltoid repairs
- Rotator cuff tear repairs
- Biceps tenodesis

Foot & Ankle:

- Hallux valgus repairs
- Medial or lateral instability repairs/reconstructions
- Achilles tendon repairs/reconstruction
- Midfoot reconstructions
- Metatarsal ligament/tendon repairs/reconstructions

Knee:

- Extra-capsular repairs:
 - Medial collateral ligament
 - Lateral collateral ligament
 - Posterior oblique ligament
- Patellar realignment and tendon repairs:
 - Vastus medialis obliquous advancement
- Iliotibial band tenodesis.

Elbow:

- Ulnar or radial collateral ligament reconstructions
- Lateral epicondylitis repair
- Biceps tendon reattachment

Hip:

- Gluteal tendon repairs
- Gluteus medius and gluteus minimus repair

G. Comparison of Technological Characteristics

The Smith & Nephew Next Generation Fully Threaded PEEK Suture Anchor is substantially equivalent in intended use, technological characteristics, and is as safe and effective as its currently marketed predicate devices, the Smith & Nephew TWINFIX FT PK (K072785), the Smith & Nephew TWINFIX AB 5.0 (K011299), and the Arthrex SwiveLock (K101823) suture anchors.

H. Summary Performance Data

The performance testing conducted demonstrates that the insertion and fixation properties of the Smith & Nephew Next Generation Fully Threaded PEEK Suture Anchor is substantially equivalent to the predicate TWINFIX AB 5.0 suture anchor, cleared via 510(k) K011299. The testing demonstrates that the differences between the new device and the predicate device do not raise any new issues of safety and efficacy.



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Room -WO66-G609
Silver Spring, MD 20993-0002

Smith & Nephew, Inc.
% Ms. Elizabeth Lavelle
Sr. Regulatory Affairs Specialist
150 Minuteman Road
Andover, Massachusetts 01810

JUN 24 2011

Re: K110545

Trade/Device Name: Smith & Nephew Next Generation Fully Threaded PEEK Suture
Anchor

Regulation Number: 21 CFR 888.3040

Regulation Name: Smooth or threaded metallic bone fixation fastener

Regulatory Class: Class II

Product Code: MBI

Dated: May 5, 2011

Received: May 6, 2011

Dear Ms. Lavelle:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you; however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Page 2 – Ms. Elizabeth Lavelle

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to <http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,



for Mark N. Melkerson
Director
Division of Surgical, Orthopedic
And Restorative Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

K110545

Indications for Use

510(k) Number (if known): _____

Device Name: Smith & Nephew Next Generation Fully Threaded PEEK Suture Anchor

Indications For Use:

The Smith & Nephew Next Generation Fully Threaded PEEK Suture Anchor is intended for use for the reattachment of soft tissue to bone for the following indications:

Shoulder:

- Bankart lesion repairs
- Slap lesion repairs
- Capsular shift or capsulolabral reconstructions
- Acromioclavicular separation repairs
- Deltoid repairs
- Rotator cuff tear repairs
- Biceps tenodesis

Knee:

- Extra-capsular repairs:
 - Medial collateral ligament
 - Lateral collateral ligament
 - Posterior oblique ligament
- Patellar realignment and tendon repairs:
 - Vastus medialis obliquous advancement
- Iliotibial band tenodesis.

Foot & Ankle:

- Hallux valgus repairs
- Medial or lateral instability repairs/reconstructions
- Achilles tendon repairs/reconstruction
- Midfoot reconstructions
- Metatarsal ligament/tendon repairs/reconstructions

Elbow:

- Ulnar or radial collateral ligament reconstructions
- Lateral epicondylitis repair
- Biceps tendon reattachment

Hip: Gluteal tendon repairs

- Gluteus medius and gluteus minimus repair


Prescription Use x AND/OR Over-The-Counter Use _____

(Per 21 CFR 801 Subpart D)

(21 CFR 807 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE – CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)


 (Division Sign Off)
 Division of Surgical, Orthopedic,
 and Restorative Devices

510(k) Number K110545