

DEC 19 2011

K112056
1/2

510(k) Summary

Company Ethicon Endo-Surgery, LLC
475 Calle C
Guaynabo, PR 00969

Contact Asifa Vonhof, RAC
Regulatory Affairs Associate II
Ethicon Endo-Surgery, Inc.
Telephone: (513) 337-3118
Fax: (513) 337-2314
Email: avonhof@its.jnj.com

Date Prepared July 18, 2011

Device Name Trade Name: Echelon Endoscopic Linear Cutter Reload, Black
Common or Usual Name: Endoscopic and Accessory
Classification Name: Staple, Implantable

Predicate Device Echelon Endoscopic Linear Cutters Reload, Green
(cleared under K081146, K070887, K051002)

Device Description

The Echelon Endoscopic Linear Cutter Reload, Black is a thick tissue reload for use with the Echelon Endoscopic Linear Cutter devices (Echelon, Echelon Flex, Echelon Powered Flex) in both 45mm and 60mm configurations. It is an addition to the current Echelon reload family and will provide a closed staple height of 2.3 mm for use in thick tissue such as gastric, lung and solid organ.

The Echelon Endoscopic Linear Cutter Reload, Black is loaded into an Echelon Linear Cutter and delivers staples into the tissue when the instrument is fired.

The design and performance of the new device, Echelon Endoscopic Linear Cutter Reload, Black, is based on the predicate, Echelon Endoscopic Linear Cutter Reload, Green.

Indications for Use The ECHELON families of endoscopic linear cutters are intended for transection, resection, and/or creation of anastomoses. The instruments have application in multiple open or minimally invasive general, gynecologic, urologic, thoracic, and pediatric surgical procedures. They can be used with staple line or tissue buttressing materials. The instruments may also be used for transection and resection of liver parenchyma (hepatic vasculature and biliary structures), pancreas, kidney and spleen.

Technological Characteristics

The Echelon Endoscopic Linear Cutter Reload, Black has a larger staple size than the predicate device. All design differences between the new and predicate device can be attributed to the need to accommodate the larger size staple in the new device.

Performance Data Bench and Animal testing was performed to demonstrate that the new features do not affect safety and effectiveness and that the device will perform as intended.



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Room -WO66-G609
Silver Spring, MD 20993-0002

DEC 19 2011

Ethicon Endo-Surgery, LLC
% Ethicon Endo-Surgery, Inc.
Ms. Asifa Vonhof
4545 Creek Road
Cincinnati, Ohio 45242

Re: K112056
Trade/Device Name: Echelon Endoscopic Linear Cutter Reload, Black
Regulation Number: 21 CFR 878.4750
Regulation Name: Implantable staple
Regulatory Class: Class II
Product Code: GDW
Dated: December 07, 2011
Received: December 08, 2011

Dear Ms. Vonhof:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you; however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.


Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act

or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to <http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,



for Mark N. Melkerson

Director

Division of Surgical, Orthopedic
and Restorative Devices

Office of Device Evaluation

Center for Devices and

Radiological Health

Enclosure

Indications for Use

510(k) Number (if known): _____

Device Name: Echelon Endoscopic Linear Cutter Reload, Black

Indications for Use:

The ECHELON families of endoscopic linear cutters are intended for transection, resection, and/or creation of anastomoses. The instruments have application in multiple open or minimally invasive general, gynecologic, urologic, thoracic, and pediatric surgical procedures. They can be used with staple line or tissue buttressing materials. The instruments may also be used for transection and resection of liver parenchyma (hepatic vasculature and biliary structures), pancreas, kidney and spleen.

Prescription Use X
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use _____
(21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

David Krone for MXM
(Division Sign-Off)
Division of Surgical, Orthopedic,
and Restorative Devices

510(k) Number K112056