

510(k) Summary

 We are **smith&nephew**

FEB 27 2012

Submitted by: Smith & Nephew, Inc.
Orthopaedic Division
1450 East Brooks Road
Memphis, Tennessee 38116

Date of Summary: February 21, 2012

Contact Person and Address: Shereen Myers, Senior Regulatory Affairs Specialist
T (901) 399-6325 F (901) 566-7075

Name of Device: Smith & Nephew, Inc Journey II Deep Dished Articular Inserts

Common Name: Knee prosthesis

Device Classification Name and Reference: 21 CFR 888.3560 Knee joint patellofemorotibial polymer/metal/polymer semi-constrained cemented prosthesis

Device Class: Class II

Panel Code: Orthopaedics/87

Product Code: JWH

Device Description

Subject of this Traditional Premarket Notification are the Journey II Deep Dished articular inserts. The Journey II Deep Dished articular insert is a cruciate stabilizing tibial insert that is intended to be used with the Journey II BCS (K111711) and Journey BCS (K042515) knee systems when the posterior cruciate ligament (PCL) is sacrificed and an insert with a post is not a viable option for the patient. Components of this premarket notification include:

- Cruciate substituting (deep dished) articular inserts which will initially be available in sizes 1-2, 3-4, 5-6, and 7-8 in right and left designs. Journey II Deep Dished inserts will be offered in 9-21 mm thicknesses and manufactured from cross-linked polyethylene (7.5 Mrad XLPE) material and conventional non-cross-linked Ultra-High Molecular Weight Polyethylene (UHMWPE) material.
- Cruciate substituting (deep dished) articular insert trial implants manufactured from Radel.

Technological Characteristics

A review of the mechanical data indicates that the Journey II Deep Dished Articular Insert is capable of withstanding expected *in vivo* loading without failure. The following mechanical testing of the Journey II Deep Dished inserts was performed:

- Tibiofemoral Contact Area Analysis
- Tibiofemoral Constraint Testing

A review of this testing has demonstrated that there are no new issues related to the safety and effectiveness of the subject devices. Due to the similarities to predicate devices, no additional biocompatibility or mechanical testing was required to support the safety and effectiveness of the subject devices. Clinical data was not needed to support the safety and effectiveness of the subject devices.

Intended Use

Total knee components are indicated for rheumatoid arthritis; post-traumatic arthritis, osteoarthritis, or degenerative arthritis; failed osteotomies, unicompartmental replacement, or total knee replacement. Posterior stabilized knee systems are designed for use in patients in primary and revision surgery, where the anterior and posterior cruciate ligaments are incompetent and the collateral ligaments remain intact.

The Journey II Total Knee system components are indicated for use only with cement and are single use devices.

Substantial Equivalence Information

The substantial equivalence of the Journey II Deep Dished Articular Inserts is based on its similarities in indications for use, design features, and operational principles to the predicate systems listed in the following table.

Table 1: Comparison to Substantially Equivalent Devices

Design Aspect Reviewed	Journey II Deep Dished Inserts	Journey II BCS Knee System	High Performance Knee System	VKS Knee System	Profix Conforming Plus Tibial Insert	Profix Flex Cruciate Retaining Articular Insert	Genesis II Total Knee System
510(k) Number	Subject 510(k)	K111711	K042515	K022204	K946236	K051229	K951987
Manufacturer	Smith & Nephew, Inc	Smith & Nephew, Inc	Smith & Nephew, Inc	Plus Orthopaedics	Smith & Nephew, Inc.	Smith & Nephew, Inc.	Smith & Nephew, Inc.
Similar Indications for Use	Subject device	Y	Y	N	Y	Y	Y
Dished Insert (Cruciate Substituting)	Y	N	N	Y	Y	N	Y
Similar Sterilization Method	Subject device	Y	Y	Y	Y	Y	Y
Material	XLPE and UHMWPE	Articular Inserts - XLPE and UHMWPE	Articular Inserts - UHMWPE	Articular Inserts - UHMWPE	Articular Inserts - UHMWPE	Articular Inserts - UHMWPE	Articular Inserts - UHMWPE
Similar Manufacturing Process	Subject device	Y	Y	Y	Y	Y	Y

Conclusion

As previously noted, this Traditional 510(k) Premarket Notification is being submitted to request clearance for the Journey II Deep Dished Articular Inserts. Based on the similarities to the predicate components and a review of the validation testing performed, the device is substantially equivalent to above predicate systems.

device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to <http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,



M Mark N. Melkerson
Director
Division of Surgical, Orthopedic,
and Restorative Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

**Premarket Notification
Indications for Use Statement**

510(k) Number (if known): K113482

Device Name: Journey II Deep Dished Articular Inserts

Indications for Use:


Total knee components are indicated for rheumatoid arthritis; post-traumatic arthritis, osteoarthritis, or degenerative arthritis; failed osteotomies, unicompartmental replacement, or total knee replacement. Posterior stabilized knee systems are designed for use in patients in primary and revision surgery, where the anterior and posterior cruciate ligaments are incompetent and the collateral ligaments remain intact.

The Journey II Total Knee system components are indicated for use only with cement and are single use devices.

Prescription Use	<u>X</u>	AND/OR	Over-The-Counter Use
(Part 21 CFR 801 Subpart D)			(21 CFR 807 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE – CONTINUE ON ANOTHER PAGE IF
NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)



(Division Sign-Off)
Division of Surgical, Orthopedic,
and Restorative Devices

510(k) Number K113482