

MAR 19 2012

Section 5. 510(k) Summary

K120197

1. Administrative

Device Information

Device Name: ABL90 Flex
Common Name: Blood gases and blood pH test system
Product Code: CHL (JGS, CEM; JFP, CGZ, CGA, KHP, GKR, GHS, KQI, JJY, JIX)
Registration Number: 21 CFR 862.1120
Classification: Class II
Classification Panel: Clinical Chemistry

Submitter

Company Name: Radiometer Medical ApS
ER Number: 3002807968
Address: Aakandevj 21
2700 Broenshoej
Denkmark
Phone: +45 3827 3827
Fax: +45 3827 2727

2. Description of Device Modification

The ABL90 FLEX is a portable, automated system intended for in vitro testing of samples of whole blood for the parameters pH, pO₂, pCO₂, potassium, sodium, calcium, chloride; glucose, lactate, and co-oximetry parameters (total hemoglobin, oxygen saturation, and the hemoglobin fractions FO₂Hb, FCOHb, F Methb, FHHb and FHbF). The integration with Netop™ allows remote operators to access a specific ABL90 Flex analyser and to execute a number of predefined functions.

3. Intended Use

The ABL90 FLEX is a portable, automated analyzer that measures pH, blood gases, electrolytes, glucose, lactate and oximetry in heparinised whole blood. The ABL90 FLEX is intended for use by trained technologists, nurses, physicians and therapists. It is intended for use in a laboratory environment, near patient or point-of-care setting.

4. Substantial Equivalence

The ABL90 FLEX with the integration with Netop™ is substantially equivalent in Intended Use, fundamental scientific technology, features, and characteristics to the predicate:

510(k) Number/Device Manufacturer:

K092686 ABL90 Flex, Radiometer Medical ApS

Predicate: ABL90 Flex (K092686)	
Similarities	Differences
Intended Use The ABL90 FLEX is a portable, automated analyzer that measures pH, blood gases, electrolytes, glucose, lactate and oximetry in heparinised whole blood. The ABL90 FLEX is intended for use by trained technologists, nurses, physicians and therapists. It is intended for use in a laboratory environment, near patient or point-of-care setting.	Remote access to the analyser by the Netop host/client OTS software supporting the following functions: <ul style="list-style-type: none"> - Perform calibrations, - Perform replacements, - Perform QC measurements, - Edit data in the log files, and - Approve patient results.
Blood Gas Measurement pH, pO ₂ , pCO ₂ by potentiometry	
Electrolyte Measurement cK ⁺ , cNa ⁺ , cCa ²⁺ , cCl ⁻ by potentiometry	
Metabolite Measurement cGlu, cLac by amperometry	
Oximetry Measurement ctHb, sO ₂ FO ₂ Hb, FHHb, FCOHb, FMetHb, FHbF	
Hemoglobin Measurement Spectrophotometry	
Identical Performance Characteristics	
Two-Point liquid calibration	
Menu driven touch screen	
Software operating system Microsoft XPE	
Sample Introduction Aspiration	
Dimensions (length x width x depth)	
External Power Source 230/120 V mains	

5. Performance Data

No performance characteristics are affected by the change. The performance data submitted in K092686 still apply.



10903 New Hampshire Avenue
Silver Spring, MD 20993

Radiometer Medical ApS
c/o Martin Gabler
Aakandevej 21
Denkmark DK-2700

MAR 19 2012

Re: k120197
Trade Name: ABL90 Flex
Regulation Number: 21 CFR §862.1120
Regulation Name: Blood Gases and blood pH test system
Regulatory Class: Class II
Product Codes: CHL, CEM, CGA, CGZ, GHS, GKR, JFP, JGS, JIX, KQI, JYJ, KHP
Dated: February 28, 2012
Received: March 2, 2012

Dear Mr. Gabler:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in Title 21, Code of Federal Regulations (CFR), Parts 800 to 895. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Parts 801 and 809); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); and good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820).

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of In Vitro Diagnostic Device Evaluation and Safety at (301) 796-5450. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding postmarket surveillance, please contact CDRH's Office of Surveillance and Biometric's (OSB's) Division of Postmarket Surveillance at (301) 796-5760. For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance...

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-5680 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>

Sincerely yours,



Courtney H. Lias, Ph.D.
Director
Division of Chemistry and Toxicology Devices
Office of *In Vitro* Diagnostic Device
Evaluation and Safety
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known): k120197

Device Name: ABL90 FLEX Analyzer

Indications For Use:

The ABL90 FLEX is a portable, automated analyzer that measures pH, blood gases, electrolytes, glucose, lactate and oximetry in heparinized whole blood. The ABL90 FLEX is intended for use by trained technologists, nurses, physicians and therapists. It is intended for use in a laboratory environment, near patient or point-of-care setting. These tests are only performed under a physician's order.

pH, pO₂ and pCO₂: pH, pCO₂ and pO₂ measurements are used in the diagnosis and treatment of life-threatening acid-base disturbances.

Potassium (cK⁺): potassium measurements are used to monitor electrolyte balance in the diagnosis and treatment of disease conditions characterized by low or high blood potassium levels.

Sodium (cNa⁺): sodium measurements are used in the diagnosis and treatment of aldosteronism, diabetes insipidus, adrenal hypertension, Addison's disease, dehydration, inappropriate antidiuretic secretion, or other diseases involving electrolyte imbalance.

Calcium (cCa²⁺): calcium measurements are used in the diagnosis and treatment of parathyroid disease, a variety of bone diseases, chronic renal disease and tetany.

Chloride (cCl⁻): chloride measurements are used in the diagnosis and treatment of electrolyte and metabolic disorders such as cystic fibrosis and diabetic acidosis.

Prescription Use X
(21 CFR Part 801 Subpart D)

And/Or

Over the Counter Use
(21 CFR Part 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE; CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of In Vitro Diagnostic Device Evaluation and Safety (OIVD)



Division Sign-Off
Office of In Vitro Diagnostic Device
Evaluation and Safety

510(k) k120197

Indications for Use

510(k) Number (if known): k120197

Device Name: ABL90 FLEX Analyzer

Glucose (cGlu): glucose measurements are used in the diagnosis and treatment of carbohydrate metabolism disorders including diabetes mellitus and idiopathic hypoglycemia, and of pancreatic islet cell carcinoma.

Lactate (cLac): The lactate measurements measure the concentration of lactate in plasma. Lactate measurements are used to evaluate the acid-base status and are used in the diagnosis and treatment of lactic acidosis (abnormally high acidity of the blood.)

Total Hemoglobin (ctHb): total hemoglobin measurements are used to measure the hemoglobin content of whole blood for the detection of anemia.

sO₂: oxygen saturation, more specifically the ratio between the concentration of oxyhemoglobin and oxyhemoglobin plus reduced hemoglobin.

FO₂Hb: oxyhemoglobin as a fraction of total hemoglobin.

FCO₂Hb: carboxyhemoglobin measurements are used to determine the carboxyhemoglobin content of human blood as an aid in the diagnosis of carbon monoxide poisoning.

FMetHb: methemoglobin as a fraction of total hemoglobin.

FHHb: reduced hemoglobin as a fraction of total hemoglobin.

Fraction of Fetal Hemoglobin (FHbF): FHbF indicates the amount of fetal hemoglobin. FHbF is seldom used clinically.

Prescription Use X
(21 CFR Part 801 Subpart D)

And/Or

Over the Counter Use
(21 CFR Part 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE; CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of In Vitro Diagnostic Device Evaluation and Safety (OIVD)



Division Sign-Off
Office of In Vitro Diagnostic Device
Evaluation and Safety

510(k) 120197