



June 6, 2013

Oricare, Incorporated
Mr. David Jamison
Executive Vice President
1900 AM Drive
Quakertown, PA 18951

Re: K120931

Trade/Device Name: Oricare V8800 Critical Care Ventilator
Regulation Number: 21 CFR 868.5895
Regulation Name: Continuous Ventilator
Regulatory Class: II
Product Code: CBK, CCK, BTI, BTT
Dated: May 16, 2013
Received: May 17, 2013

Dear Mr. Jamison:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to <http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

  Mary S.
Runner -S

Kwame Ulmer, M.S.
Acting Director
Division of Anesthesiology, General Hospital,
Respiratory, Infection Control and
Dental Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

Indications for Use Statement

510 (k) Number K120931
Device Name: Oricare V8800 Critical Care Ventilator

Indications for Use:

The V8800 Critical Care Ventilator is an electronically controlled, pneumatically powered machine intended to provide continuous ventilation to patients with respiratory failure or respiratory insufficiency, requiring respiratory support.

The V8800 Critical Care Ventilator may be used by trained professional health care providers under the supervision of a physician.

The V8800 Critical Care Ventilator is intended for patients ranging from pediatric to adult, and for use in a wide variety of clinical conditions. Specifically, the V8800 Critical Care Ventilator is applicable for adult and pediatric patients weighing at least 3.5 kg (7.7 lbs.), who require the following types of ventilator support: Positive Pressure Ventilation, delivered invasively (by ET or Tracheotomy tube) or non-invasively (by mask) via Assist/Control, SIMV, CPAP and other modes of ventilation.

The V8800 Critical Care Ventilator is not intended for use with neonatal pediatric patients weighing less than 3.5 kg (7.7 lbs.) whose approximate age range is from birth to 1 month of age.

The V8800 Critical Care Ventilator is intended for use in hospital and hospital-type facilities. It may be used during intra-hospital transport provided that electrical power and compressed gas are supplied.

Prescription Use X
(Part 21 CFR 801 Subpart D)


AND/OR

Over-The-Counter Use _____
(21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Anya C
Harry

 Digitally signed by Anya C. Harry
DN: c=US, o=U.S. Government, ou=HHS,
ou=FDA, ou=People, cn=Anya C. Harry,
g.9.2342.19200300.100.1.1=0011315590
Date: 2013.06.05 17:36:12 -0400

(Division Sign-Off)
Division of Anesthesiology, General Hospital
Infection Control, Dental Devices

510(k) Number: K120931