

510(k) Summary

K122146

APR 22 2013

Submission Date: 08 April 2013

Submitter: Spacelabs Healthcare
5150 220th Avenue SE
Issaquah, WA 98029

Submitter Contact: Spacelabs Healthcare
5150 220th Avenue SE
Issaquah, WA 98029
Mr. David J. Geraghty
Spacelabs Healthcare
Phone: +1 (425) 657-7200, ext 5889
Fax: +1 (425) 657-7210
Email: david.geraghty@spacelabs.com

Application Correspondent: Thomas Kroenke
Principal Consultant
Speed To Market, Inc.
PO Box 3018
Nederland, CO 80466 USA
tkroenke@speedtomarket.net
303 956 4232

Manufacturing Site: Spacelabs Healthcare
5150 220th Avenue SE
Issaquah, WA 98029

Trade Name: Spacelabs Healthcare Xhibit Central Station, Model 96102

Common Name: Central Station Remote Monitor

Classification Name: Monitor, Physiological, Patient (With Arrhythmia Detection or Alarms);

Primary Classification Regulation: 21 CFR §870.1025

Primary Product Code: MHX

Secondary Classification Regulation: 21 CFR §870.1025; 21 CFR §870.2300

Secondary Product Code: DSI; MLD; MSX

510(k) Summary

K122146

**Substantially
Equivalent Devices:**

<i>New Spacelabs Model</i>	<i>Predicate 510(k) Number</i>	<i>Predicate Manufacturer / Model</i>
Spacelabs Healthcare Xhibit Central Station, Model 96102	K050742	Philips Medical Systems Intellivue M3290A (Release G), M3155 Configuration, Intellivue Information Center

Device Description:

The Spacelabs Healthcare (Spacelabs) Xhibit Central Station, Model 96102 (Xhibit), is a new version of currently marketed Spacelabs product. Xhibit offers remote surveillance of patient data for those patients connected to a Spacelabs Healthcare Ultraview, Ultraview SL bedside monitor, or telemetry system. Xhibit provides for data communication using the TCP/IP network protocol employed in the Spacelabs Patient Care Management network of hardwired and/or telemetry monitored patients.

Xhibit is not the primary alarming device for the Ultraview or Ultraview SL telemetry system.

Intended Use:

The Spacelabs Healthcare Xhibit Central Station, Model 96102 intended use is to provide clinicians with central monitoring of adult, pediatric and neonatal patient data of patients connected to networked Spacelabs Healthcare patient monitors and telemetry transmitters. Data includes physiological waveforms and calculations, cardiac arrhythmia and ST data, and patient demographic information to monitor adequacy of treatment or to exclude causes of symptoms.

The Spacelabs Healthcare Xhibit Central Station, Model 96102 is a prescription device intended for use under the direct supervision of a licensed healthcare professional.

**Technology
Comparison:**

Xhibit employs the same technological characteristics as the predicate device.

<i>Characteristic</i>	<i>Predicate Device</i>	<i>Proposed Device</i>
<i>Number of Patients per Display</i>	16	Same
<i>Patients per Central Station</i>	Up to 16	Up to 48
<i>Traces per Display</i>	Up to 32	Same
<i>Dedicated Display for Data Review</i>	Yes	Same
<i>Prioritized Alarms</i>	Yes	Same

Summary of Performance Testing:

Software Testing

Xhibit contains MAJOR level of concern software. Software was designed and developed according to a robust software development process, and was rigorously verified and validated. Software information is provided in accordance with internal requirements and the following standards and guidance documents:

- *FDA guidance: The content of premarket submissions for software contained in medical devices, 11 May 05;*
- *FDA guidance: Off-the-shelf software use in medical devices, 09 Sep 99;*
- *FDA guidance: General principles of software validation; Final guidance for industry and FDA staff, 11 Jan 02;*
- *FDA guidance: Cybersecurity for networked medical devices containing off-the-shelf (OTS) software, 14 January 2005; and*
- *IEC 62304: 2006, Medical device software – Software life cycle processes.*

Test results indicate that the Xhibit complies with its predetermined specifications and the applicable standards and guidance documents.

Electromagnetic Compatibility Testing

Xhibit hardware was tested for performance in accordance with the following Standard:

- *IEC 60601-1-2: 2007, Medical electrical equipment – Part 1-2: General requirements for basic safety and essential performance - Collateral standard: Electromagnetic compatibility – Requirements and tests.*

Test results indicated that Xhibit complies with the Standard.

Performance Testing

Xhibit was tested for performance in accordance with internal requirements and the following standards:

- *IEC 60601-1-8: 2006, Medical electrical equipment – General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems.*
- *IEC 62366: 2007, Medical devices – Application of usability engineering to medical devices.*

Test results indicate that Xhibit complies with its predetermined specifications and the applicable standards.

510(k) Summary

K122146

Conclusion

Verification and validation activities were conducted to establish the performance and safety characteristics of Xhibit. The results of these activities demonstrate that Xhibit is safe and effective when used in accordance with its intended use and labeling.

Therefore, Xhibit is considered substantially equivalent to the predicate device.



DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center - WO66-G609
Silver Spring, MD 20993-002

April 22, 2013

Spacelabs Healthcare
c/o Mr. Thomas Kroenke
P.O. Box 3018
Nederland, CO 80466 US

Re: K122146
Xhibit Central Station, model 96102
Regulation Number: 21 CFR 870.1025
Regulation Name: Monitor, Physiological, Patient (With Arrhythmia Detection or Alarms)
Regulatory Class: Class II
Product Code: MHX, DSI, MLD, MSX
Dated: April 8, 2013
Received: July 19, 2012

Dear Mr. Kroenke:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Page 2 - Mr. Thomas Kroenke

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638 2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

Owen P. Faris -S

for

Bram D. Zuckerman, M.D.
Director
Division of Cardiovascular Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

Indications for Use

510(k) Number (if known): K K122146

Device Name: Spacelabs Healthcare Xhibit Central Station, Model 96102

Indications for Use: The Spacelabs Healthcare Xhibit Central Station, Model 96102 intended use is to provide clinicians with central monitoring of adult, pediatric and neonatal patient data of patients connected to networked Spacelabs Healthcare patient monitors and telemetry transmitters. Data includes physiological waveforms and calculations, cardiac arrhythmia and ST data, and patient demographic information to monitor adequacy of treatment or to exclude causes of symptoms.

The Spacelabs Healthcare Xhibit Central Station, Model 96102 is a prescription device intended for use under the direct supervision of a licensed healthcare professional.

Prescription Use X AND/OR Over-The-Counter Use _____
(Part 21 CFR 801 Subpart D) (21 CFR 807 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

 Owen P. Faris -S
2013.04.22
15:41:03 -04'00'