

K122456

510(k) Summary

MAR 22 2013

This 510(k) summary is being submitted in accordance with the requirements of 21 CFR Part 807.92 for the MRII Cranial Drill and accessories.

1. Company Making the Submission:

Name of Owner:	MRI Interventions, Inc.
Address:	5 Musick Irvine, CA 92618
Telephone:	949-900-6833
Fax:	949-900-6834
Contact:	Edward Waddell
E-mail:	ewaddell@mriinterventions.com

2. Device Name:

Common Name:	Manual Drill
Proprietary Name:	MRII Cranial Drill
Classification:	Class II
Regulation Number:	882.4300
Product Code:	HBG

3. Predicate Device

Integra Hand Drill, K961113

4. Intended Use Statement:

The MRII Cranial Drill and accessories is intended to provide access through the skull for ventriculostomy or other neurological procedures, such as biopsy or catheter placement, in or near an MR scanner of 3T maximum field strength. The MRII Cranial Drill and accessories are intended to be used only when the scanner is not performing a scan. The MRII Cranial Drill is intended for single use only.

5. Description of Device:

The MRII Cranial Drill is a hand held manual drill with a 3:1 gear ration. It is intended for use with the drill kit accessory kits consisting of a 2.0 or 3.2 mm drill bit, lancet, depth stop and ruler.

6. Summary of the Technological Characteristics of the Device Compared to the Predicate Device

	<b>MRII MRII Cranial Drill</b>	<b>Predicate Device: Integra Hand Drill K961113</b>
<b>Classification</b>	21 CFR 882.4300	21 CFR 882.4300
<b>Product Code</b>	HBG	HBG
<b>Intended Use</b>	The MRII Cranial Drill and accessories is intended to provide access through the skull for ventriculostomy or other neurological procedures, such as biopsy or catheter placement, in or near an MR scanner of 3T maximum field strength. The MRII Cranial Drill and accessories are intended to be used only when the scanner is not performing a scan. The MRII Cranial Drill is intended for single use only.	The drill is intended to be used with an external drainage and monitoring system in selected patients to reduce intracranial pressure (ICP), to monitor CSF, to provide temporary drainage of CSF and to monitor ICP. The drill is intended for single use only.
<b>Environment</b>	OR or MRI Suite	OR
<b>Drill Bit Included with Drill</b>	No	No
<b>Gear Ratio</b>	3:1	3.5:1
<b>Integral Bit</b>	Yes (316L SS)	Yes
<b>Drill Bit Sizes</b>	2.0mm 3.2mm	3/16" (4.7mm) 5/32" (3.97mm) 13/64" (5.31mm) 1/4" (6.35mm)
<b>Adjustable Depth Guard</b>	Yes	Yes
<b>Set Screw with Hex Wrench</b>	No	No
<b>Packaging</b>	Drill: Sterile, CSR Warp in Tyvek Peel Pouch. Kit: Sterile, inside tray with Tyvek Lid and external	Sterile, CSR Wrap in Tyvek Peel Pouch

	<b>MRII MRII Cranial Drill</b>	<b>Predicate Device: Integra Hand Drill K961113</b>
	Tyvek Pouch	

**7. Testing:**

Testing to applicable standards has been completed with acceptable outcomes.


Bench testing performed included design verification testing and comparison testing with the predicate Integra Hand Drill, with acceptable results. These tests demonstrated that the MR Cranial Drill functions as intended and is substantially equivalent to legally marketed devices.

**8. Rx or OTC:**

The MRI Cranial Drill is an Rx prescription device per 21 CFR Part 801, Subpart D.

**9. Substantial Equivalence:**

The MR Cranial Drill is as safe and effective as the predicate Integra Hand Drill. The MR Cranial Drill has the same intended uses and similar indications, technological characteristics, and principles of operation as its predicate device. The minor technological differences between the MR Cranial Drill and its predicate devices raise no new issues of safety or effectiveness. Performance data demonstrate that the MR Cranial Drill is as safe and effective as the Integra Hand Drill. Thus, the MRI Cranial Drill is substantially equivalent.

MRI Interventions, Inc.  
  
Edward Waddell  
Director of Regulatory Affairs

Date: 11/13/12



DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

March 22, 2013

Food and Drug Administration  
10903 New Hampshire Avenue  
Document Control Center – WO66-G609  
Silver Spring, MD 20993-0002

MRI Interventions, Inc.  
Mr. Edward Waddell  
5 Musick  
Irvine, CA, 92618

Re: K122456

Trade/Device Name: MRII Cranial Drill  
Regulation Number: 21 CFR 882.4300  
Regulation Name: Manual Cranial Drills, Burrs, Trephines, and their accessories  
Regulatory Class: Class II  
Product Code: HBG  
Dated: February 15, 2013  
Received: March 7, 2013

Dear Mr. Waddell:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you; however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set

Page 2 – Mr. Edward Waddell

forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to <http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

**Joyce M. Whang**

for Victor Krauthamer, Ph.D.  
Acting Director  
Division of Neurological  
and Physical Medicine Devices  
Office of Device Evaluation  
Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known): K122456

Device Name: MRII Cranial Drill

Indications For Use:

The MRII Cranial Drill and accessories is intended to provide access through the skull for ventriculostomy or other neurological procedures, such as biopsy or catheter placement, in or near an MR scanner of 3T maximum field strength. The MRI Cranial Drill and accessories are intended to be used only when the scanner is not performing a scan. The MRII Cranial Drill is intended for single use only.

Prescription Use    
 (Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use \_\_\_\_\_   
 (21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

---

Concurrence of CDRH, Office of Device Evaluation (ODE)

<p><b>Joyce M. Whang</b></p> <hr/> <p>(Division Sign Off) Division of Neurological and Physical Medicine Devices (DNPMD)</p> <p>510(k) Number <u>K122456</u></p>
--