

Orthofix Firebird Spinal Fixation System Special 510(k)

510(k) Summary**MAR 28 2013**

Submitter: Orthofix, Inc.
3451 Plano Parkway
Lewisville, TX 75056

Contact: Jacki Geren
Regulatory Affairs Specialist
Orthofix, Inc.
214.937.2100 (phone)
214.937.3322 (fax)
jackigeren@orthofix.com

Date Prepared: September 20, 2012

Prepared By: Jacki Geren
Regulatory Affairs Specialist
Orthofix Inc.

Product Code(s): NKB; MNI; MNH

Classification Name: Pedicle Screw Spinal System

Device Class: Class III Preamendment Device, CFR 888.3070 – *Pedicle screw spinal system* - *Class III Summary and Certification Required

Classification Panel: Orthopedics

Proprietary Name: Firebird Spinal Fixation System

Device Description: The Firebird Spinal Fixation System is a temporary, multiple component system comprised of a variety of non-sterile, single use components, made of titanium alloy or cobalt chrome alloy, that allow the surgeon to build a spinal implant construct. The system is attached to the vertebral body by means of screw fixation to the non-cervical spine. The Firebird Spinal Fixation System consists of an assortment of pedicle screws, set screws, lateral offsets, bone screws and screw bodies. The Firebird Spinal

Orthofix Firebird Spinal Fixation System Special 510(k)

Fixation System implants are not compatible with components or metal from any other manufacturer's system.

The Phoenix MIS Fixation System when used with the Firebird Spinal Fixation System as indicated to provide the surgeon with a minimally invasive approach for posterior spinal surgery.

The Firebird Spinal Fixation System components are used with certain components of the Orthofix Spinal Fixation System, including rods, rod connectors and cross-connectors.

Indications For Use:

The Firebird Spinal Fixation System is intended for posterior, non-cervical pedicle fixation. Pedicle screw fixation is limited to skeletally mature patients and is intended to be used as an adjunct to fusion using autograft or allograft. The device is indicated for all of the following indications:

- a) degenerative disc disease (defined as discogenic back pain with degeneration of the disc confirmed by history and radiographic studies)
- b) spondylolisthesis,
- c) trauma (i.e., fracture or dislocation),
- d) spinal stenosis,
- e) deformities or curvatures (i.e., scoliosis, kyphosis, and/or lordosis),
- f) tumor,
- g) pseudoarthrosis, and
- h) failed previous fusion

The Phoenix MIS Fixation System when used with the Firebird Spinal Fixation System is indicated to provide the surgeon with a minimally invasive approach for posterior spinal surgery.

The Firebird Spinal Fixation System components are used with certain components of the Orthofix Spinal Fixation System, including rods, rod connectors and cross-connectors.

Materials:

Titanium alloy per ASTM F136 and Cobalt-Chrome per ASTM F1537.

Predicate Devices:

The Firebird Spinal Fixation System is substantially equivalent to the previously cleared Firebird Spinal Fixation System (K122901)

Orthofix Firebird Spinal Fixation System Special 510(k)

Substantial Equivalence: The Orthofix Firebird Spinal Fixation System proposed rod connectors are substantially equivalent to the predicate devices based on axial grip and torque verification, FEA simulation, and material characterization.

Conclusion: The Firebird Spinal Fixation System were shown to be substantially equivalent to previously cleared devices with respect to intended use, design, function, materials, and performance.



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center - WO66-G609
Silver Spring, MD 20993-0002

March 28, 2013

Orthofix, Incorporated
% Ms. Jacki Geren
Regulatory Affairs Specialist
3451 Plano Parkway
Lewisville, Texas 75056

Re: K130176
Trade/Device Name: Firebird Spinal Fixation System
Regulation Number: 21 CFR 888.3070
Regulation Name: Pedicle screw spinal system
Regulatory Class: Class III
Product Code: NKB, MNI, MNH
Dated: February 28, 2013
Received: March 1, 2013

Dear Ms. Geren:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you; however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set

Page 2 – Ms. Jacki Geren

forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to <http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOices/ucm115809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

Erin D. Keith

Mark N. Melkerson
Director
Division of Orthopedic Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

