



### 510(k) Summary

In accordance with Title 21 of the Code of Federal Regulations, Part 807, and in particular 21 CFR §807.92, the following summary of information is provided:

#### A. Submitted by:

Olga Lewis  
Regulatory Affairs Specialist  
NuVasive, Incorporated  
7475 Lusk Blvd.  
San Diego, California 92121  
Telephone: (858) 909-1800

OCT 3 1 2013

Date Prepared: October 2, 2013

#### B. Device Name

Trade or Proprietary Name:	<i>NuVasive<sup>®</sup> GSB Global Spinal Balance System</i>
Common or Usual Name:	Pedicle Screw System
Classification Name:	Pedicle Screw Spinal System, Spinal Interlaminar Fixation Orthosis, Spinal Intervertebral Body Fixation Orthosis
Device Class:	Class III
Classification:	21 CFR § 888.3050, 888.3060, 888.3070
Product Code:	NKB, KWP, MNI, MNH, KWQ

#### C. Predicate Devices

The subject *GSB Global Spinal Balance System* is substantially equivalent to the predicate device, *DePuy Acromed Moss Miami Spinal System* (K030383), *Medtronic Sofamor Danek USA CD HORIZON<sup>®</sup> Spinal System* (K082236), *Interpore Cross International Synergy Spinal System* (K011437), *NuVasive SpheRx II Pedicle Screw System* (K061778) and *NuVasive Armada System* (K091502).

#### D. Device Description

The *NuVasive GSB Global Spinal Balance System* (hereto referenced as the *GSB System*) is a pedicle screw system that consists of a variety screws, hooks, rods, lock screws, transverse connectors, rod-to-rod connectors, iliac connectors and associated general instruments. Implant components are available in a variety sizes and can be rigidly locked into a variety of different configurations to suit the individual pathology and anatomical conditions of the patient.

#### E. Intended Use

When used as a pedicle screw fixation system, the *NuVasive GSB Global Spinal Balance System* is intended to provide immobilization and stabilization of spinal segments in skeletally mature patients as an adjunct to fusion in the treatment of the following acute and chronic instabilities or deformities of the posterior thoracic, lumbar, and sacral spine:

1. Degenerative disc disease (as defined by back pain of discogenic origin with degeneration of the disc confirmed by patient history and radiographic studies)



2. Degenerative spondylolisthesis with objective evidence of neurologic impairment
3. Fracture
4. Dislocation
5. Scoliosis
6. Kyphosis
7. Spinal tumor and/or
8. Failed previous fusion (pseudoarthrosis)

The *NuVasive GSB Global Spinal Balance System* is also indicated for the treatment of severe spondylolisthesis (Grades 3 and 4) of the L5-S1 vertebral joint in skeletally mature patients receiving fusion by autogenous bone graft, having the device fixed or attached to the lumbar and sacral spine (L3 to sacrum), with removal of the implants after attainment of a solid fusion.

When used as an anterolateral non-pedicle screw system in the thoracic and lumbar spine, the *NuVasive GSB Global Spinal Balance System* is also intended for the following indications:

1. Degenerative disc disease (as defined by back pain of discogenic origin with degeneration of the disc confirmed by patient history and radiographic studies)
2. Spinal stenosis
3. Spondylolisthesis
4. Spinal deformities
5. Fracture
6. Pseudoarthrosis
7. Tumor resection and/or
8. Failed previous fusion

#### **F. Technological Characteristics**

As was established in this submission, the subject *NuVasive GSB Global Balance System* is substantially equivalent to other predicate devices cleared by the FDA for commercial distribution in the United States. The subject device was shown to be substantially equivalent and have equivalent technological characteristics to its predicate device through comparison in areas including design, labeling/intended use, material composition, and function.

#### **G. Performance Data**

Nonclinical testing was performed to demonstrate that the subject *NuVasive GSB Global Balance System* is substantially equivalent to other predicate device. The following testing was performed:

- Static Compression Bending per ASTM F1717
- Dynamic Compression Bending per ASTM F1717
- Static Torsion per ASTM F1717
- Tulip pull-off

The results demonstrate that the subject *NuVasive GSB Global Balance System* is substantially equivalent to the predicate.

**H. Conclusions**

The subject *NuVasive GSB Global Balance System* has been shown to be substantially equivalent to legally marketed predicate devices in terms of safety and effectiveness, having similar indications for use, technological characteristics, and principles of operations as the predicate devices.



DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Food and Drug Administration  
10903 New Hampshire Avenue  
Document Control Center - WO66-G609  
Silver Spring, MD 20993-0002

October 31, 2013

NuVasive, Incorporated  
Ms. Olga Lewis  
Regulatory Affairs Specialist  
7475 Lusk Boulevard  
San Diego, California 92121

Re: K132014

Trade/Device Name: NuVasive® GSB Global Spinal Balance System  
Regulation Number: 21 CFR 888.3070  
Regulation Name: Pedicle screw spinal system  
Regulatory Class: Class III  
Product Code: NKB, MNH, MNI, KWP, KWQ  
Dated: October 2, 2013  
Received: October 3, 2013

Dear Ms. Lewis:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set

Page 2 – Ms. Olga Lewis

forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

**Ronald P. Jean -S** for

Mark N. Melkerson  
Director

Division of Orthopedic Devices  
Office of Device Evaluation  
Center for Devices and  
Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known): K132014

Device Name: NuVasive® GSB Global Spinal Balance System

### Indications For Use:

When used as a pedicle screw fixation system, the *NuVasive GSB Global Spinal Balance System* is intended to provide immobilization and stabilization of spinal segments in skeletally mature patients as an adjunct to fusion in the treatment of the following acute and chronic instabilities or deformities of the posterior thoracic, lumbar, and sacral spine: degenerative disc disease (as defined by back pain of discogenic origin with degeneration of the disc confirmed by patient history and radiographic studies), degenerative spondylolisthesis with objective evidence of neurologic impairment, fracture, dislocation, scoliosis, kyphosis, spinal tumor, and/or failed previous fusion (pseudoarthrosis).

The *NuVasive GSB Global Spinal Balance System* is also indicated for the treatment of severe spondylolisthesis (Grades 3 and 4) of the L5-S1 vertebral joint in skeletally mature patients receiving fusion by autogenous bone graft, having the device fixed or attached to the lumbar and sacral spine (L3 to sacrum), with removal of the implants after attainment of a solid fusion.

When used as an anterolateral non-pedicle screw system in the thoracic and lumbar spine, the *NuVasive GSB Global Spinal Balance System* is also intended for the following indications: degenerative disc disease (as defined by back pain of discogenic origin with degeneration of the disc confirmed by patient history and radiographic studies), spinal stenosis, Spondylolisthesis, spinal deformities, fracture, pseudoarthrosis, tumor resection, and/or failed previous fusion.

Prescription Use   
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use   
(21 CFR 807 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

---

Concurrence of CDRH, Office of Device Evaluation (ODE)

Ronald P. Jean -S

---

(Division Sign-Off)  
Division of Orthopedic Devices  
510(k) Number: K132014