

510(k) Summary

JAN 13 2014

Device Trade Name: Mixing Syringe System

Manufacturer: Zimmer Knee Creations, Inc.
841 Springdale Dr.
Exton, PA 19341

Contact: Mr. Shaun Hanson
Associate Director of Development
Phone: 484-753-5461
Fax: 202-552-5798

Prepared by: Musculoskeletal Clinical & Regulatory Advisers, LLC
1331 H Street NW, 12th Floor
Washington, DC 20005
Phone: (202) 552-5800

Date Prepared: December 5, 2013

Common Name: Piston Syringe

Classification: 21 CFR 880.5860

Class: II

Product Code: FMF

Indications For Use:

The Mixing Syringe System is intended to be used for the delivery of hydrated allograft, autograft, or synthetic bone graft material to an orthopedic surgical site.

Device Description:

The Mixing Syringe System is comprised of a commercially available disposable medical piston syringe (syringe barrel with female luer, plunger), and mixing system funnel. The Mixing Syringe System is provided sterile, for single use only.

Predicate Device:

The Mixing Syringe System has the same indications for use, design, function, materials, and is substantially equivalent to the Arthrex Mixing and Delivery System (K121124) and InFill graft Delivery System (K121476, K111632).

Performance Standards:

All necessary testing has been performed for the Mixing Syringe System to assure substantial equivalence to the predicate device and demonstrate the device performs as intended. All testing was performed on test units representative of finished devices.

The device design was qualified through the following tests:

- Simulated Use Testing
- Volume Verification
- Separation Force Testing
- Liquid Leak Testing
- Biocompatibility Evaluation

Conclusion:

The Mixing Syringe System met all specified criteria and did not raise new safety or effectiveness questions. The indications, intended use, and fundamental scientific technology of the Mixing Syringe System are the same as those for the predicate device. Therefore, the Mixing Syringe System is substantially equivalent to the Arthrex Mixing and Delivery System (K121124) and InFill graft Delivery System (K121476, K111632).



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center - WO66-G609
Silver Spring, MD 20993-0002

Zimmer Knee Creations, Incorporated
Mr. Shaun Hanson
Associate Director of Development
841 Springdale Drive
Exton, Pennsylvania 19341

January 13, 2014

Re: K133021

Trade/Device Name: Mixing Syringe System
Regulation Number: 21 CFR 880.5860
Regulation Name: Piston syringe
Regulatory Class: Class II
Product Code: FMF
Dated: December 6, 2013
Received: December 9, 2013

Dear Mr. Hanson:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you; however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set

Page 2 – Mr. Shaun Hanson

forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

Joshua C. Nipper -S

FOR
Binita S. Ashar, M.D., M.B.A., F.A.C.S.
Acting Director
Division of Surgical Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

4. Indications for Use

510(k) Number (if known): K133021

Device Name: Mixing Syringe System

The Mixing Syringe System is intended to be used for the delivery of hydrated allograft, autograft, or synthetic bone graft material to an orthopedic surgical site.

Prescription Use ✓
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use _____
(21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE
IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Long H.
Chen -A

Digitally signed by Long H. Chen -A
DN: cn=US, o=U.S. Government, ou=HHS,
ou=FDA, ou=People, cn=Long H. Chen -
A
0.9.2342.19200300.100.1.1+1300369056
Date: 2014.01.10 12:36:10 -0500

for BSA

(Division Sign-Off)
Division of Surgical Devices
510(k) Number: K133021