

510(k) Summary

JAN - 9 2014

Applicant:	Spineology Inc. 7800 3 rd Street N., Suite 600 Saint Paul, MN 55128 Phone: 651-256-8500 Fax: 651-256-8505
Contact Person:	Bryan Becker
Date Prepared:	October 30, 2013
Trade Name:	Rampart™-L
Product Classification and Code:	Class II Medical Device, Product Code MAX, 21 CFR 888.3080
Predicate Device(s):	Spineology PEEK Lumbar Interbody Fusion Devices/Rampart™-O (K111880, K113030, K120293, K123652, K130396, K131216, and K132053)
Device Description:	Rampart™ L devices are designed for use with autograft as an adjunct to fusion and are intended for use with supplemental fixation systems cleared for use in the lumbar spine. The device is available in a range of lengths and heights and is made of PEEK-OPTIMA LT-1 with tantalum markers. The device is provided sterile and associated instruments are provided non-sterile.
Intended Use:	Rampart™ L is an intervertebral body fusion device indicated for intervertebral body fusion at one level or two contiguous levels in the lumbar spine from L2 to S1 in patients with degenerative disc disease (DDD) with up to Grade I spondylolisthesis at the involved level(s). DDD is defined as back pain of discogenic origin with degeneration of the disc confirmed by patient history and radiographic studies. These patients should be skeletally mature and have had six months of non-operative treatment. Rampart™ L is designed for use with autograft as an adjunct to fusion and is intended for use with supplemental fixation systems cleared by the FDA for use in the lumbar spine.
Purpose of this 510(k):	The purpose of this 510(k) is to extend the range of sizes of Rampart O to create the Rampart-L devices.
Summary of Technological Characteristics:	The device is shown to be substantially equivalent to the intended use, materials, configuration, and performance characteristics of the predicate products.
Testing	Preclinical testing according to ASTM F2077 (static compression and static shear compression) and ASTM F2267 (subsidence), was used to justify substantial equivalence. This testing demonstrated similar performance characteristics to the identified predicate devices.

Conclusion:	The information submitted in this premarket notification supports a determination that Rampart™ L is substantially equivalent in technological characteristics and intended use to the predicate device.
--------------------	--



January 9, 2014

Spineology, Incorporated
Mr. Bryan Becker
Regulatory Affairs Manager and Compliance Officer
7800 Third Street North, Suite 600
Saint Paul, Minnesota 55128

Re: K133371

Trade/Device Name: Rampart™-L
Regulation Number: 21 CFR 888.3080
Regulation Name: Intervertebral body fusion device
Regulatory Class: Class II
Product Code: MAX
Dated: December 13, 2013
Received: December 16, 2013

Dear Mr. Becker:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set

Page 2 – Mr. Bryan Becker

forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

Lori A. Wiggins

for

Mark N. Melkerson

Director

Division of Orthopedic Devices

Office of Device Evaluation

Center for Devices and

Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K133371

Device Name

Rampart-L

Indications for Use (Describe)

Rampart™ L is an intervertebral body fusion device indicated for intervertebral body fusion at one level or two contiguous levels in the lumbar spine from L2 to S1 in patients with degenerative disc disease (DDD) with up to Grade I spondylolisthesis at the involved level(s). DDD is defined as back pain of discogenic origin with degeneration of the disc confirmed by patient history and radiographic studies. These patients should be skeletally mature and have had six months of non-operative treatment. Rampart™ L is designed for use with autograft as an adjunct to fusion and is intended for use with supplemental fixation systems cleared by the FDA for use in the lumbar spine.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

PLEASE DO NOT WRITE BELOW THIS LINE – CONTINUE ON A SEPARATE PAGE IF NEEDED.

FOR FDA USE ONLY

Concurrence of Center for Devices and Radiological Health (CDRH) (Signature)

Anton E. Dmitriev, PhD

Division of Orthopedic Devices

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."