

JUL 30 2014

### **510(k) Summary**

**Date:** 5 November 2013

**Sponsor:** K2M Inc.  
751 Miller Drive SE  
Leesburg, Virginia 20175  
Phone: 703.777.3155

**Contact Person:** Nancy Giezen

**Trade Names:** RANGE (DENALI / MESA) Spinal Systems

**Device Classification** Class II, Class III

**Classification Names (Regulations):** Spinal Interlaminar Fixation Orthosis (888.3050)  
Spinal intervertebral body fixation orthosis (888.3060)  
Pedicule Screw Spinal System (888.3070)

**Device Product Codes:** KWP, KWQ, MNI, MNH, NKB, OSH

**Submission Purpose:** This submission adds the K2M Modular Hooks to the Range (Denali/Mesa) Spinal System.

**Device Description:** The RANGE / DENALI / MESA Spinal Systems are top-loading, multiple component, posterior (thoracic-lumbar) spinal fixation systems which consists of pedicle screws, rods, hooks and rod connectors.  
The K2M Modular Hook is comprised of a hook body and a variety of adjustable blade options.

**Intended Use:** RANGE /DENALI/MESA and SMALL STATURE and ARI are cleared for the following indications:  
Non-cervical fixation as an adjunct to fusion for the following indications: degenerative disc disease (defined as back pain of discogenic origin with degeneration of the disc confirmed by history and radiographic studies); spondylolisthesis; trauma (i.e., fracture or dislocation); spinal stenosis; curvatures (i.e. scoliosis, kyphosis and/or lordosis); tumor; pseudarthrosis; and/or failed previous fusion.  
Except for hooks, when used as an anterolateral thoracic/lumbar system the Range Spinal System may also be used for the same indications as an adjunct to fusion.  
Except for the ARI staples, the Range Spinal System is indicated as an adjunct to fusion to treat adolescent idiopathic scoliosis when used for posterior noncervical fixation in pediatric patients. The Range Spinal System for pediatric use is intended to be used with autograft and/or allograft. Pediatric pedicle screw fixation is limited to a posterior approach.

**Materials:** The K2M Modular Hooks components are manufactured from titanium alloy (Ti-6Al-4V ELI) as described by ASTM and ISO standards.

**Predicate Devices:** K2M MESA® Spinal System (K052398/K080611)  
K2M Range (Mesa/Denali) Spinal System (K141147)  
Medtronic CD Horizon (K962708/K031655)

**Performance Data:** The worst case K2M Modular Hook was evaluated using static and dynamic compression bending tests according to ASTM F1717, and static A-P and disassembly tests according to ASTM F1798. The mechanical test results demonstrate that the K2M Modular Hook performance is substantially equivalent to the predicate MESA Spinal System performance.

**Technological Characteristics:** The K2M Modular Hook components possess the same technological characteristics as the predicate devices. These include:

- basic design (rod-based hook attachment),
- material (titanium alloy),
- anatomic location (non-cervical spine) and
- sizes (widths, lengths and heights are within the range(s) offered by the predicate systems).

Technological characteristics which are different have been supported with descriptive information and/or performance data which demonstrate the safety and effectiveness has not been diminished.

**Conclusion:** In comparison to the predicate devices, the K2M Modular Hooks has

- the same intended use (as described above),
- the same technological characteristics or different without raising safety and effectiveness issues (as described above)

Therefore the K2M Modular Hooks can be found substantially equivalent to the predicate devices.



Food and Drug Administration  
10903 New Hampshire Avenue  
Document Control Center – WO66-G609  
Silver Spring, MD 20993-0002

July 30, 2014

K2M, Incorporated  
% Karen E. Warden, Ph.D.  
BackRoads Consulting, Incorporated  
8202 Sherman Road  
Chesterland, Ohio 44026

Re: K133402

Trade/Device Name: RANGE (DENALI/MESA) Spinal Systems  
Regulation Number: 21 CFR 888.3070  
Regulation Name: Pedicle screw spinal system  
Regulatory Class: Class III  
Product Code: NKB, OSH, MNI, MNH, KWP, KWQ  
Dated: June 26, 2014  
Received: June, 27, 2014

Dear Dr. Warden:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set

Page 2 – Karen E. Warden, Ph.D.

forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

**Ronald P. Jean -S** for

Mark N. Melkerson  
Director  
Division of Orthopedic Devices  
Office of Device Evaluation  
Center for Devices and  
Radiological Health

Enclosure

**Indications for Use**

510(k) Number (if known)  
K133402

Device Name  
RANGE (DENALI/MESA) Spinal Systems

Indications for Use (Describe)

RANGE/DENALI/MESA and SMALL STATURE and ARI are cleared for the following indications:

Non-cervical fixation as an adjunct to fusion for the following indications: degenerative disc disease (defined as back pain of discogenic origin with degeneration of the disc confirmed by history and radiographic studies); spondylolisthesis; trauma (i.e., fracture or dislocation); spinal stenosis; curvatures (i.e., scoliosis, kyphosis, and/or lordosis); tumor; pseudarthrosis; and/or failed previous fusion.

Except for hooks, when used as an anterolateral thoracic/lumbar system the Range Spinal System may also be used for the same indications as an adjunct to fusion.

Except for the ARI staples, the Range Spinal System is indicated as an adjunct to fusion to treat adolescent idiopathic scoliosis when used for posterior noncervical fixation in pediatric patients. The Range Spinal System for pediatric use is intended to be used with autograft and/or allograft. Pediatric pedicle screw fixation is limited to a posterior approach.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

**PLEASE DO NOT WRITE BELOW THIS LINE – CONTINUE ON A SEPARATE PAGE IF NEEDED.**

**FOR FDA USE ONLY**

Concurrence of Center for Devices and Radiological Health (CDRH) (Signature)

**Zane W. Wyatt**

**Division of Orthopedic Devices**

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
PRAStaff@fda.hhs.gov

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*