



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center – WO66-G609
Silver Spring, MD 20993-0002

October 30, 2014

KJ Meditech Co. Ltd.
C/O Priscilla Juhee Chung
LK Consulting Group USA, Inc.
2651 East Chapman Avenue, Suite 110
Fullerton, California 92831

Re: K140052

Trade/Device Name: J2I Implant System
Regulation Number: 21 CFR 872.3640
Regulation Name: Endosseous Dental Implant
Regulatory Class: II
Product Code: DZE, NHA
Dated: September 23, 2014
Received: September 29, 2014

Dear Ms. Chung:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA).

You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

A handwritten signature in black ink that reads "Susan Runna DDS, MA". The signature is written in a cursive style. In the background, there is a faint, large watermark of the letters "FDA".

Erin I. Keith, M.S
Director
Division of Anesthesiology, General Hospital,
Respiratory, Infection Control and Dental Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

Indications for Use

510(k) Number (if known) K140052

Device Name
J2I Implant System

Indications for Use (Describe)

J2I Implant System is indicated for use in partially or fully edentulous mandibles and maxillae, in support of single or multiple-unit restorations including; cemented retained, screw retained, or overdenture restorations, and terminal or intermediate abutment support for fixed bridgework. J2I Implant System is for single and two stage surgical procedures. It is intended for delayed loading.

Type of Use (Select one or both, as applicable)

☒ Prescription Use (Part 21 CFR 801 Subpart D)

☐ Over-The-Counter Use (21 CFR 801 Subpart C)

PLEASE DO NOT WRITE BELOW THIS LINE – CONTINUE ON A SEPARATE PAGE IF NEEDED.

FOR FDA USE ONLY

Concurrence of Center for Devices and Radiological Health (CDRH) (Signature)

510(k) Summary - K140052

This summary of 510(K) – safety and effectiveness information is being submitted in accordance with requirements of 21 CFR Part 807.92.

Date: 10/29/2014

1. Applicant / Submitter

KJ Meditech Co., Ltd.
959-21 Daechon-dong, Buk-gu, Gwang-ju, 500-470, South Korea
Tel: +82-62-972-5476
Fax: +82-62-973-2809

2. US Agent / Submission Contact Person

LK Consulting Group USA, Inc.
2651 E Chapman Ave Ste 110, Fullerton, CA 92831
Priscilla Juhee Chung
Phone: 714.202.5789 Fax: 714-409-3357
Email: juhee.c@lkconsultinggroup.com

3. Device

- Trade Name: J2I Implant System
- Common Name: Dental Implant System
- Classification Name: Endosseous Dental Implant System
- Product Code: DZE, NHA
- Classification regulation: 21CFR872.3640

4. Predicate Device:

SS SYSTEM by OSSTEM IMPLANT CO., LTD (K062051)
Hero II Dental Implant System, IS Dental Implant System by KJ Meditech Co., Ltd.
(K121047)

5. Description:

The J2I Implant System is a dental implant system made of Titanium 6AL 4V ELI alloy intended to be surgically placed in the bone of the upper or lower jaw arches for loading after a conventional healing period. The implant may be used to replace one or more missing teeth. The system is similar to other commercially available products based on the intended use, the technology used, the claims, the material composition employed and performance characteristics. The surface of the system has been treated with R.B.M (Resorbable Blast Media).

The J2I Implant is offered in the following sizes.

| Material: Ti 6Al 4V ELI, ASTM F136 | | |
|------------------------------------|----------|------------|
| Platform Dia(mm) | Dia.(mm) | Length(mm) |
| 4.8 | ø4.1 | 8.0 |
| | | 10.0 |
| | | 11.5 |
| | | 13.0 |
| | ø4.3 | 8.0 |
| | | 10.0 |
| | | 11.5 |
| | | 13.0 |
| | ø4.8 | 8.0 |
| | | 10.0 |
| | | 11.5 |
| | | 13.0 |
| 6.5 | ø4.8 | 8.0 |
| | | 10.0 |
| | | 11.5 |
| | | 13.0 |

6. Indication for use:

J2I Implant System is indicated for use in partially or fully edentulous mandibles and maxillae, in support of single or multiple-unit restorations including; cemented retained, screw retained, or overdenture restorations, and terminal or intermediate abutment support for fixed bridgework. J2I Implant System is for single and two stage surgical procedures. It is intended for delayed loading.

7. Basis for Substantial Equivalence

Similarities

J2I Implant System has the same intended use as the identified predicate device (K062051). J2I Implant System and SS FIXTURE SYSTEM are similar in fundamental scientific technology in that they are all threaded, root form implants constructed of titanium with RBM roughened surfaces. They all share same internal octagon abutment connection system with internal beveled interface for maximum prosthetic stability. The subject and predicate devices are both bone-level implants that share similar body shape design such as straight walled neck and tapered body design.

Difference

The difference is in the body shape: cutting edge design and taper angle. However they are only the minor design factors, so they do not raise new questions on safety and effectiveness of the subject device.

8. Non-Clinical Testing

- Sterilization validating testing has been performed in accordance with ISO 11137, ISO 11737-1 & ISO 11737-2 for gamma sterilization and ISO 17665-1 and ISO 17665-2 for steam sterilization.
- Three year of shelf life has been validated through accelerating testing.
- Chemical and SEM image analyses have been performed to verify that there is no residual after RBM treatment on the fixtures.

9. Conclusion

The subject devices and the predicate devices have the same intended use and have the same technological characteristics. The subject and predicate implants are all made of commercially pure titanium and have the same surface treatments.

Overall, the J2I Implant System has the following similarities to the predicate devices:

- * has the same intended use,
- * uses the same operating principle,
- * incorporates the same basic design,
- * incorporates the same material and the surface treatment.

Based on the similarities, we conclude that the J2I Implant System is substantially equivalent to the predicate device.