



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center – WO66-G609
Silver Spring, MD 20993-0002

November 25, 2014

Merit Medical Systems, Inc.
Dan Lindsay, RAC
Senior Regulatory Affairs Specialist
1600 West Merit Parkway
South Jordan, UT 84095

Re: K140382

Trade/Device Name: AEROMini™ Tracheobronchial Stent System
Regulation Number: 878.3720
Regulation Name: Tracheal Prosthesis
Regulatory Class: Class II
Product Code: JCT
Dated: October 24, 2014
Received: October 27, 2014

Dear Mr. Lindsay,

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA).

You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

Tejashri Purohit-Sheth, M.D.

Tejashri Purohit-Sheth, M.D.
Clinical Deputy Director
DAGRID/ODE/CDRH FOR

Erin I. Keith, M.S.
Director
Division of Anesthesiology, General Hospital,
Respiratory, Infection Control and
Dental Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

4.0 Indications for Use

510(k) Number (if known): K140382

Device Name:

AEROMini™ Tracheobronchial Stent System

Indications for Use:

The MERIT ENDOTEK AEROMini Tracheobronchial Stent System is indicated for use in the treatment of tracheobronchial strictures produced by malignant neoplasms.

Prescription Use

AND/OR

Over-The-Counter Use

(Part 21 CFR 801 Subpart D)

(21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE—CONTINUE ON ANOTHER PAGE IF
NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

5.0 510(k) Summary

General Provisions	Correspondent Name:	Merit Medical Systems, Inc.
	Address:	1600 West Merit Parkway South Jordan, UT 84095
	Telephone Number:	801-208-4408
	Fax Number:	801-253-6945
	Contact Person:	Dan W. Lindsay
	Date of Preparation:	June 3, 2014
	Registration Number:	1721504

Subject Device	Trade Name:	AEROMini™
	Common/Usual Name:	Tracheobronchial Stent
	Classification Name:	Tracheal Prosthesis

Predicate Device	Trade Name:	AERO™ Tracheobronchial Stent Technology System
	Classification Name:	Tracheal Prosthesis
	Premarket Notification:	K082284
	Manufacturer:	Merit Medical Systems, Inc.

Classification	Class II
	21 CFR § 878.3720
	FDA Product Code: JCT
	Review Panel: General and Plastic Surgery

Intended Use	The Merit ENDOTEK AEROMini™ Tracheobronchial Stent System is indicated for use in the treatment of tracheobronchial strictures produced by malignant neoplasms.
---------------------	---

Device Description	<p>The MERIT ENDOTEK AEROMini Tracheobronchial Stent System is comprised of two components: the radiopaque self-expanding nitinol stent and the delivery system. The stent is completely covered with a biocompatible polyurethane membrane. The stent expansion results from the physical properties of the metal and the proprietary geometry. The overall stent geometry is designed to maintain a constant length over the entire range of possible diameters. As a result of this unique design the stent has virtually no foreshortening, thus facilitating the selection of the appropriate stent length. The stent is deployed endoscopically with a dedicated delivery system with or without the aid of fluoroscopic imaging. The delivery system consists of two coaxial sheaths. The exterior sheath serves to constrain the stent until the sheath is retracted during deployment. The stent remains constrained by the delivery system until the trigger is pulled beyond the white deployment threshold mark located between the trigger and hand grip. This feature allows for repositioning of the stent proximally. In addition, the procedure can be aborted and the entire system can be withdrawn en bloc at any time before the trigger is pulled beyond the white deployment threshold mark located between the trigger and hand grip. A radiopaque tip and marker on the inner shaft aid the operator in determining stent position in relation to the deployment threshold mark, where repositioning or en bloc withdrawal is no longer possible. The inner sheath of the delivery system contains a central lumen that will accommodate a 0.035" guide wire. This feature is designed to allow guidance of the delivery system to the intended implant site while minimizing the risk of airway injury from the delivery system tip.</p>
Comparison to Predicate Device	<p>The AEROMini is similar in design to the AERO Tracheobronchial Stent Technology System. Both stents are fabricated as a single, integral framework by laser-cutting a nitinol tube. A polyurethane polymer cover is then applied and a hydrophilic coating is added to the inner diameter surface of the stent. The subject and predicate devices are preloaded on a flexible one-handed delivery system.</p> <p>The AEROMini is offered in 5 new sizes with diameters as small as 8mm and lengths as short as 15mm. Other stent modifications include: changes to the stent flares, anti-migration struts, new proximal stent end suture eyelets. Changes to the delivery system include: material changes, smaller profile diameter, tapered tip, new trigger safety, and design changes intended to improve ergonomics. The AEROMini will be provided sterile while the predicate was not.</p>

FDA guidance and recognized performance standards have been established for Tracheal Prosthesis under Section 514 of the Food, Drug and Cosmetic Act. A battery of tests was performed based on the requirements of the below recognized performance standards and guidance, as well as biocompatibility, sterilization, and labeling standards and guidance. Conformity to these standards demonstrates that the proposed AEROMini met the standards' established acceptance criteria applicable to the substantial equivalence of the device. Performance testing was conducted based on the risk analysis and based on the requirements of the following documents:

The following tests were successfully conducted as per FDA *Guidance for the Content of Premarket Notifications for Esophageal and Tracheal Prosthesis*, April 28, 1998:

Deployment Testing.
Expansion Force Testing.
Compression Force Testing.
Dimensional Testing.
Tensile Strength Tests.

**Performance
Tests**

Guidance for Industry: Guidance for the Content of Premarket Notifications for Esophageal and Tracheal Prostheses (Issued April 28, 1998)
ISO 11135-1:2007, Sterilization of health care products – routine control of a sterilization process for medical devices
ASTM D4169 - 09 Standard Practice for Performance Testing of Shipping Containers and Systems
ASTM F2052-06e1:2006 Measurement of Magnetically Induced Displacement Force on Medical Devices in the MR Environment
ASTM F2119-07:2013, Standard Test Method for Evaluation of MR Image Artifacts from Passive Implants
ASTM F2182-09:2011, Standard Test Method for Measurement of Radio Frequency Induced Heating Near Passive Implants During MR Imaging
ASTM F2213-06:2011, Method for Measurement of Magnetically Induced Torque on Medical Devices in MR Environment
ASTM F2503-08:2013 Marking Medical Devices and Other Items for Safety in the Magnetic Resonance Environment
ISO 10993-1:2009, Biological Evaluation of Medical Devices - Part 1: Evaluation and Testing within a risk management process, and FDA guidance Required Biocompatibility Training and Toxicology Profiles for Evaluation of Medical Devices, May 1, 1995
ISO 10993-10:2010, Biological evaluation of medical devices – Part 10: Tests for irritation and skin sensitization
ISO 10993-17:2002, Biological evaluation of medical devices – Part 17: Methods for the establishment of allowable limits
ISO 10993-18:2005, Biological evaluation of medical devices – Part 18: Chemical characterization of materials
ISO 10993-5:2009, Biological evaluation of medical devices – Part 5: Tests for in vitro cytotoxicity

ISO 10993-6, 2007, Biological Evaluation of Medical Devices - Part 6: Tests for Local Effects After Implantation

ISO 10993-7:2008, Biological Evaluation of Medical Devices - Part 7: Ethylene Oxide Sterilization Residuals

ISO 11607-1: 2006, Packaging for terminally sterilized medical devices

United States Pharmacopeia 36, National Formulary 31, 2013. <151> Pyrogen Test

ISO 2233: 2000, Packaging — Complete, filled transport packages and unit loads — Conditioning for testing

The following list details significant testing that was successfully completed:

Performance Tests cont.

Guide Wire Compatibility
Delivery Device Working Length
Delivery Device Stent Pod OD
Delivery Device Shaft OD
Trigger Stroke
Insertion Force
Distal Tip Insertion & Flexibility/Kink Resistance
Repositioning
Delivery System Deployment Accuracy
Stent Expansion & Condition After Deployment
System Integrity
Stent Foreshortening
Removal
Migration & Removal Force
AM-Strut Height
Suture Purse String
Suture Tensile
Stent Tensile
Stent Fatigue
Cover Integrity After Fatigue
Coating Integrity After Fatigue
Stent Spring Back After Fatigue
Compression Force After Fatigue
Expansion Force After Fatigue
Trigger Safety
Delivery System Tensile Strength Tests
Fluoroscopic Visibility of Deployment Catheter
Endoscopic Visibility of Deployment Catheter
Atraumatic Tip
MR Compatibility

Biocompatibility

Cytotoxicity
Sensitization
Implantation
Irritation
Material Mediated Pyrogenicity
Chemical Characterization

**Performance
Tests cont.**

Packaging Performance

Seal Peel Strength
Visual Inspection
Bubble Emission

The results of the testing demonstrated that the subject AEROMini met the predetermined acceptance criteria applicable to the substantial equivalence of the device.

**Summary of
Substantial
Equivalence**

Based on the indications for use, design, and performance testing, the subject AEROMini meets the requirements that are considered essential for its intended use and is substantially equivalent to the predicate device, the AERO Tracheobronchial Stent Technology System, K082284.
