



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center – WO66-G609
Silver Spring, MD 20993-0002

September 3, 2014

WRP Rubber Products Sdn Bhd
Mr. Lim Chin Yit
Head of Operations
Lot 3, Jalan 3, Kawasan Perusahaan
Bandar Baru Salak Tinggi
Sepang, Selangor Darul Ehsan
MALAYSIA 43900

Re: K140418

Trade/Device Name: Powder Free Blue Nitrile Patient Examination Gloves,
Non-Sterile

Regulation Number: 21 CFR 880.6250

Regulation Name: Patient Examination Gloves

Regulatory Class: I

Product Code: LZA

Dated: July 10, 2014

Received: July 14, 2014

Dear Mr. Yit:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

Tejashri Purohit-Sheth, M.D.

Tejashri Purohit-Sheth, M.D.
Clinical Deputy Director
DAGRID/ODE/CDRH FOR

Erin I. Keith, M.S.
Director
Division of Anesthesiology, General Hospital
Respiratory, Infection Control and Dental Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

DEPARTMENT OF HEALTH AND HUMAN SERVICES
Food and Drug Administration

Form Approved: OMB No. 0910-0120
Expiration Date: January 31, 2017
See PRA Statement on last page.

Indications for Use

510(k) Number (if known)
K140418

Device Name
Powder Free Blue Nitrile Patient Examination Gloves Non-Sterile

Indications for Use (Describe)

A patient examination glove is a disposable device intended for medical purposes that is worn on the examiner's hand or finger to prevent contamination between patient and examiner.

Type of Use (Select one or both, as applicable)

- Prescription Use (Part 21 CFR 801 Subpart D) Over-The-Counter Use (21 CFR 801 Subpart C)

PLEASE DO NOT WRITE BELOW THIS LINE – CONTINUE ON A SEPARATE PAGE IF NEEDED.

FOR FDA USE ONLY

Concurrence of Center for Devices and Radiological Health (CDRH) (Signature)

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRAStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

1.0 Submitter:

Name: Mr. Lim Chin Yit
Address: WRP Rubber Products Sdn Bhd
Lot 3, Jalan 3, Kawasan Perusahaan Bandar Baru Salak Tinggi,
43900 Sepang, Selangor Darul Ehsan, MALAYSIA
Phone No.: +60166957817

Date of Summary Prepared: 18 August 2014

2.0 Name of the device:

Powder Free Blue Nitrile Patient Examination Gloves, Non-Sterile
Common Name: Exam Gloves
Classification Name: Patient Examination Gloves (21 CFR 880.6250 product code LZA)

3.0 Identification of The Legally Marketed Devices that equivalency is claimed:

Dermagrip-N Powder Free Blue Nitrile Examination Gloves
510(k) : K022904
MDL : D036500
Regulatory Class I
Product Code : LZA

4.0 Description of The Device:

Powder Free Blue Nitrile Patient Examination Glove, Non-Sterile meet all the requirements of ASTM standard D6319-10 and FDA 21 CFR 880.6250.

The powder free nitrile examination glove is manufactured from synthetic rubber latex. Inner surface of gloves undergo surface treatment process to produce a smooth surface that assists the user in donning the gloves with ease without using any lubricant such as powder on the glove surface. The glove is ambidextrous, i.e. can be worn on right hand or left hand.

5.0 Intended Use of the Device:

A patient examination glove is a disposable device intended for medical purposes that is worn on the examiner's hand or finger to prevent contamination between patient and examiner.

510(k) SUMMARY

6.0 Summary of the Technological Characteristics of the Device:

The Powder Free Blue Nitrile Examination Gloves, Non-Sterile are summarized with the following technological characteristics compared to ASTM D6319 or equivalent standards.

CHARACTERISTICS	STANDARDS	DEVICE PERFORMANCE	
		Predicate	Current
Dimensions	ASTM D6319-10	Meets	Meets Length min 230mm Width min 95±10
Physical Properties	ASTM D6319-10	Meets	Meets Before Aging Tensile Strength min 14 MPa Ultimate Elongation Min 500% After Aging Tensile Strength min 14 MPa Ultimate Elongation Min 400%
Thickness	ASTM D6319-10	Meets	Meets Finger min 0.05mm Palm min 0.05mm
Powder Free	ASTM D6124-06	Meets ≤ 2 mg/glove	Meets ≤ 2 mg/glove
Biocompatibility	Primary Skin Irritation – ISO 10993-10:2010(E) & Consumer Product Safety Commission, Title 16, Chapter II, Part 1500	Under the conditions of the study, not an irritant	Under the conditions of the study, not an irritant.
	Dermal Sensitization - ISO 10993-10:2010(E) & Consumer Product Safety Commission, Title 16, Chapter II, Part 1500.3(c)(4)	Under the conditions of the study, not a sensitizer.	Under the conditions of the study, not a sensitizer.

510(k) SUMMARY

CHARACTERISTICS	STANDARDS	DEVICE PERFORMANCE	
		Predicate	Current
Watertight (1000ml)	ASTM D5151-06	Passes	Passes AQL 2.5
Intended use	-	The powder free blue Nitrile examination glove, non-sterile is a disposable device and is made of synthetic rubber intended for medical purposes that is worn on the examiner's hand or finger to prevent contamination between patient and examiner.	A patient examination glove is a disposable device intended for medical purposes that is worn on the examiner's hand or finger to prevent contamination between patient and examiner.
Material	ASTM D6319-10	Nitrile	Nitrile
Color	-	Blue	Blue
Texture	-	Finger textured	Finger textured
Size	Medical Glove Guidance Manual - Labeling	Small Medium Large Extra Large	Extra Small Small Medium Large Extra Large
Single Use	Medical Glove Guidance Manual - Labeling	Single use	Single use
Manufacturer(s)	-	WRP Asia Pacific Sdn Bhd	WRP Rubber Products Sdn Bhd

There are no significant differences between the two products and are identical in terms of intended use, materials, design, manufacturing methods except for thickness. The current device is thinner than the predicate device, however it meets the ASTM standards.

510(k) SUMMARY

7.0 Substantial Equivalent Based on Assessment of Non-Clinical Performance Data

The performance test data of the non-clinical tests that support a determination of substantial equivalence is the same as mentioned immediately above (ASTM Requirements).

8.0 Substantial Equivalent Based on Assessment of Clinical Performance Data

Not applicable - Clinical data is not needed for gloves or for most devices cleared by the 510(k) process.

9.0 Conclusion

Powder Free Blue Nitrile Patient Examination Glove, Non-Sterile will perform according to the gloves performance standards referenced in section 6.0 above and meet ASTM standards, and FDA requirements for water leak test on pinhole AQL. Consequently, the device is as safe and as effective and substantially equivalent to the predicate device.