

**510(k) Summary**

In accordance with Title 21 of the Code of Federal Regulations, Part 807, and in particular 21 CFR §807.92, the following summary of information is provided:

**A. Submitted by:**

Cynthia Adams  
Regulatory Affairs Associate  
NuVasive, Incorporated  
7475 Lusk Blvd.  
San Diego, California 92121  
Telephone: (858) 909-1800

Date Prepared: July 23, 2014

**B. Device Name**

Trade or Proprietary Name:	<i>NuVasive® CoRoent® XL-F System</i>
Common or Usual Name:	Intervertebral Body Fusion Device
Classification Name:	Spinal Intervertebral Body Fixation orthosis

Device Class:	Class II
Classification:	21 CFR § 888.3080
Product Code:	OVD

**C. Predicate Devices**

The subject *NuVasive CoRoent XL-F System* is substantially equivalent to the following predicate devices: *NuVasive Single Tab System* (K131723), *NuVasive Brigade Hyperlordotic System* (K123045), and *Globus Medical InterContinental Plate-Spacer* (K103382).

**D. Device Description**

The *NuVasive CoRoent XL-F System* is manufactured from PEEK-Optima® LT-1 conforming to ASTM F2026, MP35N conforming to ASTM F562, and titanium alloy conforming to ASTM F136 and ISO 5832-3. The implants are available in a variety of different shapes and sizes to suit the individual pathology and anatomical conditions of the patient.



**E. Intended Use**

The NuVasive CoRoent XL-F System is indicated for intervertebral body fusion of the spine in skeletally mature patients. The System is designed for use with autogenous bone graft to facilitate fusion.

The CoRoent XL-F System is intended for use at either one level or two contiguous levels in the lumbar spine, from L2 to L5, for the treatment of degenerative disc disease (DDD) with up to Grade I spondylolisthesis. DDD is defined as back pain of discogenic origin with degeneration of the disc confirmed by history and radiographic studies. The lumbar devices are to be used in patients who have had at least six months of non-operative treatment. The system is intended to be used with supplemental internal spinal fixation systems (e.g., pedicle or facet screws) that are cleared by the FDA for use in the lumbar spine in addition to the integrated screws.

**F. Technological Characteristics**

As was established in this submission, the subject *CoRoent XL-F System* is substantially equivalent to other predicate devices cleared by the FDA for commercial distribution in the United States. The subject device was shown to be substantially equivalent and have the same technological characteristics to its predicate devices through comparison in areas including design, intended use, material composition, and function. This device does not contain software or electrical equipment.

**G. Performance Data**

The following activities were performed to demonstrate that the subject *CoRoent XL-F System* is substantially equivalent to other predicate devices for the following:

- Axial Compression Finite Element Analysis
- Compression Shear Finite Element Analysis
- Wear Debris Analysis
- Subsidence Analysis
- Clinical literature analysis
- Cadaveric study

The results demonstrate that the subject *CoRoent XL-F System* presents no new worst-case for performance testing, and a cadaveric study did not identify any new risks associated with the subject device. The subject device was therefore found to be substantially equivalent to its predicate devices. No clinical studies were conducted.

**H. Conclusions**

Based on the indications for use, technological characteristics, and comparison to predicate devices, the subject *CoRoent XL-F System* has been shown to be substantially equivalent to legally marketed predicate devices.

---



Food and Drug Administration  
10903 New Hampshire Avenue  
Document Control Center - WO66-G609  
Silver Spring, MD 20993-0002

July 28, 2014

NuVasive, Incorporated  
Ms. Cynthia Adams  
Regulatory Affairs Associate  
7475 Lusk Boulevard  
San Diego, California 92121

Re: K140479  
Trade/Device Name: NuVasive<sup>®</sup> CoRoent<sup>®</sup> XL-F System  
Regulation Number: 21 CFR 888.3080  
Regulation Name: Intervertebral body fusion device  
Regulatory Class: Class II  
Product Code: OVD  
Dated: June 17, 2014  
Received: June 18, 2014

Dear Ms. Adams:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set

Page 2 – Ms. Cynthia Adams

forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

**Ronald P. Jean -S** for

Mark N. Melkerson  
Director  
Division of Orthopedic Devices  
Office of Device Evaluation  
Center for Devices and  
Radiological Health

Enclosure

**Indications for Use**

510(k) Number (if known)  
K140479

Device Name  
NuVasive® CoRoent® XL-F System

*Indications for Use (Describe)*

The NuVasive CoRoent XL-F System is indicated for intervertebral body fusion of the spine in skeletally mature patients. The System is designed for use with autogenous bone graft to facilitate fusion.

The CoRoent XL-F System is intended for use at either one level or two contiguous levels in the lumbar spine, from L2 to L5, for the treatment of degenerative disc disease (DDD) with up to Grade I spondylolisthesis. DDD is defined as back pain of discogenic origin with degeneration of the disc confirmed by history and radiographic studies. The lumbar devices are to be used in patients who have had at least six months of non operative treatment. The system is intended to be used with supplemental internal spinal fixation systems (e.g., pedicle or facet screws) that are cleared by the FDA for use in the lumbar spine in addition to the integrated screws.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

**PLEASE DO NOT WRITE BELOW THIS LINE – CONTINUE ON A SEPARATE PAGE IF NEEDED.**

**FOR FDA USE ONLY**

Concurrence of Center for Devices and Radiological Health (CDRH) (Signature)

**Katherine D. Kavlock, PhD**

**Division of Orthopedic Devices**

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
PRAStaff@fda.hhs.gov

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*