



JUL 18 2014

**510(k) Summary****NanoCross™ Elite 0.014" Over-the-Wire PTA Balloon Dilatation Catheter**

<b>510(k) Summary</b>	This 510(k) summary information is submitted in accordance with the requirements of 21 CFR §807.92.
<b>Applicant</b>	ev3 Inc.
<b>Submitter</b>	ev3 Inc. 3033 Campus Drive Plymouth, MN 55441-2651  Tel: 763.398.7000 Fax: 763.591.3248
<b>Contact Person</b>	Laura J. Lind
<b>Date Prepared</b>	June 18, 2014
<b>Device Trade Name</b>	NanoCross™ Elite 0.014" Over-the-Wire PTA Balloon Dilatation Catheter
<b>Device Common Name</b>	PTA Dilatation Catheter
<b>Classification Name</b>	Catheter, Angioplasty, Peripheral, Transluminal (21 CFR §870.1250, Product Code LIT)
<b>Classification Panel</b>	Cardiovascular
<b>Predicate Devices</b>	NanoCross™ Elite 0.014" Over-the-Wire PTA Balloon Dilatation Catheter (K132777), PowerCross™ 0.018" OTW PTA Dilatation Catheter (K093286).
<b>Intended use</b>	The NanoCross™ Elite 0.014" Over-the-Wire PTA Balloon Dilatation Catheter is intended to dilate stenoses in the iliac, femoral, ilio-femoral, popliteal, infra-popliteal, and renal arteries, and for the treatment of obstructive lesions of native or synthetic arteriovenous dialysis fistulae. This device is also indicated for stent post-dilatation in the peripheral vasculature.
<b>Device Description</b>	The NanoCross catheter is an over-the-wire (OTW) coaxial lumen percutaneous transluminal angioplasty (PTA) balloon catheter compatible with 0.014" guidewires with a distally mounted semi-compliant inflatable balloon and an atraumatic tapered tip. The distal portion of the catheter has a lubricious coating. The manifold includes a lumen marked "THRU". This is the central lumen of the catheter, which terminates at the distal

## NanoCross™ Elite 0.014" Over-the-Wire PTA Balloon Dilatation Catheter Summary

tip. This lumen is used to pass the catheter over a guidewire with a maximum diameter of 0.014". The lumen marked "BALLOON" is the balloon inflation lumen, which is used to inflate and deflate the dilatation balloon with a mixture of contrast medium and saline solution. The balloon has two radiopaque markers for positioning the balloon relative to the stenosis. The radiopaque marker bands indicate the dilating or working section of the balloon.

The NanoCross Elite catheter is available in balloon sizes ranging from 1.5 mm to 6 mm in diameter, and from 20 mm to 210 mm in length; reference labeling for introducer sheath compatibility.

**Performance data**

Bench testing was performed to support a determination of substantial equivalence. Results from this testing provide assurance that the proposed device has been designed and tested to assure conformance to the requirements for its intended use.

Using internal Risk Analysis procedures, the following tests were performed for one or more model sizes:

• Dimensional Verification: Crossing Profile	• Dimensional Verification: Balloon OD
• Balloon Rated Burst Pressure	• Balloon Rated Burst Pressure (In Stent)
• Balloon Compliance	• Radiopacity
• Balloon Pull-back Force	• Presence of Coating
• Dimensional Verification: Balloon Length	• Dimensional Verification: Tip/Lesion Entry Profile
• Inflation/Deflation Time	• Pushability
• Balloon Fatigue	• Dimensional Verification: Tip ID
• Catheter Bond Strength	• Wire Movement
• Kink	• Re-Insertion Force
• Device Tracking	• Catheter Working Length
• Insertion Force	• Torque Strength
• Particle Generation	• Coating Durability

Using the same Risk Analysis procedures, the following tests were leveraged from predicate devices for one or more model sizes:

Biocompatibility per ISO 10993-1
----------------------------------

The NanoCross Elite catheter met all acceptance criteria for the bench testing with results similar to the predicates. Based on the bench test results, comparison to legally marketed predicates, and non-clinical test results, the NanoCross Elite catheter is determined to perform as safely and effectively as the predicates for its intended use.

## NanoCross™ Elite 0.014" Over-the-Wire PTA Balloon Dilatation Catheter Summary

**Summary of  
Substantial  
Equivalence**

---

The NanoCross Elite catheter has the following similarities to one or more of the predicate devices:

- Same fundamental scientific technology
- Identical intended use
- Same operating principle
- Identical balloon rated burst pressures
- Identical balloon nominal pressure
- Similar balloon diameters
- Similar balloon lengths
- Identical catheter lengths
- A lubricious coating
- Packaged with the same materials and processes
- Same sterility assurance level and sterilization method

The devices are compatible with 0.014" guidewires. All devices have similar construction and principles of operation. All devices are used by the physician in a similar manner typical of PTA balloon catheters.

The NanoCross Elite catheter and the predicates have the same intended use - all devices are intended to treat peripheral arteries. All devices are intended to treat the same target population. The manner in accessing and treating lesions is the same.

**Conclusion**

---

Based on the intended use, technological characteristics, and results from safety and performance testing, the modified NanoCross™ Elite 0.014" Over-the-Wire PTA Balloon Dilatation Catheter is considered substantially equivalent to the legally marketed predicate devices NanoCross™ Elite 0.014" Over-the-Wire PTA Balloon Dilatation Catheter (K132777) and PowerCross™ .018" OTW PTA Dilatation Catheter (K093286).

---



Food and Drug Administration  
10903 New Hampshire Avenue  
Document Control Center - WO66-G609  
Silver Spring, MD 20993-0002

July 18, 2014

ev3 Inc.  
Ms. Brenda Johnson  
Regulatory Affairs Manager  
3033 Campus Drive  
Suite N550  
Plymouth, MN 55441

Re: K141118  
Trade/Device Name: NanoCross Elite 0.014" Over-the-Wire PTA Balloon Dilatation Catheter  
Regulation Number: 21 CFR 871.1250  
Regulation Name: Catheter, Angioplasty, Peripheral, Transluminal  
Regulatory Class: Class II  
Product Code: LIT  
Dated: June 18, 2014  
Received: June 19, 2014

Dear Ms. Johnson,

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in

Page 2 - Ms. Brenda Johnson

the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638 2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

  
**Kenneth J. Cavanaugh -S**  
for  
Bram D. Zuckerman, M.D.  
Director  
Division of Cardiovascular Devices  
Office of Device Evaluation  
Center for Devices and Radiological Health

Enclosure

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
Food and Drug Administration

Form Approved: OMB No. 0910-0120  
Expiration Date: January 31, 2017  
See PRA Statement below.

## Indications for Use

510(k) Number (if known)

n/a

Device Name

NanoCross™ Elite 0.014" Over-the-Wire PTA Balloon Dilatation Catheter

Indications for Use (Describe)

The NanoCross™ Elite 0.014" Over-the-Wire PTA Balloon Dilatation Catheter is intended to dilate stenoses in the iliac, femoral, ilio-femoral, popliteal, infra-popliteal, and renal arteries, and for the treatment of obstructive lesions of native or synthetic arteriovenous dialysis fistulae. This device is also indicated for stent post-dilatation in the peripheral vasculature.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

PLEASE DO NOT WRITE BELOW THIS LINE – CONTINUE ON A SEPARATE PAGE IF NEEDED.

FOR FDA USE ONLY

Concurrence of Center for Devices and Radiological Health (CDRH) (Signature)

Kenneth J. Cavanaugh -S

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
PRASStaff@fda.hhs.gov

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*