

Food and Drug Administration 10903 New Hampshire Avenue Document Control Center – WO66-G609 Silver Spring, MD 20993-0002

September 11, 2014

Cynosure Incorporated Ms. Kelli McMillan Regulatory Affairs Specialist 5 Carlisle Road Westford, Massachusetts 01886

Re: K141425

Trade/Device Name: Cynosure Elite+ Laser Regulation Number: 21 CFR 878.4810

Regulation Name: Laser surgical instrument for use in general and

plastic surgery and in dermatology

Regulatory Class: Class II Product Code: GEX Dated: August 11, 2014 Received: August 13, 2014

Dear Ms. McMillan:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you; however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the <u>Federal Register</u>.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21

CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm.

Sincerely yours,

David Krause -S

for Binita S. Ashar, M.D., M.B.A., F.A.C.S.
Director
Division of Surgical Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

DEPARTMENT OF HEALTH AND HUMAN SERVICES Food and Drug Administration

Indications for Use

Form Approved: OMB No. 0910-0120 Expiration Date: January 31, 2017 See PRA Statement below.

510(k) Number <i>(if known)</i> K141425
Device Name Cynosure Elite+ Laser
Indications for Use (Describe) 755 nm:
The Cynosure Elite+ Laser is indicated for stable long-term, or permanent hair reduction. Permanent hair reduction is defined as long-term stable reduction in the number of hairs regrowting when measured 6, 9, or 12 months after the completion of a treatment regime. It is used for all skin types (Fitzpatrick I-VI) including tanned skin. It is also indicated for the treatment of vascular lesions, benign pigmented lesions, and wrinkles.
The Cynosure Elite+ Laser is intended for the coagulation and hemostasis of benign vascular lesions such as, but not limited to, port wine stains, hemangiomas, warts, telangectasia, rosacea, venus lake, leg veins, spider veins and poikiloderma of civatte; and treatment of benign cutaneous lesions such as warts, scars, striae and psoriasis. The laser is also intended for the treatment of benign pigmented lesions such as, but not limited to, lentigos (age spots), solar lentigos (sun spots), cafe au lait macules, seborrheic keratoses, nevi, chloasma, verrucae, skin tags, keratosis and plaques. The laser is also indicated for the treatment of wrinkles such as, but not limited to, periocular and perioral wrinkles. Additionally, the laser is indicated for the treatment of pseudofolliculitis barbae (PFB) and for stable long-term, or permanent hair reduction. Permanent hair reduction is defined as long-term stable reduction in the number of hairs regrowing when measured at 6, 9 and 12 months after the completion of a treatment regime.
Type of Use (Select one or both, as applicable)
Prescription Use (Part 21 CFR 801 Subpart D) Over-The-Counter Use (21 CFR 801 Subpart C)
CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services Food and Drug Administration Office of Chief Information Officer Paperwork Reduction Act (PRA) Staff PRAStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510k Summary Elite+ Laser Special 510(k) K141425

807.92(a)(1) Submitter Information			
Applicant	Cynosure, Inc		
Address	5 Carlisle Road		
	Westford, MA 01886		
Phone Number	(781) 993-2454		
Fax Number	(978) 256-6556		
Establishment Registration Number	1222993		
Contact Person	Kelli McMillan		
Preparation Date	September 10, 2014		
807.92(a)(2) Name of Device			
Trade or Proprietary Name	Elite+ Laser		
Common or Usual Name	Medical Laser System		
Classification	II		
Regulation Name	Laser Surgical Instrument for Use in General and Plastic Surgery and in Dermatology		
Classification Panel	General & Plastic Surgery		
Regulation	878.4810		
Product Code(s)	GEX		
807.92 (a)(3) Legally marketed device	(s) to which equivalence is claimed		
	Apogee Elite K034030		
807.92(2)(4) Device Description			
	Cynosure Elite+ laser has an Alexandrite crystal rod as the lasing medium, with a wavelength of 755 nm. The 1064 nm wavelength is generated from the Nd:YAG laser head.		
	Laser activation is by either by finger or foot switch. The overall weight of the laser is 180 lbs and the size is 41 x 15 x 25 in (HxWxD).		
	Electrical equipment is 208/240 VAC, single phase, 30A, 50/60 Hz.		
	The modifications to this device are four handpieces allowing for a larger treatment area; 18 mm, 20 mm, 22 mm and 24 mm. The 1064 Nd:YAG laser did not require any modifications to support the new handpieces. Minor modifications, including rod diameter change and lamp fill pressure change, were made to the Alexandrite laser to increase the energy to support the larger spot sizes. The software continues to control the energy based on the handpiece attached. This prevents the increase in available energy from the laser from increasing the fluence of		
	the existing spot sizes.		
807/92(a)(5) Intended Use of the Dev	rice		
	755 nm:		
	The Cynosure Elite+ Laser is indicated for stable long-term, or		
	permanent hair reduction. Permanent hair reduction is defined		
	as long-term stable reduction in the number of hairs regrowting		

510k Summary Elite+ Laser Special 510(k) K141425

when measured 6, 9, or 12 months after the completion of a treatment regime. It is used for all skin types (Fitzpatrick I-VI) including tanned skin. It is also indicated for the treatment of vascular lesions, benign pigmented lesions, and wrinkles.

1064 nm:

The Cynosure Elite+ Laser is intended for the coagulation and hemostasis of benign vascular lesions such as, but not limited to, port wine stains, hemangiomas, warts, telangectasia, rosacea, venus lake, leg veins, spider veins and poikiloderma of civatte; and treatment of benign cutaneous lesions such as warts, scars, striae and psoriasis. The laser is also intended for the treatment of benign pigmented lesions such as, but not limited to, lentigos (age spots), solar lentigos (sun spots), cafe au lait macules, seborrheic keratoses, nevi, chloasma, verrucae, skin tags, keratosis and plaques.

The laser is also indicated for the treatment of wrinkles such as, but not limited to, periocular and perioral wrinkles. Additionally, the laser is indicated for the treatment of pseudofolliculitis barbae (PFB) and for stable long-term, or permanent hair reduction. Permanent hair reduction is defined as long-term stable reduction in the number of hairs regrowing when measured at 6, 9 and 12 months after the completion of a treatment regime.

510k Summary Elite+ Laser Special 510(k) K141425

	Elite+ K141425 Alexandrite and Nd:YAG Lasers		Apogee Elite K034030 Alexandrite and Nd:YAG Lasers		
Laser Type					
Wavelength	755 nm	1064 nm	755 nm	1064 nm	
Maximum Fluence	60 J/cm	300 J/cm	60 J/cm	300 J/cm	
Repetition Rate	1 to 5 Hz	1 to 10 Hz	1 to 5 Hz	1 to 10 Hz	
Pulse Duration	0.1-300 ms		0.1-300 ms		
Spot Sizes (mm)	3, 5, 7, 10, 12, 15, 18, 20, 22, 24		3, 5, 7, 10, 12.5, 15		
807.92(b)(1) Non-clinical tests submitted					
Test	Result				
Software	The testing confirmed that the software specifications were met.				

None

807.92(b)(3) Conclusions drawn from non-clinical data

Testing confirmed that the performance of the Elite+ Laser meets the product system requirements, which is based on the predicate device. Therefore, the modification resulted in a device that performs within the same specifications of the predicate device, and is therefore substantially equivalent.