



Food and Drug Administration  
10903 New Hampshire Avenue  
Document Control Center – WO66-G609  
Silver Spring, MD 20993-0002

GE Medical Systems, LLC  
% Ms. Jenny Wong  
Regulatory Affairs Leader, Magnetic Resonance  
3200 N. Grandview Blvd.  
WAUKESHA WI 53188

December 5, 2014

Re: K142361  
Trade/Device Name: Discovery MR750 3.0T  
Regulation Number: 21 CFR 892.1000  
Regulation Name: Magnetic resonance diagnostic device  
Regulatory Class: II  
Product Code: LNH  
Dated: November 21, 2014  
Received: November 24, 2014

Dear Ms. Wong:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638 2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, “Misbranding by reference to premarket notification” (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH’s Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,



Michael D. O'Hara  
for

Janine M. Morris  
Director  
Division of Radiological Health  
Office of In Vitro Diagnostics  
and Radiological Health  
Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)

K142361

Device Name

Discovery MR750 3.0T

Indications for Use (Describe)

The Discovery MR750 3.0T system is a whole body magnetic resonance scanner designed to support high resolution, high signal-to-noise ratio, and short scan times. It is indicated for use as a diagnostic imaging device to produce axial, sagittal, coronal, and oblique images, spectroscopic images, parametric maps, and/or spectra, dynamic images of the structures and/or functions of the entire body, including, but not limited to, head, neck, TMJ, spine, breast, heart, abdomen, pelvis, joints, prostate, blood vessels, and musculoskeletal regions of the body. Depending on the region of interest being imaged, contrast agents may be used.

The images produced by the Discovery MR750 3.0T system reflect the spatial distribution or molecular environment of nuclei exhibiting magnetic resonance. These images and/or spectra when interpreted by a trained physician yield information that may assist in diagnosis.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

**PLEASE DO NOT WRITE BELOW THIS LINE – CONTINUE ON A SEPARATE PAGE IF NEEDED.**

### FOR FDA USE ONLY

Concurrence of Center for Devices and Radiological Health (CDRH) (Signature)

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*“An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number.”*



GE Healthcare  
510(k) Premarket Notification Submission

**510(k) Summary**

In accordance with 21 CFR 807.92 the following summary of information is provided:

Date: August 22, 2014

Submitter: GE Healthcare, (GE Medical Systems, LLC)  
3200 N Grandview Blvd.  
Waukesha, WI 53188

Primary Contact Person: Jenny Wong  
Regulatory Affairs Leader, MR  
GE Healthcare, (GE Medical Systems, LLC)  
Phone: 262-548-2899  
Fax: 414-908-9585

Secondary Contact Person: Glen Sabin  
Regulatory Affairs Director, MR  
GE Healthcare, (GE Medical Systems, LLC)  
Phone: 262-521-6848  
Fax: 414-908-9585

Device: Trade Name: Discovery MR750 3.0T

Common/Usual Name: Magnetic Resonance Diagnostic Device

Classification Names: 892.1000

Product Code: LNH

Predicate Device(s): Discovery MR750 3.0T and Discovery MR750w 3.0T [K132376]

Device Description: The Discovery MR750 3.0T features a superconducting magnet operating at 3.0 Tesla. The data acquisition system accommodates up to 32 independent receive channels in various increments, and multiple independent coil elements per channel during a single acquisition series. The system uses a combination of time-varying magnetic fields (gradients) and RF transmissions to obtain information regarding the density and position of elements exhibiting magnetic resonance. The system can image in the sagittal, coronal, axial, oblique and double oblique planes, using various pulse sequences and reconstruction algorithms. The Discovery MR750 3.0T uses multi-drive RF transmit for imaging of the head and body regions.

The Discovery MR750 3.0T is designed to conform to NEMA DICOM



GE Healthcare  
510(k) Premarket Notification Submission

standards.

**Intended Use:** The Discovery MR750 3.0T system is a whole body magnetic resonance scanner designed to support high resolution, high signal-to-noise ratio, and short scan times. It is indicated for use as a diagnostic imaging device to produce axial, sagittal, coronal, and oblique images, spectroscopic images, parametric maps, and/or spectra, dynamic images of the structures and/or functions of the entire body, including, but not limited to, head, neck, TMJ, spine, breast, heart, abdomen, pelvis, joints, prostate, blood vessels, and musculoskeletal regions of the body. Depending on the region of interest being imaged, contrast agents may be used.

The images produced by the Discovery MR750 3.0T system reflect the spatial distribution or molecular environment of nuclei exhibiting magnetic resonance. These images and/or spectra when interpreted by a trained physician yield information that may assist in diagnosis.

**Technology:** The Discovery MR750 3.0T employs the same fundamental scientific technology as its predicate devices.

The following is a summary of the differences between the proposed Discovery MR750 3.0T and its predicate Discovery MR750 3.0T and Discovery MR750w 3.0T (K132376).

- New base software (version DV25)
- Software and hardware modifications to enable of the multi-drive RF transmit technology for body and head
- Hardware modifications to the host computer and reconstruction engine for obsolescence and compatibility with a new operating system
- Revised and updated labeling to support new feature

**Determination of Substantial Equivalence:**

**Summary of Non-Clinical Tests:**

The Discovery MR750 3.0T system with the addition of the new multi-drive RF transmit function complies with the following voluntary standards:

- IEC 60601-1
- IEC 60601-1-2
- IEC 60601-2-33



## GE Healthcare 510(k) Premarket Notification Submission

In addition, the Discovery MR750 3.0T system in compliance with the applicable NEMA standards, including NEMA PS3.1-3.20 for DICOM conformance.

The following quality assurance measures were applied to the development of the system:

- Risk Analysis
- Requirements Reviews
- Design Reviews
- Testing on unit level (Module verification)
- Integration testing (System verification)
- Performance testing (Verification)
- Safety testing (Verification)
- Simulated use testing (Validation)

The non-clinical tests have been summarized in the verification testing that was completed for the addition of the multi-drive RF transmit mode. The testing was completed with passing results per the pass/fail criteria defined in the test cases. This supports substantial equivalence to its predicates because it was also developed under quality assurance Design Controls. In addition, Discovery MR750 3.0T is in compliance to the same Standards.

### **Summary of Clinical Tests:**

The subject of this premarket submission, Discovery MR750 3.0T, did not require external clinical studies to support substantial equivalence. Internal scans were conducted as part of validation for workflow and image quality for the addition of the new feature. The clinical results demonstrated that the Discovery MR750 3.0T maintains the same imaging performance results as its predicate device (K132376). Sample clinical images are included in this submission.

**Conclusion:** GE Healthcare considers the Discovery MR750 3.0T to be as safe, as effective, and performance is substantially equivalent to the predicate device(s).