



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center – WO66-G609
Silver Spring, MD 20993-0002

Sonoscape Company Limited
% Mrs. Toki Wu
Regulatory Affairs Manager
Yizhe Building, Yuquan Road, Nanshan
Shenzhen 518051
P.R. CHINA

January 29, 2015

Re: K142714
Trade/Device Name: S8 Exp/S9 Pro Portable Digital Color Doppler Ultrasound System
Regulation Number: 21 CFR 892.1550
Regulation Name: Ultrasonic pulsed doppler imaging system
Regulatory Class: II
Product Code: IYN, IYO, ITX
Dated: September 18, 2014
Received: January 09, 2015

Dear Mrs. Wu:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638 2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, “Misbranding by reference to premarket notification” (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH’s Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

A handwritten signature in black ink that reads "Robert A. Ochs". The signature is written in a cursive style. A faint, large "FDA" watermark is visible in the background behind the signature.

Robert Ochs, Ph.D.
Acting Director
Division of Radiological Health
Office of In Vitro Diagnostics
and Radiological Health
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)
K142714

Device Name
S8 Exp/S9 Pro Portable Digital Color Doppler Ultrasound System

Indications for Use (Describe)

The SonoScape S8 Exp/S9 Pro system is a general-purpose ultrasonic imaging instrument intended for use by a qualified physician for evaluation of Fetal, Abdominal, Pediatric, Small Organ (breast, testes, thyroid), Cephalic (neonatal and adult), Trans-rectal, Trans-vaginal, Peripheral Vascular, Cerebral Vascular, Musculo-skeletal (Conventional and Superficial), Cardiac (neonatal and adult), Trans-esoph.(Cardiac), Laparoscopic, OB/Gyn and Urology.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) Summary

1. Submitter [21 CFR807.92 (a) (1)]

Submitter: SonoScape Company Limited
 Address: Yizhe Building, Yuquan Road, Nanshan, Shenzhen
 518051, P.R.China
 Contact Person: Toki Wu
 Tel: +86 755 26722890
 Fax: +86 755 26722850
 Email: wusq@sonoscape.net
 Date Prepared September 18, 2014

2. Device [21 CFR807.92 (a)(2)]

Trade Name: S8 Exp/S9 Pro Portable Digital Color Doppler Ultrasound System
 Common Name: Diagnostic Ultrasound System and Transducers

Classification Regulatory:

	<u>FR Number</u>	<u>Product Code</u>
Ultrasonic Pulsed Doppler Imaging System	892.1550	90-IYN
Ultrasonic Pulsed Echo Imaging System	892.1560	90-IYO
Diagnostic Ultrasound Transducer	892.1570	90-ITX

Classification Panel: Radiology

Device Class: II

3. Predicate Device(s) [21 CFR 807.92(a) (3)]

The identified predicate devices within this submission are as follows:

SonoScape S8 Exp Portable Digital Color Doppler Ultrasound System	K132768
SonoScape S30 Digital Color Doppler Ultrasound System	K132527
Philips EPIQ Diagnostic Ultrasound System	K132304
Philips CX50 3.0 Diagnostic Ultrasound System	K123754

4. Device Description [21 CFR 807.92(a) (4)]

This SonoScape S8 Exp/S9 Pro Portable Digital Color Doppler Ultrasound System is an integrated preprogrammed color ultrasound imaging system, capable of producing high detail resolution intended for clinical diagnostic imaging applications.

The basic principle is that system transmits ultrasonic energy into patient body and implements post processing of received echoes to generate onscreen display of anatomic structures and fluid flow within the body.

This system is a Track 3 device that employs a wide array of probes that include linear array, convex array and phased array with a frequency range of 2.0 MHz to 15.0 MHz. This system consists of a portable console with keyboard control panel, power supply module, color LCD monitor and optional probes.

This system is a portable, general purpose, software controlled, color diagnostic ultrasound system. Its basic function is to acquire ultrasound data and to display the image in B-Mode (including Tissue Harmonic Image), M-Mode, TDI, Color-Flow Doppler, Pulsed Wave Doppler, Continued Wave Doppler and Power Doppler, or the combination of these modes, Elastography, 3D/4D.

The subject of this submission is the addition of new indications, new probes and special function to original SonoScape S8 Exp and the addition of SonoScape S9 Pro.

New indications	Trans-esoph.(Cardiac), Cerebral Vascular and Laparoscopic.
New probes	Add C613, MPTEE, MPTEE mini, CWD2.0, CWD5.0, PWD2.0 and LAP7 probes.
Special function	Elastography.
New Model	Add S9 Pro model which is identical to S8 Exp except for the difference of probes configuration.

5. Intended Use [21 CFR 807.92(a)(5)]

The SonoScape S8 Exp/S9 Pro system is a general-purpose ultrasonic imaging instrument intended for use by a qualified physician for evaluation of Fetal, Abdominal, Pediatric, Small Organ (breast, testes, thyroid), Cephalic(neonatal and adult), Trans-rectal, Trans-vaginal, Peripheral Vascular, Cerebral Vascular, Musculo-skeletal (Conventional and Superficial), Cardiac (neonatal and adult), Trans-esoph.(Cardiac), Laparoscopic, OB/Gyn and Urology.

6. Comparison with the Predicate Devices [21 CFR 807.92(a) (6)]

S8 Exp/S9 Pro Portable Digital Color Doppler Ultrasound System is comparable with and substantially equivalent to the predicate devices:

SonoScape S8 Exp Portable Digital Color Doppler Ultrasound System	K132768
SonoScape S30 Digital Color Doppler Ultrasound System	K132527
Philips EPIQ Diagnostic Ultrasound System	K132304

Philips CX50 3.0 Diagnostic Ultrasound System

K123754

Intended Use Comparison:

Compared with SonoScape S30 (K132527), the Subject Device S8 Exp/S9 Pro has the same intended uses except for Laparoscopic indication which is comparable and equivalent to the Predicate Device Philips CX50 3.0 (K123754) and Cerebral Vascular indication which is comparable and equivalent to the Predicate Device Philips EPIQ (K132304). The detailed analysis can be found in the probes comparison.

Technical Characteristics Comparison:

The basic and main technical features of the Subject Device S8 Exp/S9 Pro are the same as the original SonoScape S8 Exp (K132768), including Design, Operation Controls, Display Modes, Measurement Items, Cine Loop, Power Supply, Operating and Storage Condition and Screen Size. For Operation Modes, Elastography is a special Operation Mode for the Subject Device S8 Exp/S9 Pro, but already employed by many marketed devices and considered Substantially Equivalent to the Predicate Device Philips EPIQ (K132304) and detailed comparison analysis can be found in **Substantial Equivalence Comparison**. The detailed technical features can be found in **General Device Descriptions** of the submission.

Probes Comparison:

Subject device S8 Exp/S9 Pro has the similar probes as the predicated device original SonoScape S8 Exp (K132768), SonoScape S30 (K132527), Philips EPIQ (K132304) and Philips CX50 3.0 (K123754).

Table 1 Probes Comparison

Subject device SonoScape S8 Exp/S9 Pro	Predicate Device SonoScape S8 Exp	Remark
C322 Micro-curved Array C344 Curved Array C353 Curved Array C542 Curved Array	C322 Micro-curved Array C344 Curved Array C353 Curved Array C354 Curved Array C362 Curved Array C542 Curved Array	Same
VC6-2 Curved Array	VC6-2 Curved Array	Same
C613 Micro-curved Array	C611 Micro-curved Array C311 Micro-curved Array	SE Analysis1a)
2P1 Phased Array 2P2 Phased Array 3P1 Phased Array	2P1 Phased Array 2P2 Phased Array 3P1 Phased Array	Same

5P1 Phased Array 5P2 Phased Array 8P1 Phased Array	5P1 Phased Array 5P2 Phased Array 8P1 Phased Array	
L741 Linear Array L742 Linear Array L743 Linear Array L752 Linear Array 10L1 Linear Array	L741 Linear Array L742 Linear Array L743 Linear Array L752 Linear Array 10L1 Linear Array	Same
6V1 Micro-curved Array 6V3 Micro-curved Array EC9-5 Micro-curved Array BCC9-5 Micro-curved Array BCL10-5 Biplane (Curved + Linear Array)	6V1 Micro-curved Array 6V3 Micro-curved Array EC9-5 Micro-curved Array BCC9-5 Micro-curved Array BCL10-5 Biplane (Curved + Linear Array)	Same
Subject device SonoScape S8 Exp/S9 Pro	Predicate Device SonoScape S30	Remark
MPTEE Phased Array (Multi-plane) MPTEE mini Phased Array (Multi-plane)	MPTEE Phased Array (Multi-plane) MPTEE mini Phased Array (Multi-plane)	Same
Subject device SonoScape S8 Exp/S9 Pro	Predicate Device Philips CX50 3.0	Remark
CWD2.0, 2.0 MHz CWD5.0, 5.0 MHz Continuous Wave Doppler	D2cwc, 2.0 MHz D5cwc, 5.0 MHz Continuous Wave Doppler	SE Analysis1b)
LAP7, Linear Array, 10.0-5.0 MHz	L10-4 lap, Linear Array, 10.0-4.0 MHz	
Subject device SonoScape S8 Exp/S9 Pro	Predicate Device Philips EPIQ	Remark
PWD2.0, 2.0 MHz Pulsed Wave Doppler	D2tcd, 2.0 MHz Pulsed Wave Doppler	SE Analysis1c)

Note: C344, 2P1, 2P2, 5P1, L741 and 6V1 only for S8 Exp; and 3P1, 5P2 and 8P1 only for S9 Pro.

SE Analysis 1:

- Compared with the predicate device original SonoScape S8 Exp (K132768), there is a new probe (C613) for the subject device. And the frequency, performance and clinical application of C613 probe are the same as C611 for predicate device.
- Compared with the predicate device Philips CX50 3.0 (K123754), the frequency and clinical application of CWD2.0 and CWD5.0 probe is the same as D2cwc and D5cwc cleared for use with predicate device; the clinical application of LAP7 probe is the same

as L10-4 lap cleared for use with predicate device, and though the frequency of them is different, they both meet the clinical use.

- c) Compared with the predicate device Philips EPIQ (K132304), the frequency and clinical application of PWD2.0 probe is the same as D2tcd cleared for use with predicate device.

*Note: detailed description information about the new probes can be found in **Substantial Equivalence Comparison**.*

Moreover, compared with predicated devices, the subject device (S8 Exp/S9 Pro) complies with the same regulation and safety standards and has the consistent acoustic output levels. Therefore they can be considered Substantially Equivalent in safety and effectiveness, and no new risk is raised, so the SE is not affected.

7. Non-Clinical Tests [21 CFR 807.92(b) (1)]

The S8 Exp/S9 Pro Portable Digital Color Doppler Ultrasound System has been evaluated for electrical, mechanical, thermal and electromagnetic compatibility safety, biocompatibility and acoustic output.

Laboratory tests were conducted to verify that the S8 Exp/S9 Pro system met all design specifications and the S8 Exp/S9 Pro system conformed to applicable medical device standards. Phantom test was conducted to verify that the strain Elastography function was effective and Elastography performance met design specifications, including accuracy and repeatability of strainratio measurement and etc.

The S8 Exp/S9 Pro system has been designed and manufactured to meet the following standards: IEC 60601-1, IEC 60601-1-2, IEC 60601-2-37, ISO 10993-5, ISO10993-10, UD2, and UD3.

8. Clinical Test [21 CFR 807.92(b) (2)]

No clinical testing was required.

9. Substantially Equivalent Conclusions [21 CFR 807.92(b) (3)]

In accordance with the 21 CFR Part 807 and based on the information provided in this premarket notification, SonoScape Company Limited concludes that S8 Exp/S9 Pro Portable Digital Color Doppler Ultrasound System is substantially equivalent to the predicate devices with regard to safety and effectiveness.