



Food and Drug Administration  
10903 New Hampshire Avenue  
Document Control Center – WO66-G609  
Silver Spring, MD 20993-0002

October 28, 2014

ALPINION MEDICAL SYSTEMS Co., Ltd.

% Mr. Donghwan Kim

QARA Manager

1, 6 and 7FL Verdi Tower, 72,

Digital-ro(St) 26-gil(Rd), Guro-gu

Seoul 152-848

REPUBLIC OF KOREA

Re: K142733

Trade/Device Name: E-CUBE 5

Regulation Number: 21 CFR 892.1550

Regulation Name: Ultrasonic pulsed doppler imaging system

Regulatory Class: II

Product Code: IYN, IYO, ITX

Dated: September 26, 2014

Received: September 29, 2014

Dear Mr. Kim:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you; however, that device labeling must be truthful and not misleading.

This determination of substantial equivalence applies to the following transducers intended for use with the E-CUBE 5 Diagnostic Ultrasound System, as described in your premarket notification:

Transducer Model Number

C1-6T

L3-12T

EV3-10T

EC3-10T

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be

found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638 2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,



for

Janine M. Morris  
Director  
Division of Radiological Health  
Office of In Vitro Diagnostics  
and Radiological Health  
Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)

K142733

Device Name

E-CUBE 5 Diagnostic Ultrasound System

Indications for Use (Describe)

The device is intended for use by a qualified physician for the evaluation of soft tissue and blood flow in the clinical applications; Fetal; Abdominal (renal & GYN/pelvic); Pediatric; Small Organ(breast, testes, thyroid); Trans-rectal(TR); Trans-vaginal(TV); Musculo-skeletal(Conventional); Musculo-skeletal (Superficial); Peripheral Vascular (PV); and Urology (including prostate).

Type of Use (Select one or both, as applicable)

☒ Prescription Use (Part 21 CFR 801 Subpart D)

☐ Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

## Diagnostic Ultrasound Indications for Use

**E-CUBE 5 Ultrasound System**

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation								
	B	M	PWD	CWD	Color Doppler	Power Doppler	Tissue Harmonic Imaging	Combined* (Specify)	Other** (Specify)
Ophthalmic									
Fetal	N	N	N		N	N	N	N	
Abdominal	N	N	N		N	N	N	N	
Intra-operative (Specify)									
Intra-operative (Neuro)									
Laparoscopic									
Pediatric	N	N	N		N	N	N	N	
Small Organ (breast, testes, thyroid)	N	N	N		N	N	N	N	
Neonatal Cephalic									
Adult Cephalic									
Trans-rectal	N	N	N		N	N	N	N	
Trans-vaginal	N	N	N		N	N	N	N	
Trans-urethral									
Trans-esoph. (non-Card.)									
Musculo-skeletal ( <b>Conventional</b> )	N	N	N		N	N	N	N	
Musculo-skeletal ( <b>Superficial</b> )	N	N	N		N	N	N	N	
Intravascular									
Cardiac Adult									
Cardiac Pediatric									
Intravascular (Cardiac)									
Trans-esoph. (Cardiac)									
Intra-cardiac									
Peripheral vessel	N	N	N		N	N	N	N	
Urology (including prostate)	N	N	N		N	N	N	N	

N = new indication; P = previously cleared by FDA ; E = added under appendix

\* Combined: B/Color Doppler, B/PWD, B/Color Doppler/PWD; \*\*Other: 3D, 4D

**(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)**

Concurrence of CDRH

Prescription User (Per 21 CFR 801.109)

## Diagnostic Ultrasound Indications for Use

**E-CUBE 5 with C1-6T Transducer**

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation								
	B	M	PWD	CWD	Color Doppler	Power Doppler	Tissue Harmonic Imaging	Combined* (Specify)	Other** (Specify)
Ophthalmic									
Fetal	N	N	N		N	N	N	N	
Abdominal	N	N	N		N	N	N	N	
Intra-operative (Specify)									
Intra-operative (Neuro)									
Laparoscopic									
Pediatric	N	N	N		N	N	N	N	
Small Organ (breast, testes, thyroid)									
Neonatal Cephalic									
Adult Cephalic									
Trans-rectal									
Trans-vaginal									
Trans-urethral									
Trans-esoph. (non-Card.)									
Musculo-skeletal ( <i>Conventional</i> )									
Musculo-skeletal ( <i>Superficial</i> )									
Intravascular									
Cardiac Adult									
Cardiac Pediatric									
Intravascular (Cardiac)									
Trans-esoph. (Cardiac)									
Intra-cardiac									
Peripheral vessel									
Urology (including prostate)	N	N	N		N	N	N	N	

N = new indication; P = previously cleared by FDA; E = added under appendix

\* Combined: B/Color Doppler, B/PWD, B/Color Doppler/PWD; \*\*Other: 3D, 4D

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH

Prescription User (Per 21 CFR 801.109)

## Diagnostic Ultrasound Indications for Use

**E-CUBE 5 with L3-12T Transducer**

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation								
	B	M	PWD	CWD	Color Doppler	Power Doppler	Tissue Harmonic Imaging	Combined* (Specify)	Other** (Specify)
Ophthalmic									
Fetal									
Abdominal									
Intra-operative (Specify)									
Intra-operative (Neuro)									
Laparoscopic									
Pediatric	N	N	N		N	N	N	N	
Small Organ (breast, testes, thyroid)	N	N	N		N	N	N	N	
Neonatal Cephalic									
Adult Cephalic									
Trans-rectal									
Trans-vaginal									
Trans-urethral									
Trans-esoph. (non-Card.)									
Musculo-skeletal ( <b>Conventional</b> )	N	N	N		N	N	N	N	
Musculo-skeletal ( <b>Superficial</b> )	N	N	N		N	N	N	N	
Intravascular									
Cardiac Adult									
Cardiac Pediatric									
Intravascular (Cardiac)									
Trans-esoph. (Cardiac)									
Intra-cardiac									
Peripheral vessel	N	N	N		N	N	N	N	
Urology (including prostate)									

N = new indication; P = previously cleared by FDA; E = added under appendix

\* Combined: B/Color Doppler, B/PWD, B/Color Doppler/PWD; \*\*Other: 3D, 4D

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH

Prescription User (Per 21 CFR 801.109)

## Diagnostic Ultrasound Indications for Use

**E-CUBE 5 with EV3-10T Transducer**

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation								
	B	M	PWD	CWD	Color Doppler	Power Doppler	Tissue Harmonic Imaging	Combined* (Specify)	Other** (Specify)
Ophthalmic									
Fetal									
Abdominal									
Intra-operative (Specify)									
Intra-operative (Neuro)									
Laparoscopic									
Pediatric									
Small Organ (breast, testes, thyroid)									
Neonatal Cephalic									
Adult Cephalic									
Trans-rectal	N	N	N		N	N		N	
Trans-vaginal	N	N	N		N	N		N	
Trans-urethral									
Trans-esoph. (non-Card.)									
Musculo-skeletal ( <i>Conventional</i> )									
Musculo-skeletal ( <i>Superficial</i> )									
Intravascular									
Cardiac Adult									
Cardiac Pediatric									
Intravascular (Cardiac)									
Trans-esoph. (Cardiac)									
Intra-cardiac									
Peripheral vessel									
Urology (including prostate)	N	N	N		N	N		N	

N = new indication; P = previously cleared by FDA; E = added under appendix

\* Combined: B/Color Doppler, B/PWD, B/Color Doppler/PWD; \*\*Other: 3D, 4D

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH

Prescription User (Per 21 CFR 801.109)

## Diagnostic Ultrasound Indications for Use

**E-CUBE 5 with EC3-10T Transducer**

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation								
	B	M	PWD	CWD	Color Doppler	Power Doppler	Tissue Harmonic Imaging	Combined* (Specify)	Other** (Specify)
Ophthalmic									
Fetal									
Abdominal									
Intra-operative (Specify)									
Intra-operative (Neuro)									
Laparoscopic									
Pediatric									
Small Organ (breast, testes, thyroid)									
Neonatal Cephalic									
Adult Cephalic									
Trans-rectal	N	N	N		N	N		N	
Trans-vaginal	N	N	N		N	N		N	
Trans-urethral									
Trans-esoph. (non-Card.)									
Musculo-skeletal ( <i>Conventional</i> )									
Musculo-skeletal ( <i>Superficial</i> )									
Intravascular									
Cardiac Adult									
Cardiac Pediatric									
Intravascular (Cardiac)									
Trans-esoph. (Cardiac)									
Intra-cardiac									
Peripheral vessel									
Urology (including prostate)	N	N	N		N	N		N	

N = new indication; P = previously cleared by FDA; E = added under appendix

\* Combined: B/Color Doppler, B/PWD, B/Color Doppler/PWD; \*\*Other: 3D, 4D

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH

Prescription User (Per 21 CFR 801.109)



## Section F 510(k) Summary

In accordance with 21CFR807.92, the following summary of information is provided;

Date Sep 5<sup>th</sup> 2014

Submitter: ALPINION MEDICAL SYSTEMS Co., Ltd.  
Address: 1, 6 and 7FL Verdi Tower, 72, Digital-ro(St) 26-gil(Rd), Guro-gu,  
Seoul, Republic of Korea 152-848

Primary Contact Donghwan Kim  
Person QARA Manager  
Address: 1, 6 and 7FL Verdi Tower, 72, Digital-ro(St) 26-gil(Rd), Guro-gu,  
Seoul, Republic of Korea 152-848  
Phone: +82 70 7465 2068  
Fax: +82 2 851 5594  
Email: donghwan.kim@alpinion.com

Secondary Contact JULIAN LEE  
Person Address: 21312 30th Dr SE Ste 100 Bothell, WA 98021, United States  
Phone: 425 949 1059  
Fax: 425 949 4910  
Email: julian.lee@alpinionusa.com

Device Trade Name: E-CUBE 5

Common/Usual Ultrasonic Pulsed Doppler Imaging System  
Name:  
Classification Names System, Imaging, Pulsed Doppler Ultrasonic

Product Code: Ultrasonic Pulsed Doppler Imaging System, 21CFR 892.1550 90-IYN  
Ultrasonic Pulsed Echo Imaging System, 21CFR 892.1560, 90-IYO  
Diagnostic Ultrasound Transducer, 21CFR 892.1570, 90-ITX

Predicate Device(s) K132687 E-CUBE 7 Diagnostic Ultrasound System

Device Description: E-CUBE 5 product is an ultrasound imaging system for medical diagnosis. The system platform provides optimal patient diagnosis workflow with the 15.6" wide flat panel display, ergonomic control panel with easy user interface, optimal image quality.

Modes of operation:

**1. Signal Mode:**

B(2D) mode, M mode, Color Flow(CF) mode, Power Doppler(PD) mode,  
Pulsed Wave

Doppler(PWD) mode, Tissue Harmonic Imaging(THI)

**2. Combination Mode:**

B/M, B/CF, B/PD, B/PWD, B/CF/PWD, B/PD/PWD, B/CF/M

Acoustic output track:

Track 3

Types of transducers compatible with the device:

	C1-6T	L3-12T	EV3-10T	EC3-10T
<b>Applicable frequency</b>	1~6MHz	3~12MHz	3~10MHz	3~10 MHz
<b>Intended Usage</b>	Fetal, Abdominal Pediatric, Urology	Pediatric, Small Organ, Musculo- skeletal (Conventional), Musculo-skeletal (Superficial), Peripheral vessel	Trans-rectal, Trans-vaginal, Urology	Trans-rectal, Trans-vaginal, Urology
<b>Foot print size (mm)</b>	71.6 x 16.8	44.8 x 7.8	21.5 x 18.6	21.5 x 18.6
<b>Applicable mode</b>	B/M/PWD/ Color Doppler/ Power Doppler/ Tissue Harmonic Imaging	B/M/PWD/ Color Doppler/ Power Doppler/ Tissue Harmonic Imaging	B/M/PWD/ Color Doppler/ Power Doppler Imaging	B/M/PWD/ Color Doppler/ Power Doppler Imaging
<b>Scanning depth(mm)</b>	300	100	100	100
<b>FOV</b>	60(°)	142(°)	142(°)	142(°)
<b>Steer Angle</b>	N/A	Max 9(°)	N/A	N/A
<b>Total number of element</b>	128	128	128	128
<b>Element spacing</b>	0.484mm	0.3mm	0.195mm	0.195mm
<b>elevating length</b>	13.5mm	4.5mm	6.0mm	6.0mm

Indications For Use: The device is intended for use by a qualified physician for the evaluation of soft tissue and blood flow in the clinical applications; Fetal; Abdominal (renal & GYN/pelvic); Pediatric; Small Organ (breast, testes, thyroid); Trans-rectal(TR); Trans-vaginal(TV); Musculo-skeletal(Conventional); Musculo-skeletal (Superficial); Peripheral Vascular (PV); and Urology (including prostate).

Determination of  
Substantial  
Equivalence:

Comparison with Predicate device:

Feature	Proposed E-CUBE 5	Predicate E-CUBE 7 (K132687)
Indications for use	The device is intended for use by a qualified physician for the evaluation of soft tissue and blood flow in the clinical applications; Fetal; Abdominal (renal & GYN/pelvic); Pediatric; Small Organ (breast, testes, thyroid); Trans-rectal(TR); Trans-vaginal(TV); Musculo-skeletal(Conventional); Musculo-skeletal Superficial); Peripheral Vascular (PV); Urology (including prostate).	The device is intended for use by a qualified physician for the evaluation of soft tissue and blood flow in the clinical applications; Fetal; Abdominal (renal & GYN/pelvic); Pediatric; Small Organ (breast, testes, thyroid); Trans-rectal(TR); Trans-vaginal(TV); Musculo-skeletal(Conventional); Musculo-skeletal Superficial); Cardiac (adult & pediatric); Peripheral Vascular (PV); Urology (including prostate).

## 510(k) E-CUBE 5

Transducer	C1-6T  L3-12T  EV3-10T EC3-10T	C1-6 C5-8  L3-12 L3-12H L3-12H <sup>WD</sup> L3-8 L8-17  EV3-10 EC3-10 E3-10 EN3-10  SP1-5 SP3-8  SC1-6 VC1-6  CW2.0 CW5.0  IO3-12
Electrical power	Voltage: 100~120V, 200~240V Frequency: 50/60Hz Power: Max. 450 VA with Built-in and On-Board Peripherals	Voltage: 100~120V, 200~240V Frequency: 50/60Hz Power: Max. 600 VA with Built-in and On-Board Peripherals
Operating Mode	B Mode M Mode Color Flow Mode Power Doppler Mode Pulsed Wave Doppler Mode Tissue Harmonic Imaging Mode  Xpeed™ Full SRI™ Spatial compounding Frequency Compounding Panoramic Auto IMT Measurement	B Mode M Mode Color Flow Mode PW Doppler Mode Power Doppler Mode Continuous wave Doppler mode Tissue Harmonic Imaging Mode 3D/4D Mode  Xpeed™ Full SRI™ Spatial compounding Frequency Compounding Panoramic Auto IMT Measurement
Thermal, mechanical and electrical safety	The E-CUBE 5 has been designed to conform to the following standards: - NEMA UD2, UD3 - AIUM Medical Ultrasound Safety - IEC60601-1 - IEC60601-1-2 - IEC60601-2-37	The E-CUBE 7 has been designed to conform to the following standards: - NEMA UD2, UD3 - AIUM Medical Ultrasound Safety - IEC60601-1 - IEC60601-1-2 - IEC60601-2-37

### Summary of Non-Clinical Tests:

E-CUBE 5 has been evaluated for biocompatibility, acoustic output as well as thermal, electrical, electromagnetic, and mechanical safety, and has been found to conform to applicable medical device safety standards. E-CUBE 5 and its application comply with voluntary standards as detailed in this premarket submission. The following quality management system measures were applied to the development of E-CUBE 5:

- ♦ Medical Device Risk Management
- ♦ Requirements Reviews
- ♦ Design Reviews
- ♦ Component Verification
- ♦ Integration Review (System Verification)
- ♦ Performance Testing (System Verification)
- ♦ Safety Testing (Compliance Test)
- ♦ Design Validation

Transducer materials and other patient contact materials are biocompatible.

Summary of Clinical Tests:

The subject of this premarket submission, E-CUBE 5, did not require clinical studies to support substantial equivalence.

Discussion:

E-CUBE 5 and the predicate device have differences in clinical applications and operating modes. Several transducers are changed for these purposes. These design changes have been verified via non-clinical testing. The subject device is in conformance with applicable safety standards. Therefore, the differences between E-CUBE 5 and the predicate would not affect the safety, effectiveness and essential performance of E-CUBE 5.

Conclusion: ALPINION MEDICAL SYSTEMS Co., Ltd. considers E-CUBE 5 to be as safe, as effective. Performance, technology and software are substantially equivalent to the predicate device.

ALPINION MEDICAL SYSTEMS Co., Ltd. will update and include in this summary any other information deemed reasonably necessary by the FDA or the requirements will be published in guidance documents.