



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center – WO66-G609
Silver Spring, MD 20993-0002

Medartis AG
% Mr. Kevin A. Thomas
PaxMed International, LLC
12264 El Camino Real, Suite 400
San Diego, California 92130

June 9, 2015

Re: K142906
Trade/Device Name: APTUS[®] Wrist 2.5 System
Regulation Number: 21 CFR 888.3030
Regulation Name: Single/multiple component metallic bone fixation appliances and accessories
Regulatory Class: Class II
Product Code: HRS, HWC
Dated: May 7, 2015
Received: May 8, 2015

Dear Mr. Thomas:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set

forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

Mark N. Melkerson -S

Mark N. Melkerson
Director
Division of Orthopedic Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K142906

Device Name

APTUS® Wrist 2.5 System

Indications for Use (Describe)

APTUS® Wrist 2.5 System is intended for use in hand and forearm fractures, osteotomies and arthrodeses.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) Summary

APTUS® Wrist 2.5 System

510(k) Summary**Medartis AG****APTUS® Wrist 2.5 System**

June 2, 2015

ADMINISTRATIVE INFORMATION

Manufacturer Name	Medartis AG Hochbergerstrasse 60E CH-4057 Basel, Switzerland Telephone: +41 61 633 34 34 Fax: +41 61 633 34 00
Official Contact	Andrea Schweizer Head of Quality Management, Medartis AG
Representative/Consultant	Kevin A. Thomas, PhD Floyd G. Larson PaxMed International, LLC 12264 El Camino Real, Suite 400 San Diego, CA 92130 Telephone: +1 (858) 792-1235 Fax: +1 (858) 792-1236 Email: kthomas@paxmed.com flarson@paxmed.com

DEVICE NAME AND CLASSIFICATION

Trade/Proprietary Name	APTUS® Wrist 2.5 System
Common Name	Plate, fixation, bone Screw, fixation, bone
Classification Names	Single/multiple component metallic bone fixation appliances and accessories Smooth or threaded metallic bone fixation fastener
Classification Regulations	21 CFR 888.3030, 21 CFR 888.3040, Class II
Product Codes	HRS, HWC
Classification Panel	Orthopedic Products Panel
Reviewing Branch	Joint Fixation Devices Branch Two (JFDB2)

510(k) Summary

APTUS[®] Wrist 2.5 System

INTENDED USE

APTUS[®] Wrist 2.5 System is intended for use in hand and forearm fractures, osteotomies and arthrodeses.

DEVICE DESCRIPTION

APTUS Wrist 2.5 System consists of titanium locking plates. The plates are secured using previously cleared titanium screws (K103332 and K051567), both locking and non-locking. This submission includes additional lengths of non-locking screws. Plates have a low overall height, rounded edges, polished surfaces and incorporate TriLock Technology with use of TriLock (locking) screws. All plates are made from unalloyed titanium conforming to ASTM F67, and all screws are made from titanium alloy conforming to ASTM F136.

The subject device is substantially equivalent in indications and design principles to the following predicate devices:

K051567, APTUS[®] Titanium Fixation System, Medartis, Inc.

K103332, APTUS[®] Ulna Plates, Medartis AG

K040022, Stryker[®] Leibinger Universal Distal Radius System, Stryker Leibinger

K070946, Synthes LCP Diaphyseal-Metaphyseal (Dia-Meta) Volar Distal Radius Plate, Synthes (USA)

The subject device and the predicate devices have the same intended use and have the same technological characteristics. The subject and predicate devices are all fabricated from the same or similar materials and share similar design characteristics, including plate screw holes to accommodate locking and non-locking screws. The subject and predicate devices encompass the same range of physical dimensions, and the subject device is compatible with screws from the predicate devices K103332 and K051567. The subject and predicate devices are packaged using the same materials, and are to be sterilized by the same methods. Any differences in the technological characteristics do not raise new issues of safety or efficacy.

Performance data provided to demonstrate substantial equivalence included detailed dimensional analysis and engineering analysis of the subject and predicate device designs. Mechanical testing conducted on the subject screws in comparison with the predicate screws demonstrated substantial equivalence in screw performance.

Overall, APTUS[®] Wrist 2.5 System has the following similarities to the predicate devices:

- has the same intended use,
- uses the same operating principle,
- incorporates the same basic design,
- incorporates the same or very similar materials, and
- has similar packaging and is sterilized using the same materials and processes.