

Food and Drug Administration 10903 New Hampshire Avenue Document Control Center - WO66-G609 Silver Spring, MD 20993-0002

August 21, 2015

TCM Associates Ltd Mr. Iain Alligan Technical Director 3 Hillgrove Business Park Nazeing Road Essex EN9 2HB United Kingdom

Re: K143692

Trade/Device Name: GranOS - beta tricalcium phosphate synthetic bone granules Regulation Number: 21 CFR 888.3045 Regulation Name: Resorbable Calcium Salt Bone Void Filler Device Regulatory Class: Class II Product Code: MQV Dated: July 13, 2015 Received: July 23, 2015

Dear Mr. Alligan:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the <u>Federal Register</u>.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in

the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<u>http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm</u> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm.

Sincerely yours,

## Mark N. Melkerson -S

Mark N. Melkerson Director Division of Orthopedic Devices Office of Device Evaluation Center for Devices and Radiological Health

Enclosure

#### DEPARTMENT OF HEALTH AND HUMAN SERVICES Food and Drug Administration

### Indications for Use

Form Approved: OMB No. 0910-0120 Expiration Date: January 31, 2017 See PRA Statement below.

510(k) Number (*if known*) K143692

### Device Name

GranOSTM - beta tricalcium phosphate synthetic bone granules

#### Indications for Use (Describe)

GranOS<sup>™</sup> devices are indicated only for bony voids or gaps that are not intrinsic to the stability of the bony structure. GranOS<sup>™</sup> is indicated to be gently placed into bony voids or gaps of the skeletal system (ie the extremities, posterolateral spine and pelvis). These defects may be surgically created osseous defects or osseous defects created from traumatic injury to the bone. The product provides a bone void filler that resorbs and is replaced with bone during the healing process.

Type of Use (Select one or both, as applicable)	
Prescription Use (Part 21 CFR 801 Subpart D)	Over-The-Counter Use (21 CFR 801 Subpart C)

#### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

#### \*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\*

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services Food and Drug Administration Office of Chief Information Officer Paperwork Reduction Act (PRA) Staff *PRAStaff@fda.hhs.gov* 

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

## Section 5

## Premarket Notification 510(k) Summary

## beta tricalcium phosphate synthetic bone granules (Trade Name: GranOS™)

510k (Tradition	al) Premarket Submission	TCM Associates Ltd
Device:	GranOS <sup>™</sup> - beta tricalcium phosphate synthetic bone granules	Issue: 03
Classification:	Resorbable calcium salt bone void filler device	Issue Date: 20.08.2015
Page Title:	Section 5: Premarket Notification 510(k) Summary	

# 5.1Premarket Notification 510(k) Summary

GranOS<sup>™</sup> - beta tricalcium phosphate synthetic bone granules (per 21 CFR 807.92)

Submitter (Owner) TCM Associates Ltd 3 Hillgrove Business Park Nazeing Road Nazeing Essex EN9 2HB United Kingdom **Contact Person** lain Alligan , Technical Director Phone: +44 (0) 1992 892 085 Fax: +44 (0) 1992 893 721 Date prepared 20th August 2015 Trade Name GranOS<sup>™</sup> - beta tricalcium phosphate synthetic bone granules Common Name Synthetic bone void filler **Device Classification** Resorbable calcium salt bone void filler device Regulation Class II Special Controls as per 21 CFR 888.3045 **Device Product Code** MQV  $\beta \text{Gran}^{\$}$  Synthetic Osteoconductive Scaffold (K041616) Cerasorb  ${}^{\$}$  ORTHO (K014156) Legally Marketed Predicate Devices Vitoss <sup>®</sup> Scaffold Synthetic Cancellous Bone Void Filler K032409 **Device Description** GranOS<sup>™</sup> is a porous, resorbable osteoconductive scaffold constructed of highly pure beta tricalcium phosphate granules for use in the repair of bony defects. Intended Use GranOS<sup>™</sup> devices are indicated only for bony voids or gaps that are not intrinsic to the stability of the bony structure. GranOS<sup>™</sup> is indicated to be gently placed into bony voids or gaps of the skeletal system (i.e. the extremities, posterolateral spine, and pelvis). These defects may be surgically created osseous defects or osseous defects created from traumatic injury to the bone. The product provides a bone void filler that resorbs and is replaced with bone during the healing process

510k (Tradition	nal) Premarket Submission	TCM Associates Ltd	
Device:	GranOS™ - beta tricalcium phosphate synthetic bone granules	lssue: 03	
Classification:	Resorbable calcium salt bone void filler device	Issue Date: 20.08.2015	
Page Title:	Section 5: Premarket Notification 510(k) Summary		

<b>Technological Characteristics</b>	<ul> <li>Comparison to Predicates a</li> </ul>	and Substantial Equivalence
--------------------------------------	--	-----------------------------

Trade Name	GranOS™ - beta tricalcium phosphate synthetic bone granules	βGran <sup>®</sup> synthetic osteoconductive scaffold	Cerasorb <sup>®</sup> ORTHO	Vitoss <sup>®</sup> Scaffold Synthetic Cancellous Bone void Filler
510(k)	K143692	K041616	К014156	K032409
Registration No:	la sta tai sa lai wa sa basa bata			
chemical	$C_{2}$ (PQ) monting the	Ca (PQ) maating the	$C_{2}$ (PQ) mosting the	Co (DO) mosting the
matorial	requirements of		requirements of ASTM E	requirements of
material	ASTM = 1088-042 (2010)	ASTM = 1088-042 (2010)	1088-042 (2010)	ASTM = 1088-042 (2010)
Patient Population	Individuals with hony	Individuals with bony defects	Patients with hone voids	Individuals with hony
i diciti i opulation	defects resulting from	resulting from surgery or	or gaps caused by surgery	defects resulting from
	surgery or trauma	trauma	trauma or degeneration	surgery or trauma
Anatomical	To be packed into the	Bony voids or gaps of the	Skeletal system	Bony voids or gaps of the
Locations	irregular shaped bony voids	skeletal system i.e. the	extremities, spine and	skeletal system i.e. the
	of the skeletal system i.e.	extremities, spine and pelvis	pelvis	extremities, spine and
	extremities, spine and pelvis			pelvis
Physical Structure	Interconnected porosity	Interconnected porosity	Interconnected porosity	Trabecular structure
of the Material				similar to cancellous
				bone
Porosity	Approximately 70%	Approximately 70%	Approximately 60% - 70%	Approximately 90%
Pore Size (range)	Micropores <1 - 700µm	Micropores <1 - 700µm	Micropores >0<80 μm	1-1000 μm
Performance				
Ostoosondustivity	Ostaasandustiva	Ostaasandustiva	Octooconductivo	Octooconductivo
Osteoconductivity	Osteoconductive	Osteoconductive	Osteoconductive	Osteoconductive
Sterility	Sterilised by Gamma	Sterilised by Gamma	Sterilised by Gamma	Sterilised by Gamma
	Radiation, single use only.	Radiation, single use only.	Radiation, single use only.	Radiation, single use
	Sterility Assurance Level	Sterility Assurance Level	Sterility Assurance Level	only. Sterility Assurance
	(SAL) of 1 x 10 <sup>-6</sup>	(SAL) of 1 x 10 <sup>-6</sup>	(SAL) of 1 x 10 <sup>-6</sup>	Level (SAL) of $1 \times 10^{-6}$
Biocompatibility	Established	Established	Established	Established
Presentation	Granule Sizes	Granule Sizes	Granules and Blocks	Morsels 1-4mm and
	250-500μm	250-500μm	500 - 1000μm	cylinders 9 x 23mm
	1 – 2.8mm	1 – 2.8mm	1000 - 2000μm	
	2 – 4mm	2 – 4mm		

Results of non-clinical performance tests has demonstrated that GranOS<sup>™</sup> - beta tricalcium phosphate granules have the same critical specification (i.e. chemistry, crystallinity, physical form, porosity, dissolution/solubility) and the same intended use as the predicate device.

Key requirements and recommendations have also been met within The Class II Special Controls Guidance: Document Resorbable Calcium Salt Bone Void Filler Device; Guidance for Industry and FDA.

### Non-clinical testing

GranOS<sup>™</sup> - beta tricalcium phosphate granules have been tested and shown to be in compliance to the following recognised performance standard: ASTM F1088-04a 2010 - Standard Specification for beta tricalcium phosphate for Surgical Implantation. In addition, GranOS<sup>™</sup> - beta tricalcium phosphate granules have the same Contraindications, Instructions, Warnings, Precautions and possible adverse reactions/complications when directly compared to the predicate devices and so does not raise new questions regarding safety and effectiveness and therefore substantial equivalence is claimed under Section 513(i) of the FD&C Act (21 U.S.C. § 360c(i)).

### [21CFR 807.92(b)(1) and 21CFR 807.92(b)(2])

### Conclusion

Based on the results of the non-clinical testing and the data gathered from the clinical evaluation of the product, we conclude that GranOS<sup>™</sup> - beta tricalcium phosphate granules are as safe, as effective, and performs as well as or better than the predicate device. [21CFR 807.92(b)(3)