



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center - WO66-G609
Silver Spring, MD 20993-0002

May 2, 2015

Grandway Technology (Shenzhen) Limited
Patrick Chow
Block 6 and 7, Zhu Keng Industrial Zone, Ping Shan
Long Gang District
Shenzhen, Guang Dong, 518118 CN

Re: K143734
Trade/Device Name: Digital Automatic Blood Pressure Monitor BPM28 Series -
MD2800, MD2810
Regulation Number: 21 CFR 870.1130
Regulation Name: Noninvasive Blood Pressure Measurement System
Regulatory Class: Class II
Product Code: DXN
Dated: March 27, 2015
Received: March 27, 2015

Dear Patrick Chow,

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply

with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

A handwritten signature in black ink, appearing to read "Bram D. Zuckerman", is written over a faint, large, stylized "FDA" logo.

for Bram D. Zuckerman, M.D.

Director

Division of Cardiovascular Devices

Office of Device Evaluation

Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K143734

K143734

Page 1 of 1

Device Name

Digital Automatic Blood Pressure Monitor BPM28 Series, MD2800, MD2810

Indications for Use (Describe)

The upper arm blood pressure monitor is used for non-invasive measurement and monitoring of adults' arterial blood pressure. You can use it to measure your systolic and diastolic pressure, and pulse rate through an inflatable cuff wrapped around the upper arm. Quickly and easily, storing the results and displaying the progression of readings together with the average.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

1. Submitter Identification

510(k) Submitter	GRANDWAY TECHNOLOGY (SHENZHEN) LIMITED
Address	Block 6 and 7, Zhu Keng Industrial Zone, Ping Shan, Long Gang District, Shenzhen, Guang Dong, People's Republic of China
Phone Number	(00852)-2851-6789
Fax Number	(00852)-2851-6278
Contact Person	Mr. Patrick Chow
Date of Submission	30-Dec-2014

2. Device Identification

Trade Name	Digital Automatic Blood Pressure Monitor BPM28 Series [Model No.: MD28x0] x --- The first character (0 and 1) is for the identification of add-onfeatures.
Common Name	Non-invasive Blood Pressure Measurement System
Classification Name	Non-invasive Blood Pressure Measurement System (CFR 870.1130, Class II, Product Code DXN)

3. Predicate Device

Predicate Device	Digital Automatic Blood Pressure Monitor BPM18 Series
Manufacturer	GRANDWAY TECHNOLOGY (SHENZHEN) LIMITED
510(k) Number	K133619

4. Device Description

Digital Automatic Blood Pressure Monitor BPM28 Series is a non-invasive blood pressure measurement system for use by medical professional or at home. It is designed to measure the systolic and diastolic blood pressure, and pulse rate of an individual in each measurement and then display the readings on a digital panel.

The device utilizes the oscillometric methodology, in which an inflatable cuff is wrapped around the [upper arm](#) of an individual, for blood pressure measurement. This means the monitor detects your blood's movement through your brachial artery and converts the movement into a digital reading.

The table below illustrate the feature presence in Digital Automatic Blood Pressure Monitor BPM28 Series.

Model	BP Measurement	Pulse Rate Measurement	WHO	IHB	LCD Type	Key Type	DC Jack	Key Ring Illustration
MD2800	✓	✓	✓	✓	Positive Reflective	Hard	✓	✓
MD2810	✓	✓	✓	✓	Positive Reflective	Soft	✗	✗

5. Indication for Use

The upper arm blood pressure monitor is used for non-invasive measurement and monitoring of adults' arterial blood pressure. You can use it to measure your systolic and diastolic pressure, and pulse rate through an inflatable cuff wrapped around the upper arm. Quickly and easily, storing the results and displaying the progression of readings together with the average.

6. Comparison of Technological Characteristics between New Device and Predicate Devices

Digital Automatic Blood Pressure Monitor BPM28 Series is compared to the predicate device, [BPM18 Series \(K133619\)](#) in the device comparison table below.

Comparison between BPM28 Series and Predicate device			
Item	Predicate Device	BPM28 Series	Comment
Indication for Use	Digital Automatic Blood Pressure Monitor BPM18 Series is for use by medical professional or home user. The BPM18 Series is intended to measure the systolic and diastolic blood pressure, and pulse rate of an adult individual by using a non-invasive technique, in which an inflatable cuff is wrapped around the upper arm.	The upper arm blood pressure monitor is used for non-invasive measurement and monitoring of adults' arterial blood pressure. You can use it to measure your systolic and diastolic pressure, and pulse rate through an inflatable cuff wrapped around the upper arm. Quickly and easily, storing the results and displaying the progression of readings together with the average.	Equivalent
Measurement Method	Non-invasive, Oscillometric	Non-invasive, Oscillometric	Identical
IHB Detection	Yes	Yes	Identical
Patient Population	Age 16 or above	Age 16 or above	Identical
BP Measurement Range	Cuff Pressure: 0 - 300 mmHg Systolic Pressure: 50 - 250 mmHg Diastolic Pressure: 30 - 200 mmHg	Cuff Pressure: 0 - 300 mmHg Systolic Pressure: 50 - 250 mmHg Diastolic Pressure: 30 - 200 mmHg	Identical
Number of User	2 independent users	2 independent users	Identical
Memory Space	2 users × 120 memory space	2 users × 60 memory space	Equivalent
Resolution of Measurement	Blood Pressure: 1 mmHg or 0.1kPa Pulse Rate: 1 beat/ min	Blood Pressure: 1 mmHg Pulse Rate: 1 beat/ min	Identical
Blood Pressure Measurement Accuracy	± 3 mmHg or 2% of reading	± 3 mmHg	Equivalent
Pulse Rate Measurement Range	30 - 180 beats/min	40 - 180 beats/min	Equivalent
Pulse Rate Measurement Accuracy	± 5 % of the reading	± 5 % of the reading	Identical
Display Type	LCD	LCD	Identical
Power Source	4 × 1.5 V AAA-batteries; and/or AC adaptor (6V/600mA)	4 × 1.5 V AAA-batteries; and/or AC adaptor (6V/600mA)	Identical
Pressurization Mode	Automatic Inflation	Automatic Inflation	Identical
Deflation Mode	Automatic Exhaust/ Deflation	Automatic Exhaust/ Deflation	Identical
Operating Condition	Temperature: 10 - 40 °C Humidity: 30 - 85 % R.H. max Atmospheric Pressure: 700 - 1060 kPa	Temperature: +5 to +40 °C Humidity: 15 to 93 % R.H. max Atmospheric Pressure: 700 - 1060 hPa	Equivalent

Comparison between BPM28 Series and Predicate device			
Item	Predicate Device	BPM28 Series	Comment
Storage and Transportation Condition	Temperature: -20 - 60 °C Humidity: 10 - 95 % R.H. max Atmospheric Pressure: 700 - 1060 kPa	Temperature: -25 to +70 °C Humidity: up to 93% R.H. max Atmospheric Pressure: 700 - 1060 hPa	Equivalent
Material	Resistances, capacitance, transistors, amplifiers, pressure sensor, CPU, PCB, cuff ABS button, ABS cabinet, batteries and packaging	Resistances, capacitance, transistors, amplifiers, pressure sensor, CPU, PCB, cuff ABS button, ABS cabinet, batteries and packaging	Identical
Compatibility with Environment and Other Devices	No influence with environment and other device	No influence with environment and other device	Identical
Applicable Standard	- EN 1060-1:1995+A2:2009 - EN 1060-3:1997+A2:2009 - IEC 60601-1:2012 - EN 60601-1-2:2007 - FCC Part 15 Subpart B - ISO 10993-5:2009 - ISO 10993-10:2010 - IEC 62304:2006 - IEC 81060-2:2009	- EN 1060-1:1995+A2:2009 - EN 1060-3:1997+A2:2009 - IEC 60601-1:2012 - EN 60601-1-2:2007 - FCC Part 15 Subpart B - ISO 10993-5:2009 - ISO 10993-10:2010 - IEC 62304:2006 - IEC 81060-2:2013	Equivalent

Digital Automatic Blood Pressure Monitor BPM28 Series is a non-invasive measuring device and utilize the oscillometric methodology to measure the blood pressure reading. The key components of device are: a pressure sensor, an electric valve and an electronic control module together with an electric pump. The electric pump inflate (and deflate) the inflatable cuff automatically according to our designed architecture. The predicate device adopts exactly same methodology and key components for measuring blood pressure.

7. Clinical and Non-clinical Tests

Clinical Test Summary

Testing to insure clinical accuracy of the device in accordance with [ISO 81060-2:2013](#) is documented in Clinical Test report.

One hundred patients ([46 males and 54 females](#)) were invited for the study. Standard auscultation method was used as the reference blood pressure monitor measuring in the left upper arm. Blood pressure measurements were repeated alternatively with the device and auscultation in the same arm according to the sequence in [ISO 81060-2:2013](#).

Non-Clinical Test Summary

Digital Automatic Blood Pressure Monitor BPM28 Series has performed several non-clinical tests to show that all requirement specifications and standard requirements are met. The tests includes the follows:

- ✧ EN 1060-1:1995+A2:2009
- ✧ EN 1060-3:1997+A2:2009
- ✧ IEC 60601-1:2012
- ✧ EN 60601-1-2:2007
- ✧ FCC Part 15 Subpart B
- ✧ ISO 10993-5:2009

- ✧ ISO 10993-10:2010
- ✧ IEC 62304:2006

As all of the clinical and non-clinical testing performed on Digital Automatic Blood Pressure Monitor BPM28 Series are as same as the predicate device.

Also, bench testing, IEC 80601-2-30, is conducted to show the performance of Digital Automatic Blood Pressure Monitor BPM28 Series is equivalent to the predicate device.

8. Conclusion

Digital Automatic Blood Pressure Monitor BPM28 Series has the same intended use and same technological characteristics as the predicate device, [BPM18 Series \(K133619\)](#). Moreover both clinical and non-clinical testing has demonstrated that no differences in the technological characteristics and questioning on safety or effectiveness to be raised. Thus, Digital Automatic Blood Pressure Monitor BPM28 Series is substantially equivalent to the predicate device.