



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center - WO66-G609
Silver Spring, MD 20993-0002

April 14, 2015

Roxwood Medical, Inc.
Grace Li
Director, QA
400 Seaport Ct, Suite #103
Redwood City, California 94063

Re: K143744
Trade/Device Name: MicroCross Catheter
Regulation Number: 21 CFR 870.1250
Regulation Name: Percutaneous Catheter
Regulatory Class: Class II
Product Code: DQY
Dated: March 4, 2015
Received: March 6, 2015

Dear Grace Li,

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the [Federal Register](#).

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in

the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

Bram D. Zuckerman -S

Bram D. Zuckerman, M.D.
Director
Division of Cardiovascular Devices
Office of Device Evaluation
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

Device Name

MicroCross Catheter

Indications for Use (Describe)

The MicroCross Catheter is intended for use as a conduit for the exchange/support of guidewires in the peripheral and coronary vasculatures. The MicroCross Catheter is also intended to infuse and deliver saline and contrast agents.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) SUMMARY

510(k) Notification K143744

GENERAL INFORMATION

Applicant:

Roxwood Medical, Inc.
400 Seaport Ct., Suite #103
Redwood City, CA 94063
Phone: (650) 779-4555
FAX: (650) 779-4554

Contact Person:

Grace Li
Roxwood Medical, Inc.
400 Seaport Ct., Suite #103
Redwood City, CA 94063
Phone: (650) 779-4555
FAX: (650) 779-4554

Date Prepared: December 30, 2014

DEVICE INFORMATION

The MicroCross Catheter is a percutaneous catheter for use in the coronary and peripheral vasculature.

Trade Name:

MicroCross Catheter

Generic/Common Name:

Percutaneous Catheter

Classification:

21 CFR§870.1250, Class II

Product Code:

DQY

PREDICATE DEVICE(S)

Valet Micro Catheter (Volcano Corporation, K112035)

INDICATIONS FOR USE

The MicroCross Catheter is intended for use as a conduit for the exchange/support of guidewires in the peripheral and coronary vasculatures. The MicroCross Catheter is also intended to infuse and deliver saline and contrast agents.

PRODUCT DESCRIPTION

The MicroCross Catheter is a sterile, single-use, single lumen over-the-wire catheter to be used for the exchange and support of guidewires in the peripheral and coronary vasculatures. The MicroCross Catheter also infuses and delivers saline and contrast agents.

The MicroCross Catheter consists of a catheter shaft and a proximal hub that provides strain relief. The MicroCross Catheter is offered in working lengths of 150cm and 155cm and is compatible with 0.014” diameter guidewires, 5F guide catheters, and 4F introducer sheaths.

Subsequent to conventional guidewire placement, interventional tools such as angioplasty, stent, and atherectomy devices, may be used to provide therapeutic benefit. The MicroCross Catheter in and of itself does not provide therapeutic benefit beyond simple facilitation of guidewire support. The MicroCross Catheter is similar in its design and it achieves its intended use by means of the same mechanisms as the predicate device.

TECHNOLOGICAL CHARACTERISTICS

The technological characteristics of the MicroCross Catheter are similar to the predicate device. Performance data is provided to support the determination of substantial equivalence.

SUBSTANTIAL EQUIVALENCE

The MicroCross Catheter is substantially equivalent to the Valet Micro Catheter (Volcano Corporation, K112035). The subject device and the predicate devices are percutaneous catheters. The proposed indications for use for the MicroCross Catheter are substantially equivalent to the indications for use for the predicate device. Any differences in the technological characteristics between the devices do not raise any new issues of safety or effectiveness. Thus, the MicroCross Catheter is substantially equivalent to the predicate device.

TESTING IN SUPPORT OF SUBSTANTIAL EQUIVALENCE DETERMINATION

Extensive bench testing was conducted on the MicroCross Catheter to evaluate the performance of the device and to support a determination of substantial equivalence to the predicate device. Non-clinical testing assessed the following aspects of the device:

Nonclinical Testing Summary:

- Dimensional Inspection
- Hydrophilic Coating
- Simulated Use
- Contrast Flush Injection
- Leak
- Torque Strength
- Kink Resistance & Flexibility
- Pressurized Flow Rate
- Tensile Strength
- Luer Compliance
- Corrosion
- Particulate
- LAL
- Biocompatibility
- Sterilization adoption into a validated sterilization process
- Packaging and shelf-life

All testing was performed in accordance with recognized standards. The collective results of the non-clinical testing demonstrate that the MicroCross Catheter meets the established specifications necessary for consistent performance for its intended use and is substantially equivalent to the predicate device.

CONCLUSION

The MicroCross Catheter is a percutaneous vascular catheter and shares its design and mechanism of action with the identified predicate device. The results of the performance testing confirm that the MicroCross Catheter functions to its specifications and intended use and exhibit the appropriate characteristics of a percutaneous vascular catheter. The MicroCross Catheter is substantially equivalent to the predicate device in terms of technological characteristics, intended use and performance. No new issues of safety or effectiveness are raised by the MicroCross Catheter.

SUMMARY

The MicroCross Catheter is substantially equivalent to the predicate device.