



Food and Drug Administration  
10903 New Hampshire Avenue  
Document Control Center - WO66-G609  
Silver Spring, MD 20993-0002

July 21, 2015

Boston Scientific Corporation  
Ms. Renee Mitchell  
Principal Regulatory Affairs Specialist  
One Scimed Place  
Maple Grove, MN 55311

Re: P140028

Trade/Device Name: Innova™ Vascular Self-Expanding Stent System

Filed: December 10, 2014

Amended: January 29, 2015, March 9, 2015, April 16, 2015

Product Code: NIP

Dear Ms. Mitchell:

The Center for Devices and Radiological Health (CDRH) of the Food and Drug Administration (FDA) has completed its review of your premarket approval application (PMA) for the Innova Vascular Self-Expanding Stent System. This device is indicated for: The Innova™ Vascular Self-Expanding Stent System is indicated to improve luminal diameter in the treatment of symptomatic de-novo or restenotic lesions in the native superficial femoral artery (SFA) and/or proximal popliteal artery (PPA) with reference vessel diameters from 4.0 mm to 7.0 mm and lesion lengths up to 190 mm. We are pleased to inform you that the PMA is approved. You may begin commercial distribution of the device in accordance with the conditions of approval described below.

The sale and distribution of this device are restricted to prescription use in accordance with 21 CFR 801.109 and under section 515(d)(1)(B)(ii) of the Federal Food, Drug, and Cosmetic Act (the act). The device is further restricted under section 515(d)(1)(B)(ii) of the act insofar as the labeling must specify the specific training or experience practitioners need in order to use the device. FDA has determined that these restrictions on sale and distribution are necessary to provide reasonable assurance of the safety and effectiveness of the device. Your device is therefore a restricted device subject to the requirements in sections 502(q) and (r) of the act, in addition to the many other FDA requirements governing the manufacture, distribution, and marketing of devices.

Expiration dating for this device has been established and approved at 6 months. This is to advise you that the protocol you used to establish this expiration dating is considered an approved protocol for the purpose of extending the expiration dating as provided by 21 CFR 814.39(a)(7).

Continued approval of this PMA is contingent upon the submission of periodic reports, required under 21 CFR 814.84, at intervals of one year (unless otherwise specified) from the date of approval of the original PMA. Two copies of this report, identified as "Annual Report" and bearing the applicable PMA reference number, should be submitted to the address below. The Annual Report should indicate the beginning and ending date of the period covered by the report and should include the information required by 21 CFR 814.84. This is a reminder that as of September 24, 2014, class III devices are subject to certain provisions of the final UDI rule. These provisions include the requirement to provide a UDI on the device label and packages (21 CFR 801.20), format dates on the device label in accordance with 21 CFR 801.18, and submit data to the Global Unique Device Identification Database (GUDID) (21 CFR 830 Subpart E). Additionally, 21 CFR 814.84 (b)(4) requires PMA annual reports submitted after September 24, 2014, to identify each device identifier currently in use for the subject device, and the device identifiers for devices that have been discontinued since the previous periodic report. It is not necessary to identify any device identifier discontinued prior to December 23, 2013. For more information on these requirements, please see the UDI website, <http://www.fda.gov/udi>.

In addition to the above, and in order to provide continued reasonable assurance of the safety and effectiveness of the device, the Annual Report must include, separately for each model number (if applicable), the number of devices sold and distributed during the reporting period, including those distributed to distributors. The distribution data will serve as a denominator and provide necessary context for FDA to ascertain the frequency and prevalence of adverse events, as FDA evaluates the continued safety and effectiveness of the device.

In addition to the Annual Report requirements, you must provide the following data in post-approval study (PAS) reports for each PAS listed below. Two (2) copies of each report, identified as an "ODE Lead PMA Post-Approval Study Report" in accordance with how the study is identified below and bearing the applicable PMA reference number, should be submitted to the address below.

ODE Lead PMA Post-Approval Study – *SuperNOVA Study*: The Office of Device Evaluation (ODE) will have the lead for this clinical study, which was initiated prior to device approval. The *SuperNOVA study* is described below:

*SuperNOVA Continued Follow-Up Study*: This study must be conducted per Protocol 90875294, Version AE, dated July 12, 2013, and Statistical Analysis Plan 90875301, Version AB, dated July 2, 2014. This study is a multi-center, single arm, prospective continued follow-up of the SuperNOVA global pivotal study. It will evaluate the long-term safety and effectiveness of the Innova stent. All 256 remaining patients (10 patients exited due to death) of the 299 SuperNOVA global pivotal study enrolled from 49 investigational sites will be followed annually through 36 months post-procedure (60 months for subjects enrolled in Japan) with expected no more than 20% attrition, not including attrition due to death.

The safety and effectiveness endpoints to be assessed through 36 months (60 months for subjects enrolled in Japan) post-procedure are: (1) the composite Major Adverse Event (MAE) Rate and its individual components; (2) primary and assisted primary patency; (3) stent fracture rate; and (4) other efficacy rates as defined in the protocol.

## A. Primary Safety Endpoint -

- The safety endpoint assesses a composite MAE rate and will be reported at 2 and 3 years post-procedure. The MAEs considered and adjudicated by the Clinical Events committee are defined as:
  - Death
  - Target limb major amputation
  - Target Lesion Revascularization
  - Target Vessel Revascularization
  - Stent Thrombosis

## B. Co-Primary Efficacy Endpoints -

- The efficacy endpoints assess vessel primary patency as determined by Duplex Ultrasound (DUS). Primary patency will be reported at 2 years post-procedure for the core stent matrix, the long stents and the entire matrix.
  - Vessel primary patency is defined as freedom from more than 50% stenosis based on DUS peak systolic velocity ratio comparing data within the target segment to the proximal normal arterial segment in the absence of Target Lesion Revascularization or bypass.
  - The DUS will be conducted at each site per protocol requirements and sent to an independent core laboratory for analysis.
  - A systolic velocity ratio >2.4 suggests >50% stenosis.

Be advised that the failure to conduct any such study in compliance with the good clinical laboratory practices in 21 CFR part 58 (if a non-clinical study subject to part 58) or the institutional review board regulations in 21 CFR part 56 and the informed consent regulations in 21 CFR part 50 (if a clinical study involving human subjects) may be grounds for FDA withdrawal of approval of the PMA. In addition, the results from any post approval study should be included in the labeling as these data become available. Any updated labeling must be submitted to FDA in the form of a PMA Supplement. For more information on post-approval studies, see the FDA guidance document entitled, "Procedures for Handling Post-Approval Studies Imposed by PMA Order"

<http://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/ucm070974.htm>).

Within 30 days of your receipt of this letter, you must submit a PMA supplement that includes a complete protocol of your post-approval study described above. Your PMA supplement should be clearly labeled as an "ODE Lead PMA Post-Approval Study Report" as noted above and submitted in triplicate to the address below. Please reference the PMA number above to facilitate processing. If there are multiple protocols being finalized after PMA approval, please submit each protocol as a separate PMA supplement.

Before making any change affecting the safety or effectiveness of the device, you must submit a PMA supplement or an alternate submission (30-day notice) in accordance with 21 CFR 814.39. All PMA supplements and alternate submissions (30-day notice) must comply with the applicable requirements in 21 CFR 814.39. For more information, please refer to the FDA

guidance document entitled, "Modifications to Devices Subject to Premarket Approval (PMA) - The PMA Supplement Decision-Making Process"

<http://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/ucm089274.htm>

You are reminded that many FDA requirements govern the manufacture, distribution, and marketing of devices. For example, in accordance with the Medical Device Reporting (MDR) regulation, 21 CFR 803.50 and 21 CFR 803.52, you are required to report adverse events for this device. Manufacturers of medical devices, including in vitro diagnostic devices, are required to report to FDA no later than 30 calendar days after the day they receive or otherwise becomes aware of information, from any source, that reasonably suggests that one of their marketed devices:

1. May have caused or contributed to a death or serious injury; or
2. Has malfunctioned and such device or similar device marketed by the manufacturer would be likely to cause or contribute to a death or serious injury if the malfunction were to recur.

Additional information on MDR, including how, when, and where to report, is available at <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm>

In accordance with the recall requirements specified in 21 CFR 806.10, you are required to submit a written report to FDA of any correction or removal of this device initiated by you to: (1) reduce a risk to health posed by the device; or (2) remedy a violation of the act caused by the device which may present a risk to health, with certain exceptions specified in 21 CFR 806.10(a)(2). Additional information on recalls is available at <http://www.fda.gov/Safety/Recalls/IndustryGuidance/default.htm>

CDRH does not evaluate information related to contract liability warranties. We remind you; however, that device labeling must be truthful and not misleading. CDRH will notify the public of its decision to approve your PMA by making available, among other information, a summary of the safety and effectiveness data upon which the approval is based. The information can be found on the FDA CDRH Internet HomePage located at <http://www.fda.gov/MedicalDevices/ProductsandMedicalProcedures/DeviceApprovalsandClearances/PMAApprovals/default.htm>. Written requests for this information can also be made to the Food and Drug Administration, Dockets Management Branch, (HFA-305), 5630 Fishers Lane, Rm. 1061, Rockville, MD 20852. The written request should include the PMA number or docket number. Within 30 days from the date that this information is placed on the Internet, any interested person may seek review of this decision by submitting a petition for review under section 515(g) of the act and requesting either a hearing or review by an independent advisory committee. FDA may, for good cause, extend this 30-day filing period.

Failure to comply with any post-approval requirement constitutes a ground for withdrawal of approval of a PMA. The introduction or delivery for introduction into interstate commerce of a device that is not in compliance with its conditions of approval is a violation of law.

You are reminded that, as soon as possible and before commercial distribution of your device, you must submit an amendment to this PMA submission with copies of all approved labeling in final printed form. Final printed labeling that is identical to the labeling approved in draft form will not routinely be reviewed by FDA staff when accompanied by a cover letter stating that the final printed labeling is identical to the labeling approved in draft form. If the final printed labeling is not identical, any changes from the final draft labeling should be highlighted and explained in the amendment.

All required documents should be submitted in 6 copies, unless otherwise specified, to the address below and should reference the above PMA number to facilitate processing.

U.S. Food and Drug Administration  
Center for Devices and Radiological Health  
PMA Document Control Center - WO66-G609  
10903 New Hampshire Avenue  
Silver Spring, MD 20993-0002

If you have any questions concerning this approval order, please contact Eleni Whatley at 301-796-6372 or [Eleni.Katsanevakis@fda.hhs.gov](mailto:Eleni.Katsanevakis@fda.hhs.gov).

Sincerely yours,

for  
Bram D. Zuckerman, M.D.  
Director  
Division of Cardiovascular Devices  
Office of Device Evaluation  
Center for Devices and Radiological Health