



Food and Drug Administration  
10903 New Hampshire Avenue  
Document Control Center - WO66-G609  
Silver Spring, MD 20993-0002

May 12, 2015

Cook Incorporated  
Chad Schulenburg  
Regulatory Affairs Specialist  
750 Daniels Way  
P.O. Box 489  
Bloomington, Indiana 47404

Re: K150970  
Trade/Device Name: Coda LP Balloon Catheter  
Regulation Number: 21 CFR 870.1250  
Regulation Name: Percutaneous Catheter  
Regulatory Class: Class II  
Product Code: DQY  
Dated: April 10, 2015  
Received: April 13, 2015

Dear Chad Schulenburg:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the [Federal Register](#).

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-

related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

**Kenneth J. Cavanaugh -S**

for

Bram D. Zuckerman, M.D.

Director

Division of Cardiovascular Devices

Office of Device Evaluation

Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)

K150970

Device Name

Coda® LP Balloon Catheter

Indications for Use (Describe)

The Coda® LP Balloon Catheter is intended for the temporary occlusion of large vessels, or to expand vascular prostheses.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*



## 510(k) SUMMARY

**Submitted By:** Chad Schulenburg, MA  
Cook Incorporated  
750 Daniels Way  
P.O. Box 489  
Bloomington, IN 47404  
Phone: (812) 335-3575 x104073  
Fax: (812) 332-0281  
Date Prepared: May 12, 2015

### Device:

Trade Name: Coda<sup>®</sup> LP Balloon Catheter  
Common Name: Occlusion Balloon Catheter  
Classification Name: Catheter, Percutaneous  
DQY (21 CFR §870.1250)

### Indications for Use:

The Coda<sup>®</sup> Low Profile Balloon Catheter is intended for the temporary occlusion of large vessels, or to expand vascular prostheses.

### Predicate Devices:

The device subject of this submission is substantially equivalent to the predicate devices, cleared for market under 510(k) numbers K032869 on November 19, 2003 and K122917 on June 28, 2013.

### Comparison to Predicate Devices:

It has been demonstrated that the Coda LP Balloon Catheters are comparable to the predicate devices. The Coda LP Balloon Catheters are identical in terms of intended use, principles of operation, materials of construction, and basic technological characteristics to the predicate devices. The device changes consist of a balloon with a smaller diameter and volume, and a smaller catheter French size. The Coda LP Balloon Catheters are also used through a smaller introducer sheath. The safety and effectiveness of the modifications are supported by testing.



### **Device Description:**

The Coda LP Balloon Catheters are over-the-wire catheters that are available with an inflated balloon diameter of 32 millimeters and a balloon volume of 30 cc. The catheters are 9.0 French and are available in lengths of 100 or 120 centimeters. The Coda LP Balloon Catheter shaft contains two independent lumens within a single extrusion. The distal lumen extends the length of the catheter and is used for placement over a 0.035 inch diameter wire guide. The balloon (inflation and deflation) lumen extends from the proximal hub to the two inflation media exit ports within the balloon. The catheters are supplied sterile and are intended for one-time use.

### **Test Data:**

The following tests were performed to demonstrate that the Coda LP Balloon Catheters met applicable design and performance requirements and support a determination of substantial equivalence.

- Balloon Burst Testing – Testing showed that the balloons burst above the maximum rated volume, with all failures being non-fragmentary. The acceptance criteria were met.
- Balloon Inflation/Deflation Testing – Testing showed that the balloons inflated to rated burst pressure within 30 seconds and fully deflated within 30 seconds. The acceptance criteria were met.
- Simulated Use Testing – Testing showed that the catheters were capable of being inserted into and retracted from an appropriately sized sheath without experiencing excessive resistance. Testing showed that this device was able to expand an endovascular graft by inflating the balloon. The acceptance criteria were met.
- Fatigue Testing – Testing showed that the balloons were free from leakage and damage on inflation, withstanding 40 cycles of inflation/deflation. The acceptance criterion was met.
- Occlusion and Migration Testing – Testing showed that this device was able to occlude fluid flow without migrating. The acceptance criteria were met.
- Tensile Strength Testing – Testing showed that under proper clinical use of the device, the peak load values were in accordance with the acceptance criteria.



COOK INCORPORATED  
750 DANIELS WAY, P.O. BOX 489  
BLOOMINGTON, IN 47402-0489 U.S.A.  
PHONE: 812.339.2235 TOLL FREE: 800.457.4500  
WWW.COOKMEDICAL.COM

**Conclusion:**

In conclusion, the results of these tests support a determination of substantial equivalence to the predicate devices.