



Food and Drug Administration  
10903 New Hampshire Avenue  
Document Control Center – WO66-G609  
Silver Spring, MD 20993-0002

July 9, 2015

Stryker Corporation  
Garry T. Hayeck, Ph.D.  
Senior Regulatory Affairs Specialist  
2 Pearl Court  
Allendale, New Jersey 07401

Re: K151755  
Trade/Device Name: OASYS<sup>®</sup> System  
Regulatory Class: Unclassified  
Product Code: NKG, KWP  
Dated: June 26, 2015  
Received: June 29, 2015

Dear Dr. Hayeck:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Page 2 – Garry T. Hayeck, Ph.D.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

**Lori A. Wiggins -S**

for

Mark N. Melkerson

Director

Division of Orthopedic Devices

Office of Device Evaluation

Center for Devices and

Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)

K151755

Device Name

OASYS® System

Indications for Use (Describe)

The Stryker Spine OASYS® System is intended to provide immobilization and stabilization of spinal segments as an adjunct to fusion for the following acute and chronic instabilities of the craniocervical junction, the cervical spine (C1 to C7) and the thoracic spine (T1-T3): traumatic spinal fractures and/or traumatic dislocations; instability or deformity; failed previous fusions (e.g. pseudoarthrosis); tumors involving the cervical/thoracic spine; and degenerative disease, including intractable radiculopathy and/or myelopathy, neck and/or arm pain of discogenic origin as confirmed by radiographic studies, and degenerative disease of the facets with instability.

The Stryker Spine OASYS® System is also intended to restore the integrity of the spinal column even in the absence of fusion for a limited time period in patients with advanced stage tumors involving the cervical spine in whom life expectancy is of insufficient duration to permit achievement of fusion.

The Stryker Spine OASYS® System can be linked to the Xia® System, SR90D System and Xia® 4.5 Spinal System via the rod-to-rod connectors and transition rods.

The Stryker Spine OASYS® System can also be linked to the polyaxial screws of the Xia® II and Xia® 3 Systems via the saddle connector.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

**PLEASE DO NOT WRITE BELOW THIS LINE – CONTINUE ON A SEPARATE PAGE IF NEEDED.**

### FOR FDA USE ONLY

Concurrence of Center for Devices and Radiological Health (CDRH) (Signature)

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
PRAStaff@fda.hhs.gov

*“An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number.”*

<b>510(k) Summary: OASYS® System</b>	
Submitter	Stryker Corporation 2 Pearl Court Allendale, NJ 07401
Contact Person	Garry T. Hayeck, Ph.D. Senior Regulatory Affairs Specialist Phone: 201-760-8043 Fax: 201-962-4043 E-mail: garry.hayeck@stryker.com
Date Prepared	June 26, 2015
Trade Name	OASYS® System
Common Name	Spinal Fixation Appliances
Proposed Class	Unclassified, Class II
Classification Name, Codification	Orthosis, Cervical Pedicle Screw Spinal Fixation Spinal Interlaminar Fixation Orthosis, 21 CFR § 888.3050
Product Codes	NKG, KWP
Predicate Devices	<b>Primary Predicate:</b> Stryker Spine OASYS® System: K150539  <b>Additional Predicate:</b> Stryker Spine OASYS® System: K150753
Device Description	The Stryker Spine OASYS® System is comprised of rods, polyaxial screws, bone screws, hooks, connectors, and occiput plates. The components are available in a variety of lengths in order to accommodate patient anatomy. The components are fabricated from Titanium alloy and CP Titanium and are provided non-sterile. The subject system also offers Vitallium® rods. The Stryker Spine OASYS® System can be linked to the Stryker Spine Xia® family and Xia 4.5 Systems and SR90D System.
Indications for Use	<p>The Stryker Spine OASYS® System is intended to provide immobilization and stabilization of spinal segments as an adjunct to fusion for the following acute and chronic instabilities of the craniocervical junction, the cervical spine (C1 to C7) and the thoracic spine (T1-T3): traumatic spinal fractures and/or traumatic dislocations; instability or deformity; failed previous fusions (e.g. pseudoarthrosis); tumors involving the cervical/thoracic spine; and degenerative disease, including intractable radiculopathy and/or myelopathy, neck and/or arm pain of discogenic origin as confirmed by radiographic studies, and degenerative disease of the facets with instability.</p> <p>The Stryker Spine OASYS® System is also intended to restore the integrity of the spinal column even in the absence of fusion for a limited time period in patients with advanced stage tumors involving the cervical spine in whom life expectancy is of insufficient duration to permit achievement of fusion.</p> <p>The Stryker Spine OASYS® System can be linked to the Xia® System, SR90D System and Xia® 4.5 Spinal System via the rod-to-rod connectors and transition rods.</p>

	The Stryker Spine OASYS® System can also be linked to the polyaxial screws of the Xia® II and Xia® 3 Systems via the saddle connector.
Summary of Technological Characteristics	The subject OASYS® System shares the same materials, geometries, and fundamental scientific technologies as the predicate OASYS® System. None of the aforementioned characteristics have been altered, augmented, or otherwise changed.
Summary of Performance Data	This submission seeks to reconcile the expansion of indications cleared under K150539 with the new implants cleared under K150753. Therefore, no additional performance data is necessary.
Conclusion	The devices, methodologies, and materials used in this system are equivalent to previously cleared OASYS® Systems. As such, this system is substantially equivalent to the predicate systems.