



Food and Drug Administration  
10903 New Hampshire Avenue  
Document Control Center - WO66-G609  
Silver Spring, MD 20993-0002

December 1, 2015

Edan Instruments, Inc.  
% Doug Worth  
Sr. Director, US Regulatory and Quality Affairs  
Edan Medical USA, Inc.  
1200 Crossman Ave, Suite 200  
Sunnyvale, CA 94086

Re: K151878  
Trade/Device Name: Video Colposcope, Models C3A, C6A  
Regulation Number: 21 CFR 884.1630  
Regulation Name: Colposcope  
Regulatory Class: II  
Product Code: HEX  
Dated: October 21, 2015  
Received: October 26, 2015

Dear Doug Worth,

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the [Federal Register](#).

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-

related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

  
**Herbert P. Lerner -S**

for Benjamin R. Fisher, Ph.D.  
Director  
Division of Reproductive, Gastro-Renal,  
and Urological Devices  
Office of Device Evaluation  
Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)

K151878

Device Name

Video Colposcope, models C3A, C6A

Indications for Use (Describe)

The C3A, C6A video colposcope is intended for gynecological examination. It provides magnified visualization of the vagina, cervix and external genitalia, which can help diagnose abnormalities and select areas for biopsy. It is intended to be used only by trained and qualified personnel in hospitals, clinics and private offices, and not intended for home use or to touch the patient.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

## 510(K) Summary

Prepared in accordance with the content and format regulatory requirements of 21 CFR Part 807.92

**Submitted by:** Edan Instruments, Inc.  
3/F - B, Nanshan Medical  
Equipments Park, Nanhai Rd 1019#,  
Shekou, Nanshan Shenzhen,  
518067 P.R. China

**Contact person:** Queena Chen  
Tel: +86(0755) 26858736  
Fax: +1 (408) 418-4059  
Email: [chenqin@edan.com.cn](mailto:chenqin@edan.com.cn)

**Date prepared:** November 30, 2015

**Device Identification:** **Trade name:** Video Colposcope, models C3A, C6A  
**Common name:** Colposcope  
**Regulation No:** 21 CFR 884.1630, Colposcope  
**Regulatory Class:** II  
**Product Code:** HEX – Colposcope

---

**Predicate Device:** Shenzhen Goldway Medical Electronics Co., Ltd,  
SLC-2000 Digital Video Colposcope Imaging System  
(K021153)

Cleared – February 10, 2003

**Device Description:** The Edan video colposcope, including models C3A and C6A, consists of the camera module (CCD color camera), stand (vertical or swing arm stand), video capture box, foot switch, and video colposcope software. The Edan video colposcope is a non-patient contacting medical device.

The C3A, C6A video colposcope is a digital imaging equipment that works as follows: The loop group LED light illumines the target, and the CCD camera takes images of the target. The images are converted into video signals, which are then captured and transmitted to a commercially available computer or monitor. The video colposcope software implements the functions of

displaying images, capturing images or taking videos, storing and managing images and providing reports.

**Indications for Use:** The C3A, C6A video colposcope is intended for gynecological examination. It provides magnified visualization of the vagina, cervix and external genitalia, which can help diagnose abnormalities and select areas for biopsy. It is intended to be used only by trained and qualified personnel in hospitals, clinics and private offices, and not intended for home use or to touch the patient.

**Predicate Device Comparison**

The subject device and the predicate device (K021153) have the same fundamental technology and similar technological characteristics, detailed in the table below:

<b>Comparison Items</b>	<b>SLC-2000A(GW)</b>	<b>C3A&amp;C6A</b>	<b>Comparison of C3A, C6A with SLC-2000A</b>
<b>Manufacturer/K#</b>	Goldway/K021153	EDAN	N/A
<b>Indications For Use</b>			
<b>Indications for Use</b>	The Goldway Digital Video Colposcope Imaging System is intended for magnified viewing of the vagina, cervix and external genitalia in order to diagnose abnormalities and select areas for biopsy. The image system is intended to provide documentation of the image in the field of view of the colposcope. The image can be viewed on a color screen, printed on a color printer or archived for storage and subsequent retrieval. The device is intended to be used in Hospitals and clinics.	The C3A, C6A video colposcope is intended for gynecological examination. It provides magnified visualization of the vagina, cervix and external genitalia, which can help diagnose abnormalities and select areas for biopsy. It is intended to be used only by trained and qualified personnel in hospitals, clinics and private offices, and not intended for home use or to touch the patient.	Different
<b>Power Supply</b>			
<b>Voltage</b>	100 V-240 V~	100 V-240 V~	Same
<b>Frequency</b>	50 Hz/60 Hz	50 Hz/60 Hz	Same

<b>Input power (Maximum)</b>	500 VA	48 VA	Different
<b>Specifications</b>			
<b>Standard Configuration</b>	Digital CCD camera, Stand	Digital CCD camera, Stand	Same
<b>Light module</b>	Double loop group LED light	Single loop group LED light (C3A) Double loop group LED light (C6A)	Different
<b>Light source</b>	White LED light	White LED light	Same
<b>Illumination</b>	2200 lux at working distance 300 mm	1600 lux at working distance 300 mm (C3A) 3000 lux at working distance 300 mm (C6A)	Different
<b>Illumination range</b>	$\geq\phi 60$ mm at working distance 200 mm	$\geq\phi 60$ mm at working distance 200 mm	Same
<b>Light source lifetime</b>	$\geq 10,000$ hours	$\geq 10,000$ hours	Same
<b>System resolution</b>	$\geq 470$ TVL	$\geq 500$ TVL	Different
<b>Space resolution</b>	Not available	$\geq 10$ lpm	Different
<b>Image geometric distortion</b>	$< 2.6\%$	$< 3\%$	Different
<b>Magnification</b>	1 ~ 40X	1 ~ 28X (C3A) 1 ~ 36X (C6A)	Different
<b>Operation Distance:</b>	200 mm-300 mm	200 mm-300 mm	Same
<b>Field of view</b>	At minimum magnification $52^\circ$ or $\geq\phi 60$ mm At maximum magnification $\geq\phi 10$ mm	3X: $\geq\phi 80$ mm OR $\geq 16.5^\circ$ 18X: $\geq\phi 12$ mm OR $\geq 2.5^\circ$	Different
<b>Depth of field:</b>	1X: $\geq 120$ mm 40X: $\geq 5$ mm	6X: $\geq 120$ mm 18X: $\geq 6$ mm	Different
<b>Focus mode</b>	Electronic control: Auto focus only	Electronic control: Manual and auto focus	Different
<b>Electronic Filter</b>	Green filter (3 grades)	Green filter (3 grades)	Same
<b>Magnification and timing display</b>	YES	YES	Same
<b>Freeze function</b>	YES	YES	Same
<b>Stand type</b>	Vertical	Vertical, swing arm (optional)	Different
<b>Video Output</b>	S-Video, Video	S-Video	Different
<b>MBTF</b>	$\leq 50,000$ h	$\leq 48,956$ h	Different

<b>Standards Compliance</b>	IEC60601-1	IEC 60601-1: 2005, IEC 60601-1-2: 2007, ISO 8600-3, ISO 8600-5	Different
-----------------------------	------------	---	-----------

The technological characteristics of the C3A, C6A video colposcope are comparable to the predicate device.

**Performance Data:**

The following performance data were provided in support of the substantial equivalence determination.

**Electrical safety and electromagnetic compatibility (EMC)**

Electrical safety and EMC testing were conducted on the C3A, C6A video colposcope. The system complies with the IEC 60601-1:2005/A1: 2012 and standards for safety and the IEC 60601-1-2: 2007 standard for EMC.

**Software Verification and Validation Testing**

Software verification and validation testing were conducted and documentation was provided as recommended by FDA’s Guidance for Industry and FDA Staff, “Guidance for the Content of Premarket Submissions for Software Contained in Medical Devices.” The software for this device was considered as a “moderate” level of concern.

**Bench Testing**

The most important performance specification is about the features associated with imaging, testing conducted to show the effectiveness of the subject device include:

- ISO 8600-3: 2003 Optics and photonics —Medical endoscopes and endotherapy devices — Part 3: Determination of field of view and direction of view of endoscopes with optics.
- ISO 8600-5: 2005 Optics and photonics —Medical endoscopes and endotherapy devices — Part 5: Determination of optical resolution of rigid endoscopes with optics.
- Thermal Testing
- Image Geometric Distortion Testing

**Conclusion**

The results of the testing demonstrate that the C3A, C6A video colposcope is as safe and effective as the predicate device and supports a determination of substantial equivalence.