

Food and Drug Administration 10903 New Hampshire Avenue Document Control Center – WO66-G609 Silver Spring, MD 20993-0002

Asclepion Laser Technologies GmbH Mrs. Antje Katzer Product Management and International Regulatory Affairs Bruesseler Str.10, Jena, Thuringia 07747 Germany

December 11, 2015

Re: K152153 Trade/Device Name: MicroSpot Handpiece Regulation Number: 21 CFR 878.4810 Regulation Name: Laser surgical instrument for use in general and plastic surgery and in dermatology
Regulatory Class: Class II Product Code: GEX Dated: November 13, 2015 Received: November 13, 2015

Dear Mrs. Katzer:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you; however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the <u>Federal Register</u>.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set

forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm.

Sincerely yours,

Joshua C. Nipper -S

Binita S. Ashar, M.D., M.B.A., F.A.C.S. Director

For

Division of Surgical Devices Office of Device Evaluation Center for Devices and Radiological Health

Enclosure

collection of	rson is not required to respond to, a currently valid OMB number."	"An agency may not conduct or sponsor, and a pe information unless it displays a
	th and Human Services ninistration mation Officer on Act (PRA) Staff gov	Department of Heal Food and Drug Adn Office of Chief Infor Paperwork Reductic PRAStaff@fda.hhs.
onse, including the led and complete any other aspect	mated to average 79 hours per respondation of the second s	burden time for this collection of information is estire to review instructions, search existing data sources review the collection of information. Send comment nis information collection, including suggestions for r
995. S BELOW.*	of the Paperwork Reduction Act of 19 THE PRA STAFF EMAIL ADDRES	This section applies only to requirements *DO NOT SEND YOUR COMPLETED FORM TO
	ATE PAGE IF NEEDED.	CONTINUE ON A SEPAR
FR 801 Subpart C)	Over-The-Counter Use (21 Ci	Select one or both, as applicable) Prescription Use (Part 21 CFR 801 Subpart D)
edures.	cedures and Skin resurfacing proce	oot Handpiece is intended for Dermatological pro
		indpiece
		r <i>(if known)</i> K152153
oved: OMB No. 0910-0120 Date: January 31, 2017 Statement below.	VICES Form Appro	DEPARTMENT OF HEALTH AND HUMAN SERV Food and Drug Administration Indications for Use



Asclepion Laser Technologies GmbH · Brüsseler Str. 10 · 07747 Jena · Germany

Traditional 510(k) SUMMARY

MicroSpot Handpiece

This Traditional 510(k) summary of safety and effectiveness for the Asclepion Laser Technologies GmbH MicroSpot Handpiece is submitted in accordance with the requirements of 21 CFR 907.92 and follows Office of Device Evaluation Guidance concerning the organization and content of a 510(k) summary.

Applicant:	ASCLEPION LASER TECHNOLOGIES GmbH Bruesseler Str. 10 07747 Jena, Germany
Contact Person:	Mrs. Antje Katzer Product Management and International Regulatory Affairs
Phone: Fax: e-mail:	+49 3641 77 00 309 +49 3641 77 00 302 antje.katzer@asclepion.com
Preparation Date:	December 10, 2015
Device Name:	MicroSpot Handpiece
Common Name:	Microbeam Handpiece for Er:YAG Laser

Our general terms and conditions: www.asclepion.com

Registered office: Jena Register of commerce court: Jena HRB 209648 UST ID Nr. DE 813678553 WEEE-Reg.-Nr. DE 33663120 Managing Director: Dr. Danilo Leggieri Bank Connections: Sparkasse Jena • SWIFT HELADEF1JEN • IBAN DE 3483053030000000094 Deutsche Bank Jena • SWIFT DEUTDE8EXXX • IBAN DE 67820700000397755000 Commerzbank Jena • SWIFT COBADEFF821 • IBAN DE 54820400000258272400

Classification Name:	Laser surgical instrument for use in general and plastic surgery and in dermatology 79-GEX 21 CFR 878.4810 Regulatory Class: Class II Product Code: GEX		
Equivalent Devices:	K081541 K101306 K132806	Dermablate Effect (Asclepion Laser Technologies) Fotona Dynamis Er:YAG Laser System (Fotona) Fotona FS-01 Laser Handpiece (Fotona)	
Device Description:	The MicroSpo used with the MCL 31 Derm 2940 nm. The is triggered by with a handpie The system in plume.	t Handpiece is a handpiece with microbeam output to be previously cleared MCL 31 Dermablate (K150140). The ablate is a pulsed Er:YAG laser emitting a wavelength of system comprises a main console unit, a handpiece and means of a footswitch. The MCL 31 can be operated ece of larger spotsize or with a microbeam handpiece. corporates a suction unit for the safe removal of laser	
Intended Use:	The MicroSpo and Skin resu	t handpiece is intended for Dermatological procedures rfacing procedures.	

Summary of Technical Characteristics

	Proposed Modified Device	Un-Modified Predicate Device	Un-Modified Predicate Device
Name Manufacturer 510(k)	MCL 31 Dermablate with MicroSpot Zoom Handpiece Asclepion Laser Technologies	Dermablate Effect with MicroSpot - Zoom handpiece Asclepion Laser Technologies	Dynamis Er:YAG Laser System Family with FS 01-Handpiece Fotona
		K081541	K101306 / K132806
Intended Use (Microbeam handpiece)	Dermatological procedures. Skin resurfacing procedures.	Dermatological procedures. Skin resurfacing procedures.	Dermatological procedures requiring resurfacing of soft tissue with fractionated handpiece.
Laser medium wavelength	Er:YAG 2940 nm	Er:YAG 2940 nm	Er :YAG 2940 nm
Energy, max.	2,5 J	1,5 J	3,0 J
Power, max.	20 W	12 W	20 W
Frequency, max.	20 Hz	20 Hz	50 Hz
Fluence, max.	10 J/cm² Up to 150 J/cm² with stacking pulses	18 J/cm ² Up to 162 J/cm ² with stacking pulses	Up to 800 J/cm ² with stacking pulses
Max. Energy per Microbeam	12 mJ	8 mJ	160 mJ
Pulse Duration	0,1 – 1,0 ms	< 1 ms	0,1 – 1,5 ms
Pulse repetition rate	Up to 20 Hz	Up to 20 Hz	Up to 50 Hz
Spot size of microbeams	350 and 600 µm	250 µm	250 µm
Creation of microbeams	By microlens array	By microlens array	By microlens array
Treatment zones	13 x 13 mm	13 x 13 mm 9 x 9 mm 7 x 7 mm	13 x 13 mm
Coverage of skin	10 and 25 %	5 – 17 %	5 – 60 %

Comparison to: The MicroSpot Handpiece of the MCL 31 Dermablate laser system is substantially equivalent to the MicroSpot Handpiece of the Dermablate Effect K081541 and the Fotona FS-01 Laser Handpiece K132806 of the Fotona Dynamis Er:YAG laser K101306 with the same principles of operation, with similar parameters and with similar indications for use. The fundamental scientific technology of the device is unchanged from the legally marketed predicates.

Nonclinical Performance Data: The MicroSpot Handpiece is designed and tested according to following standards:

ISO 14971:2009 IEC 60601-1:2005 IEC 60601-1-2:2007 IEC 60601-1-6:2010 IEC 60601-2-22:2007 IEC 60825-1:2007 IEC 62304:2006

Laboratory testing was conducted to validate and verify that the MicroSpot Handpiece met all design specifications and was substantially equivalent to the predicate devices.

The size of the single microspots was evaluated as well as the laser pulse duration, the energy release of the laser and the energy in a duration test.

The results of those testings were used to select the desired parameters with the MicroSpot handpiece.

Animal Performance Data: Histological data was submitted to support safety and effectiveness of the microbeam mode. The device was used in vivo on pig skin with low, medium and high energies, and with different pulse durations and densities. The five possible settings of the handpiece were applied to nine pigs. Three received 20 J/cm², three received 40 J/cm² and three received 60 J/cm². The histological samples have been collected at day 0, 3 and 14. The biopsied areas show the depth and width of ablation as well as the thermal damage zones and demonstrate the healing response over time.
 Conclusion: The MicroSpot Handpiece is another safe and effective device for Dermatological and Skin resurfacing procedures.