



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center – WO66-G609
Silver Spring, MD 20993-0002

Siemens Medical Solutions USA, Inc., Ultrasound Division
% Mr. Mark Job
Responsible Third Party Official
Regulatory Technology Services, LLC
1394 25th Street, NW
BUFFALO MN 55313

September 21, 2015

Re: K152469
Trade/Device Name: ACUSON NX3 Elite and ACUSON NX3 Diagnostic
Ultrasound System
Regulation Number: 21 CFR 892.1550
Regulation Name: Ultrasonic pulsed Doppler imaging system
Regulatory Class: II
Product Code: IYN, IYO, ITX
Dated: August 28, 2015
Received: August 31, 2015

Dear Mr. Job:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638 2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, “Misbranding by reference to premarket notification” (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH’s Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

A handwritten signature in black ink that reads "Robert Ochs". The signature is written in a cursive style and is positioned above the typed name.

Robert Ochs, Ph.D.
Director
Division of Radiological Health
Office of In Vitro Diagnostics
and Radiological Health
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)
K152469

Device Name
ACUSON NX3 Elite and ACUSON NX3 Diagnostic Ultrasound Systems

Indications for Use (Describe)

For ACUSON NX3

The ACUSON NX3 ultrasound imaging system, is intended for the following applications: Fetal, Abdominal (including liver), Pediatric, Small Parts(Small Organ), Transcranial, OB/GYN, Pelvic, Neonatal, Cardiac, Vascular(including Peripheral Vessel), Musculoskeletal, Superficial Musculoskeletal and Urology applications. The systems also provide for the measurement of anatomical structures and for analysis packages that provide information that is used for clinical diagnosis purposes.

Ultrasound is used as an imaging aid, but may have further restrictions specific to in-vitro fertilization (IVF), chorionic villus sampling (CVS), and percutaneous umbilical cord blood sampling (PUBS) procedures. Observe local laws and regulations.

For ACUSON NX3 Elite

ACUSON NX3 Elite ultrasound imaging systems are intended for the following applications: Fetal, Abdominal (including liver, intra-operative), Pediatric, Small Parts(Small Organ including intra-operative), Transcranial, OB/GYN, Pelvic, Neonatal, Cardiac, Vascular(including Peripheral Vessel, intra-operative), Musculoskeletal, Superficial Musculoskeletal and Urology applications. The systems also provide for the measurement of anatomical structures and for analysis packages that provide information that is used for clinical diagnosis purposes.

Ultrasound is used as an imaging aid, but may have further restrictions specific to in-vitro fertilization (IVF), chorionic villus sampling (CVS), and percutaneous umbilical cord blood sampling (PUBS) procedures. Observe local laws and regulations.

The Arterial Health Package (AHP) software provides the physician with the capability to measure Intima Media Thickness and the option to reference normative tables that have been validated and published in peer-reviewed studies. The information is intended to provide the physician with an easily understood tool for communicating with patients regarding state of their cardiovascular system. Note: This feature can be utilized according to the "ASE Consensus Statement; Use of Carotid Ultrasound to Identify Subclinical Vascular Disease and Evaluate Cardiovascular Disease Risk: A Consensus Statement from the American Society of Echocardiography Carotid Intima-Media Thickness Task Force. Endorsed by the Society for Vascular Medicine."

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

“An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number.”

Diagnostic Ultrasound Indications for Use Form

510(k) Number (if known):

Device Name: **ACUSON NX3 Elite™ Diagnostic Ultrasound System**
 Intended Use: Diagnostic imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation							
Other (Track 1 Only)	Specific (Tracks 11 & 3)	B	M	PWD	CWD	Color Doppler	Power Doppler	Combined (Specify)	Other (Specify)
Ophthalmic	Ophthalmic								
Fetal Imaging & Other	Fetal	N	N	N		N	N	BMDC	Note 2,3,4,5,6,9,10,12,14
	Abdominal	N	N	N		N	N	BMDC	Note 2,3,4,5,6,9,10,15
	Intra-operative (Note 6)	N	N	N		N	N	BMDC	Note 2,3,4,5,6,9,10
	Intra-operative (Neuro)								
	Laparoscopic								
	Pediatric	N	N	N		N	N	BMDC	Note 2,3,4,5,6,9,10
	Small Organ (Note 1)	N	N	N		N	N	BMDC	Note 2,3,4,5,6,9,10,17
	Neonatal Cephalic	N	N	N		N	N	BMDC	Note 2,3,4,5,6,9
	Adult Cephalic	N	N	N	N	N	N	BMDC	Note 2,3,4,5,6,9,11
	Trans-rectal	N	N	N		N	N	BMDC	Note 2,3,4,5,6,9,10
	Trans-vaginal	N	N	N		N	N	BMDC	Note 2,3,4,5,9,10,12,14,16
	Trans-urethral								
	Trans-esoph. (non-Card.)								
	Musculo-skel. (Convent.)	N	N	N		N	N	BMDC	Note 2,3,4,5,6,9,10,11
	Musculo-skel. (Superfic)	N	N	N		N	N	BMDC	Note 2,3,4,9,10,11
Intra -vascular									
Other (Specify)									
Cardiac	Cardiac Adult	N	N	N		N	N	BMDC	Note 2,7,8,9,13,
	Cardiac Pediatric	N	N	N		N	N	BMDC	Note 2,3,4,6,8,9,13
	Intra-vascular (Cardiac)								
	Trans-esophageal (Cardiac)								
	Intra-Cardiac								
	Other (Specify)								
Peripheral Vessel	Peripheral vessel	N	N	N		N	N	BMDC	Note 2,3,4,6,9,10
	Other (Specify)								

N = new indication; P = previously cleared

- | | | | |
|---------|---|---------|---|
| Note 1 | For example: breast, testes, thyroid, penis, prostate, etc. | Note 11 | syngo Arterial Health Package (AHP) (NX3 Elite only) |
| Note 2 | Dynamic TCE Technology | Note 12 | syngo Auto OB Measurements |
| Note 3 | SieClear | Note 13 | syngo Auto Left Heart (Auto LH) Technology (NX3 Elite only) |
| Note 4 | Advanced SieClear | Note 14 | syngo fourSight 4D imaging |
| Note 5 | 3-Scape 3D Imaging | Note 15 | Contrast Agent Imaging (NX3 Elite only) |
| Note 6 | For example: abdominal, vascular | Note 16 | syngo Auto follicle |
| Note 7 | Stress Echo Imaging | Note 17 | eSie Touch (NX3 Elite only) |
| Note 8 | Axiu Edge Assisted Ejection Fraction | | |
| Note 9 | Clarify Vascular Enhancement Technology | | |
| Note 10 | SieScape Panoramic Imaging | | |

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

 Concurrence of Center for Devices and Radiological Health (CDRH) (Signature)

 Concurrence of Center for Devices and Radiological Health (CDRH) (Signature)

510(k) _____

Diagnostic Ultrasound Indications for Use Form

510(k) Number (if known):

Device Name:

ACUSON NX3™ Diagnostic Ultrasound System

Intended Use:

Diagnostic imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation							
Other (Track 1 Only)	Specific (Tracks 11 & 3)	B	M	PWD	CWD	Color Doppler	Power Doppler	Combined (Specify)	Other (Specify)
Ophthalmic	Ophthalmic								
Fetal Imaging & Other	Fetal	N	N	N		N	N	BMDC	Note 2,3,4,5,6,9,10,12,14
	Abdominal	N	N	N		N	N	BMDC	Note 2,3,4,5,6,9,10
	Intra-operative (Note 6)								
	Intra-operative (Neuro)								
	Laparoscopic								
	Pediatric	N	N	N		N	N	BMDC	Note 2,3,4,5,6,9,10
	Small Organ (Note 1)	N	N	N		N	N	BMDC	Note 2,3,4,5,6,9,10
	Neonatal Cephalic								
	Adult Cephalic	N	N	N	N	N	N	BMDC	Note 2,3,4,5,6,9
	Trans-rectal	N	N	N		N	N	BMDC	Note 2,3,4,5,6,9,10
	Trans-vaginal	N	N	N		N	N	BMDC	Note 2,3,4,5,9,10,12,16
	Trans-urethral								
	Trans-esoph. (non-Card.)								
	Musculo-skel. (Convent.)	N	N	N		N	N	BMDC	Note 2,3,4,5,6,9,10
	Musculo-skel. (Superfic)	N	N	N		N	N	BMDC	Note 2,3,4,9,10
Intra -vascular									
Other (Specify)									
Cardiac	Cardiac Adult	N	N	N		N	N	BMDC	Note 2,7,8,9
	Cardiac Pediatric	N	N	N		N	N	BMDC	Note 2,7,8,9
	Intra-vascular (Cardiac)								
	Trans-esophageal (Cardiac)								
	Intra-Cardiac								
Other (Specify)									
Peripheral Vessel	Peripheral vessel	N	N	N		N	N	BMDC	Note 2,3,4,5,6,9,10
	Other (Specify)								

N = new indication; P = previously cleared

- Note 1 For example: breast, testes, thyroid, penis, prostate, etc.
- Note 2 Dynamic TCE Technology
- Note 3 SieClear
- Note 4 Advanced SieClear
- Note 5 3-Scape 3D Imaging
- Note 6 For example: abdominal, vascular
- Note 7 Stress Echo Imaging
- Note 8 Axius Edge Assisted Ejection Fraction
- Note 9 Clarify Vascular Enhancement Technology
- Note 10 SieScape Panoramic Imaging

- Note 11 syngo Arterial Health Package (AHP) (NX3 Elite only)
- Note 12 syngo Auto OB Measurements
- Note 13 syngo Auto Left Heart (Auto LH) Technology (NX3 Elite only)
- Note 14 syngo fourSight 4D imaging
- Note 15 Contrast Agent Imaging (NX3 Elite only)
- Note 16 syngo Auto follicle
- Note 17 eSie Touch (NX3 Elite only)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of Center for Devices and Radiological Health (CDRH) (Signature)

Concurrence of Center for Devices and Radiological Health (CDRH) (Signature)

510(k) _____

Diagnostic Ultrasound Indications for Use Form

510(k) Number (if known):

Device Name: **CH5-2 Transducer for use with:
 ACUSON NX3 Elite™ Diagnostic Ultrasound System / ACUSON NX3™ Diagnostic Ultrasound System**

Intended Use: Diagnostic imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation							
Other (Track 1 Only)	Specific (Tracks 11 & 3)	B	M	PWD	CWD	Color Doppler	Power Doppler	Combined (Specify)	Other (Specify)
Ophthalmic	Ophthalmic								
Fetal Imaging & Other	Fetal	P	P	P		P	P	BMDC	Note 2,3,4,5,6,9,10,12
	Abdominal	P	P	P		P	P	BMDC	Note 2,3,4,5,6,9,10,15,16
	Intra-operative (Note 6)								
	Intra-operative (Neuro)								
	Laparoscopic								
	Pediatric	P	P	P		P	P	BMDC	Note 2,3,4,5,6,9,10
	Small Organ (Note 1)								
	Neonatal Cephalic								
	Adult Cephalic								
	Trans-rectal								
	Trans-vaginal								
	Trans-urethral								
	Trans-esoph. (non-Card.)								
	Musculo-skel. (Convent.)								
	Musculo-skel. (Superfic)								
	Intra-vascular								
Other (Specify)									
Cardiac	Cardiac Adult								
	Cardiac Pediatric								
	Intra-vascular (Cardiac)								
	Trans-esophageal (Cardiac)								
	Intra-cardiac								
	Other (Specify)								
Peripheral Vessel	Peripheral vessel	P	P	P		P	P	BMDC	Note 2,3,4,5,6,9,10
	Other (Specify)								

N = new indication; P = previously cleared by K141846

- Note 1 For example: breast, testes, thyroid, penis, prostate, etc.
- Note 2 Dynamic TCE Technology
- Note 3 SieClear
- Note 4 Advanced SieClear
- Note 5 3-Scape 3D Imaging
- Note 6 For example: abdominal, vascular
- Note 7 Stress Echo Imaging
- Note 8 Aixus Edge Assisted Ejection Fraction
- Note 9 Clarify Vascular Enhancement Technology
- Note 10 SieScape Panoramic Imaging

- Note 11 syngo Arterial Health Package (AHP) (NX3 Elite only)
- Note 12 syngo Auto OB Measurements
- Note 13 syngo Auto Left Heart (Auto LH) Technology (NX3 Elite only)
- Note 14 syngo fourSight 4D imaging
- Note 15 Contrast Agent Imaging (NX3 Elite only)
- Note 16 syngo Auto follicle
- Note 17 eSie Touch (NX3 Elite only)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of Center for Devices and Radiological Health (CDRH) (Signature)

Concurrence of Center for Devices and Radiological Health (CDRH) (Signature)

510(k) _____

Diagnostic Ultrasound Indications for Use Form

510(k) Number (if known):

Device Name: **VF10-5 Linear Array Transducer for use with:
 ACUSON NX3 Elite™ Diagnostic Ultrasound System / ACUSON NX3™ Diagnostic Ultrasound System**

Intended Use: Diagnostic imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation							
Other (Track 1 Only)	Specific (Tracks 1 & 3)	B	M	PWD	CWD	Color Doppler	Power Doppler	Combined (Specify)	Other (Specify)
Ophthalmic	Ophthalmic								
Fetal Imaging & Other	Fetal								
	Abdominal	P	P	P		P	P	BMDC	Note 2,3,4,9,10,16
	Intra-operative (Note 6)								
	Intra-operative (Neuro)								
	Laparoscopic								
	Pediatric	P	P	P		P	P	BMDC	Note 2,3,4,9,10
	Small Organ (Note 1)	P	P	P		P	P	BMDC	Note 2,3,4,9,10,17
	Neonatal Cephalic								
	Adult Cephalic	P	P	P		P	P	BMDC	Note 2,3,4,9,10,11
	Trans-rectal								
	Trans-vaginal								
	Trans-urethral								
	Trans-esoph. (non-Card.)								
	Musculo-skel. (Convent.)	P	P	P		P	P	BMDC	Note 2,3,4,9,10
	Musculo-skel. (Superfic)	P	P	P		P	P	BMDC	Note 2,3,4,9,10
Intra-vascular									
Other (Specify)									
Cardiac	Cardiac Adult								
	Cardiac Pediatric								
	Intra-vascular (Cardiac)								
	Trans-esophageal (Cardiac)								
	Intra-cardiac								
	Other (Specify)								
Peripheral Vessel	Peripheral vessel	P	P	P		P	P	BMDC	Note 2,3,4,9,10
	Other (Specify)								

N = new indication; P = previously cleared K141846

- | | | | |
|---------|---|---------|---|
| Note 1 | For example: breast, testes, thyroid, penis, prostate, etc. | Note 11 | syngo Arterial Health Package (AHP) (NX3 Elite only) |
| Note 2 | Dynamic TCE Technology | Note 12 | syngo Auto OB Measurements |
| Note 3 | SieClear | Note 13 | syngo Auto Left Heart (Auto LH) Technology (NX3 Elite only) |
| Note 4 | Advanced SieClear | Note 14 | syngo fourSight 4D imaging |
| Note 5 | 3-Scape 3D Imaging | Note 15 | Contrast Agent Imaging (NX3 Elite only) |
| Note 6 | For example: abdominal, vascular | Note 16 | syngo Auto follicle |
| Note 7 | Stress Echo Imaging | Note 17 | eSie Touch (NX3 Elite only) |
| Note 8 | Axiu Edge Assisted Ejection Fraction | | |
| Note 9 | Clarify Vascular Enhancement Technology | | |
| Note 10 | SieScape Panoramic Imaging | | |

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of Center for Devices and Radiological Health (CDRH) (Signature)

Concurrence of Center for Devices and Radiological Health (CDRH) (Signature)

510(k) _____

Diagnostic Ultrasound Indications for Use Form

510(k) Number (if known):

Device Name: **VF12-4 Linear Array Transducer for use with:
 ACUSON NX3 Elite™ Diagnostic Ultrasound System / ACUSON NX3™ Diagnostic Ultrasound System**

Intended Use: **Diagnostic imaging or fluid flow analysis of the human body as follows:**

Clinical Application		Mode of Operation							
Other (Track 1 Only)	Specific (Tracks 11 & 3)	B	M	PWD	CWD	Color Doppler	Power Doppler	Combined (Specify)	Other (Specify)
Ophthalmic	Ophthalmic								
Fetal Imaging & Other	Fetal								
	Abdominal	P	P	P		P	P	BMDC	Note 2,3,4,9,10,16
	Intra-operative (Note 6)								
	Intra-operative (Neuro)								
	Laparoscopic								
	Pediatric	P	P	P		P	P	BMDC	Note 2,3,4,9,10
	Small Organ (Note 1)	P	P	P		P	P	BMDC	Note 2,3,4,9,10,17
	Neonatal Cephalic – NX3 Elite only	P	P	P		P	P	BMDC	Note 2,3,4,9,10
	Adult Cephalic	P	P	P		P	P	BMDC	Note 2,3,4,9,10,11
	Trans-rectal								
	Trans-vaginal								
	Trans-urethral								
	Trans-esoph. (non-Card.)								
	Musculo-skel. (Convent.)	P	P	P		P	P	BMDC	Note 2,3,4,9,10
	Musculo-skel. (Superfic)	P	P	P		P	P	BMDC	Note 2,3,4,9,10
Intra-vascular									
Other (Specify)									
Cardiac	Cardiac Adult								
	Cardiac Pediatric								
	Intra-vascular (Cardiac)								
	Trans-esophageal (Cardiac)								
	Intra-cardiac								
Other (Specify)									
Peripheral Vessel	Peripheral vessel	P	P	P		P	P	BMDC	Note 2,3,4,9,10
	Other (Specify)								

N = new indication; P = previously cleared by K141846

- | | | | |
|---------|---|---------|---|
| Note 1 | For example: breast, testes, thyroid, penis, prostate, etc. | Note 11 | syngo Arterial Health Package (AHP) (NX3 Elite only) |
| Note 2 | Dynamic TCE Technology | Note 12 | syngo Auto OB Measurements |
| Note 3 | SieClear | Note 13 | syngo Auto Left Heart (Auto LH) Technology (NX3 Elite only) |
| Note 4 | Advanced SieClear | Note 14 | syngo fourSight 4D imaging |
| Note 5 | 3-Scape 3D Imaging | Note 15 | Contrast Agent Imaging (NX3 Elite only) |
| Note 6 | For example: abdominal, vascular | Note 16 | syngo Auto follicle |
| Note 7 | Stress Echo Imaging | Note 17 | eSie Touch (NX3 Elite only) |
| Note 8 | Axius Edge Assisted Ejection Fraction | | |
| Note 9 | Clarify Vascular Enhancement Technology | | |
| Note 10 | SieScape Panoramic Imaging | | |

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of Center for Devices and Radiological Health (CDRH) (Signature)

Concurrence of Center for Devices and Radiological Health (CDRH) (Signature)

510(k) _____

Diagnostic Ultrasound Indications for Use Form

510(k) Number (if known):

Device Name: **EC10-5w** Convex Array Transducer for use with:
ACUSON NX3 Elite™ Diagnostic Ultrasound System
 Intended Use: Diagnostic imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation							
Other (Track 1 Only)	Specific (Tracks 11& 3)	B	M	PWD	CWD	Color Doppler	Power Doppler	Combined (Specify)	Other (Specify)
Ophthalmic	Ophthalmic								
Fetal Imaging & Other	Fetal	N	N	N		N	N	BMDC	Note 2,3,4,5,6,9,10,12
	Abdominal								
	Intra-operative (Note 6)								
	Intra-operative (Neuro)								
	Laparoscopic								
	Pediatric								
	Small Organ (Note 1)	N	N	N		N	N	BMDC	Note 2,3,4,5,9,10
	Neonatal Cephalic	N	N	N		N	N	BMDC	Note 2,3,4,5,6, 9,10
	Adult Cephalic								
	Trans-rectal	N	N	N		N	N	BMDC	Note 2,3,4,5,9,10
	Trans-vaginal	N	N	N		N	N	BMDC	Note 2,3,4,5,9,10,12,16
	Trans-urethral								
	Trans-esoph. (non-Card.)								
	Musculo-skel. (Convent.)								
	Musculo-skel. (Superfic)								
Intra-vascular									
Other (Specify)									
Cardiac	Cardiac Adult								
	Cardiac Pediatric								
	Intra-vascular (Cardiac)								
	Trans-esophageal (Cardiac)								
	Intra-Cardiac								
	Other (Specify)								
Peripheral Vessel	Peripheral vessel								
	Other (Specify)								

N = new indication; P = previously cleared

- Note 1 For example: breast, testes, thyroid, penis, prostate, etc.
- Note 2 Dynamic TCE Technology
- Note 3 SieClear
- Note 4 Advanced SieClear
- Note 5 3-Scape 3D Imaging
- Note 6 For example: abdominal, vascular
- Note 7 Stress Echo Imaging
- Note 8 Axius Edge Assisted Ejection Fraction
- Note 9 Clarify Vascular Enhancement Technology
- Note 10 SieScape Panoramic Imaging

- Note 11 syngo Arterial Health Package (AHP) (NX3 Elite only)
- Note 12 syngo Auto OB Measurements
- Note 13 syngo Auto Left Heart (Auto LH) Technology (NX3 Elite only)
- Note 14 syngo fourSight 4D imaging
- Note 15 Contrast Agent Imaging (NX3 Elite only)
- Note 16 syngo Auto follicle
- Note 17 eSie Touch (NX3 Elite only)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

 Concurrence of Center for Devices and Radiological Health (CDRH) (Signature)

 Concurrence of Center for Devices and Radiological Health (CDRH) (Signature)

510(k) _____

Diagnostic Ultrasound Indications for Use Form

510(k) Number (if known):

Device Name: **EC9-4 Convex Array Transducer for use with:
 ACUSON NX3 Elite™ Diagnostic Ultrasound System / ACUSON NX3™ Diagnostic Ultrasound System**

Intended Use: Diagnostic imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation							
Other (Track 1 Only)	Specific (Tracks 11 & 3)	B	M	PWD	CWD	Color Doppler	Power Doppler	Combined (Specify)	Other (Specify)
Ophthalmic	Ophthalmic								
Fetal Imaging & Other	Fetal	P	P	P		P	P	BMDC	Note 2,3,4,5,6,9,10,12
	Abdominal								
	Intra-operative (Note 6)								
	Intra-operative (Neuro)								
	Laparoscopic								
	Pediatric								
	Small Organ (Note 1)	P	P	P		P	P	BMDC	Note 2,3,4,5,9,10
	Neonatal Cephalic – NX3 Elite Only	P	P	P		P	P	BMDC	Note 2,3,4,5,6, 9,10
	Adult Cephalic								
	Trans-rectal	P	P	P		P	P	BMDC	Note 2,3,4,5,9,10
	Trans-vaginal	P	P	P		P	P	BMDC	Note 2,3,4,5,9,10,12,16
	Trans-urethral								
	Trans-esoph. (non-Card.)								
	Musculo-skel. (Convent.)								
	Musculo-skel. (Superfic)								
Intra-vascular									
Other (Specify)									
Cardiac	Cardiac Adult								
	Cardiac Pediatric								
	Intra-vascular (Cardiac)								
	Trans-esophageal (Cardiac)								
	Intra-Cardiac								
	Other (Specify)								
Peripheral Vessel	Peripheral vessel								
	Other (Specify)								

N = new indication; P = previously cleared by K121699

- Note 1 For example: breast, testes, thyroid, penis, prostate, etc.
- Note 2 Dynamic TCE Technology
- Note 3 SieClear
- Note 4 Advanced SieClear
- Note 5 3-Scape 3D Imaging
- Note 6 For example: abdominal, vascular
- Note 7 Stress Echo Imaging
- Note 8 Axius Edge Assisted Ejection Fraction
- Note 9 Clarify Vascular Enhancement Technology
- Note 10 SieScape Panoramic Imaging

- Note 11 syngo Arterial Health Package (AHP) (NX3 Elite only)
- Note 12 syngo Auto OB Measurements
- Note 13 syngo Auto Left Heart (Auto LH) Technology (NX3 Elite only)
- Note 14 syngo fourSight 4D imaging
- Note 15 Contrast Agent Imaging (NX3 Elite only)
- Note 16 syngo Auto follicle
- Note 17 eSie Touch (NX3 Elite only)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of Center for Devices and Radiological Health (CDRH) (Signature)

Concurrence of Center for Devices and Radiological Health (CDRH) (Signature)

510(k) _____

Diagnostic Ultrasound Indications for Use Form

510(k) Number (if known):

Device Name:

**CW2 Continuous Wave Doppler Transducer for use with:
 ACUSON NX3 Elite™ Diagnostic Ultrasound System**

Intended Use:

Diagnostic imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation							
Other (Track 1 Only)	Specific (Tracks 1I & 3)	B	M	PWD	CWD	Color Doppler	Power Doppler	Combined (Specify)	Other (Specify)
Ophthalmic	Ophthalmic								
Fetal Imaging & Other	Fetal								
	Abdominal								
	Intra-operative (Note 6)								
	Intra-operative (Neuro)								
	Laparoscopic								
	Pediatric								
	Small Organ (Note 1)								
	Neonatal Cephalic								
	Adult Cephalic								
	Trans-rectal								
	Trans-vaginal								
	Trans-urethral								
	Trans-esoph. (non-Card.)								
	Musculo-skel. (Convent.)								
	Musculo-skel. (Superfic)								
Intra-vascular									
Other (Specify)									
Cardiac	Cardiac Adult				P				
	Cardiac Pediatric				P				
	Intra-vascular (Cardiac)								
	Trans-esophageal (Cardiac)								
	Intra-cardiac								
Other (Specify)									
Peripheral Vessel	Peripheral vessel								
	Other (Specify)								

N = new indication; P = previously cleared by K141846

- Note 1 For example: breast, testes, thyroid, penis, prostate, etc.
- Note 2 Dynamic TCE Technology
- Note 3 SieClear
- Note 4 Advanced SieClear
- Note 5 3-Scape 3D Imaging
- Note 6 For example: abdominal, vascular
- Note 7 Stress Echo Imaging
- Note 8 Axius Edge Assisted Ejection Fraction
- Note 9 Clarify Vascular Enhancement Technology
- Note 10 SieScape Panoramic Imaging

- Note 11 syngo Arterial Health Package (AHP) (NX3 Elite only)
- Note 12 syngo Auto OB Measurements
- Note 13 syngo Auto Left Heart (Auto LH) Technology (NX3 Elite only)
- Note 14 syngo fourSight 4D imaging
- Note 15 Contrast Agent Imaging (NX3 Elite only)
- Note 16 syngo Auto follicle
- Note 17 eSie Touch (NX3 Elite only)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

 Concurrence of Center for Devices and Radiological Health (CDRH) (Signature)

 Concurrence of Center for Devices and Radiological Health (CDRH) (Signature)

510(k) _____

Diagnostic Ultrasound Indications for Use Form

510(k) Number (if known):

Device Name:

**CW5 Continuous Wave Doppler Transducer for use with:
 ACUSON NX3 Elite™ Diagnostic Ultrasound System**

Intended Use:

Diagnostic imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation							
Other (Track 1 Only)	Specific (Tracks 1I & 3)	B	M	PWD	CWD	Color Doppler	Power Doppler	Combined (Specify)	Other (Specify)
Ophthalmic	Ophthalmic								
Fetal Imaging & Other	Fetal								
	Abdominal								
	Intra-operative (Note 6)								
	Intra-operative (Neuro)								
	Laparoscopic								
	Pediatric								
	Small Organ (Note 1)								
	Neonatal Cephalic								
	Adult Cephalic					P			
	Trans-rectal								
	Trans-vaginal								
	Trans-urethral								
	Trans-esoph. (non-Card.)								
	Musculo-skel. (Convent.)								
	Musculo-skel. (Superfic)								
Intra-vascular									
Other (Specify)									
Cardiac	Cardiac Adult							P	
	Cardiac Pediatric							P	
	Intra-vascular (Cardiac)								
	Trans-esophageal (Cardiac)								
	Intra-cardiac								
Other (Specify)									
Peripheral Vessel	Peripheral vessel								
	Other (Specify)								

N = new indication; P = previously cleared by K141846

- Note 1 For example: breast, testes, thyroid, penis, prostate, etc.
- Note 2 Dynamic TCE Technology
- Note 3 SieClear
- Note 4 Advanced SieClear
- Note 5 3-Scape 3D Imaging
- Note 6 For example: abdominal, vascular
- Note 7 Stress Echo Imaging
- Note 8 Axis Edge Assisted Ejection Fraction
- Note 9 Clarify Vascular Enhancement Technology
- Note 10 SieScape Panoramic Imaging

- Note 11 syngo Arterial Health Package (AHP) (NX3 Elite only)
- Note 12 syngo Auto OB Measurements
- Note 13 syngo Auto Left Heart (Auto LH) Technology (NX3 Elite only)
- Note 14 syngo fourSight 4D imaging
- Note 15 Contrast Agent Imaging (NX3 Elite only)
- Note 16 syngo Auto follicle
- Note 17 eSie Touch (NX3 Elite only)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

 Concurrence of Center for Devices and Radiological Health (CDRH) (Signature)

 Concurrence of Center for Devices and Radiological Health (CDRH) (Signature)

510(k) _____

Diagnostic Ultrasound Indications for Use Form

510(k) Number (if known):

Device Name: **P4-2 Phased Sector Array Transducer for use with:
 ACUSON NX3 Elite™ Diagnostic Ultrasound System / ACUSON NX3™ Diagnostic Ultrasound System**

Intended Use: **Diagnostic imaging or fluid flow analysis of the human body as follows:**

Clinical Application		Mode of Operation							
Other (Track 1 Only)	Specific (Tracks 11 & 3)	B	M	PWD	CWD	Color Doppler	Power Doppler	Combined (Specify)	Other (Specify)
Ophthalmic	Ophthalmic								
Fetal Imaging & Other	Fetal	P	P	P		P	P	BMDC	Note 2,9
	Abdominal	P	P	P		P	P	BMDC	Note 2,9,16
	Intra-operative (Note 6)								
	Intra-operative (Neuro)								
	Laparoscopic								
	Pediatric	P	P	P		P	P	BMDC	Note 2,9
	Small Organ (Note 1)	P	P	P		P	P	BMDC	Note 2,9
	Neonatal Cephalic – NX3 Elite Only	P	P	P		P	P	BMDC	Note 2,9
	Adult Cephalic	P	P	P		P	P	BMDC	Note 2,9
	Trans-rectal								
	Trans-vaginal								
	Trans-urethral								
	Trans-esoph. (non-Card.)								
	Musculo-skel. (Convent.)								
	Musculo-skel. (Superfic)								
Intra-vascular									
Other (Specify)									
Cardiac	Cardiac Adult	P	P	P	P	P	P	BMDC	Note 2,7,8,9,13
	Cardiac Pediatric	P	P	P	P	P	P	BMDC	Note 2,7,8,9,13
	Intra-vascular (Cardiac)								
	Trans-esophageal (Cardiac)								
	Intra-cardiac								
	Other (Specify)								
Peripheral Vessel	Peripheral vessel								
	Other (Specify)								

N = new indication; P = previously cleared by K121699

- Note 1 For example: breast, testes, thyroid, penis, prostate, etc.
- Note 2 Dynamic TCE Technology
- Note 3 SieClear
- Note 4 Advanced SieClear
- Note 5 3-Scape 3D Imaging
- Note 6 For example: abdominal, vascular
- Note 7 Stress Echo Imaging
- Note 8 Axius Edge Assisted Ejection Fraction
- Note 9 Clarify Vascular Enhancement Technology
- Note 10 SieScape Panoramic Imaging

- Note 11 syngo Arterial Health Package (AHP) (NX3 Elite only)
- Note 12 syngo Auto OB Measurements
- Note 13 syngo Auto Left Heart (Auto LH) Technology (NX3 Elite only)
- Note 14 syngo fourSight 4D imaging
- Note 15 Contrast Agent Imaging (NX3 Elite only)
- Note 16 syngo Auto follicle
- Note 17 eSie Touch (NX3 Elite only)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of Center for Devices and Radiological Health (CDRH) (Signature)

Concurrence of Center for Devices and Radiological Health (CDRH) (Signature)

510(k) _____

Diagnostic Ultrasound Indications for Use Form

510(k) Number (if known):

Device Name:

**VF16-5 Transducer for use with:
 ACUSON NX3 Elite™ Diagnostic Ultrasound System**

Intended Use:

Diagnostic imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation							
Other (Track 1 Only)	Specific (Tracks 11& 3)	B	M	PWD	CWD	Color Doppler	Power Doppler	Combined (Specify)	Other (Specify)
Ophthalmic	Ophthalmic								
Fetal Imaging & Other	Fetal								
	Abdominal	P	P	P		P	P	BMDC	Note 2,3,4,9,10,16,17
	Intra-operative (Note 6)								
	Intra-operative (Neuro)								
	Laparoscopic								
	Pediatric	P	P	P		P	P	BMDC	Note 2,3,4,9,10
	Small Organ (Note 1)	P	P	P		P	P	BMDC	Note 2,3,4,9,10,17
	Neonatal Cephalic								
	Adult Cephalic	P	P	P		P	P	BMDC	Note 2,3,4,9,10,11
	Trans-rectal								
	Trans-vaginal								
	Trans-urethral								
	Trans-esoph. (non-Card.)								
	Musculo-skel. (Convent.)	P	P	P		P	P	BMDC	Note 2,3,4,9,10,17
	Musculo-skel. (Superfic)	P	P	P		P	P	BMDC	Note 2,3,4,9,10,17
Intra-vascular									
Other (Specify)									
Cardiac	Cardiac Adult								
	Cardiac Pediatric								
	Intra-vascular (Cardiac)								
	Trans-esophageal (Cardiac)								
	Intra-cardiac								
	Other (Specify)								
Peripheral Vessel	Peripheral vessel								
	Other (Specify)								

N = new indication; P = previously cleared by K141846

- Note 1 For example: breast, testes, thyroid, penis, prostate, etc.
- Note 2 Dynamic TCE Technology
- Note 3 SieClear
- Note 4 Advanced SieClear
- Note 5 3-Scape 3D Imaging
- Note 6 For example: abdominal, vascular
- Note 7 Stress Echo Imaging
- Note 8 Axius Edge Assisted Ejection Fraction
- Note 9 Clarify Vascular Enhancement Technology
- Note 10 SieScape Panoramic Imaging

- Note 11 syngo Arterial Health Package (AHP) (NX3 Elite only)
- Note 12 syngo Auto OB Measurements
- Note 13 syngo Auto Left Heart (Auto LH) Technology (NX3 Elite only)
- Note 14 syngo fourSight 4D imaging
- Note 15 Contrast Agent Imaging (NX3 Elite only)
- Note 16 syngo Auto follicle
- Note 17 eSie Touch (NX3 Elite only)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

 Concurrence of Center for Devices and Radiological Health (CDRH) (Signature)

 Concurrence of Center for Devices and Radiological Health (CDRH) (Signature)

510(k) _____

Diagnostic Ultrasound Indications for Use Form

510(k) Number (if known):

Device Name:

VF13-5sp Transducer for use with:
ACUSON NX3 Elite™ Diagnostic Ultrasound System

Intended Use:

Diagnostic imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation							
Other (Track 1 Only)	Specific (Tracks 11 & 3)	B	M	PWD	CWD	Color Doppler	Power Doppler	Combined (Specify)	Other (Specify)
Ophthalmic	Ophthalmic								
Fetal Imaging & Other	Fetal								
	Abdominal								
	Intra-operative (Note 6)	P	P	P		P	P	BMDC	Note 2,3,4,9,10,16,17
	Intra-operative (Neuro)								
	Laparoscopic								
	Pediatric	P	P	P		P	P	BMDC	Note 2,3,4,9,10
	Small Organ (Note 1)	P	P	P		P	P	BMDC	Note 2,3,4,9,10,17
	Neonatal Cephalic								
	Adult Cephalic	P	P	P		P	P	BMDC	Note 2,3,4,9,10,11
	Trans-rectal								
	Trans-vaginal								
	Trans-urethral								
	Trans-esoph. (non-Card.)								
	Musculo-skel. (Convent.)	P	P	P		P	P	BMDC	Note 2,3,4,9,10,17
	Musculo-skel. (Superfic)	P	P	P		P	P	BMDC	Note 2,3,4,9,10,17
	Intra-vascular								
Other (Specify)									
Cardiac	Cardiac Adult								
	Cardiac Pediatric								
	Intra-vascular (Cardiac)								
	Trans-esophageal (Cardiac)								
	Intra-cardiac								
	Other (Specify)								
Peripheral Vessel	Peripheral vessel (Note 1)	P	P	P		P	P		Note 2,3,4,9
	Other (Specify)								

N = new indication; P = previously cleared by K121699

- Note 1 For example: breast, testes, thyroid, penis, prostate, etc.
- Note 2 Dynamic TCE Technology
- Note 3 SieClear
- Note 4 Advanced SieClear
- Note 5 3-Scape 3D Imaging
- Note 6 For example: abdominal, vascular
- Note 7 Stress Echo Imaging
- Note 8 Axis Edge Assisted Ejection Fraction
- Note 9 Clarify Vascular Enhancement Technology
- Note 10 SieScape Panoramic Imaging

- Note 11 syngo Arterial Health Package (AHP) (NX3 Elite only)
- Note 12 syngo Auto OB Measurements
- Note 13 syngo Auto Left Heart (Auto LH) Technology (NX3 Elite only)
- Note 14 syngo fourSight 4D imaging
- Note 15 Contrast Agent Imaging (NX3 Elite only)
- Note 16 syngo Auto follicle
- Note 17 eSie Touch (NX3 Elite only)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

 Concurrence of Center for Devices and Radiological Health (CDRH) (Signature)

 Concurrence of Center for Devices and Radiological Health (CDRH) (Signature)

510(k) _____

Diagnostic Ultrasound Indications for Use Form

510(k) Number (if known):

Device Name: **C8F3 Transducer for use with:
 ACUSON NX3 Elite™ Diagnostic Ultrasound System / ACUSON NX3™ Diagnostic Ultrasound System**

Intended Use: Diagnostic imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation							
Other (Track 1 Only)	Specific (Tracks 11 & 3)	B	M	PWD	CWD	Color Doppler	Power Doppler	Combined (Specify)	Other (Specify)
Ophthalmic	Ophthalmic								
Fetal Imaging & Other	Fetal	P	P	P		P	P	BMDC	Note 2,3,4,5,6,9,10,12,14
	Abdominal	P	P	P		P	P	BMDC	Note 2,3,4,5,6,9,10,14,16
	Intra-operative (Note 6)								
	Intra-operative (Neuro)								
	Laparoscopic								
	Pediatric								
	Small Organ (Note 1)								
	Neonatal Cephalic								
	Adult Cephalic								
	Trans-rectal								
	Trans-vaginal								
	Trans-urethral								
	Trans-esoph. (non-Card.)								
	Musculo-skel. (Convent.)								
	Musculo-skel. (Superfic)								
	Intra-vascular								
Other (Specify)									
Cardiac	Cardiac Adult								
	Cardiac Pediatric								
	Intra-vascular (Cardiac)								
	Trans-esophageal (Cardiac)								
	Intra-cardiac								
	Other (Specify)								
Peripheral Vessel	Peripheral vessel								
	Other (Specify)								

N = new indication; P = previously cleared by K141846

- Note 1 For example: breast, testes, thyroid, penis, prostate, etc.
- Note 2 Dynamic TCE Technology
- Note 3 SieClear
- Note 4 Advanced SieClear
- Note 5 3-Scape 3D Imaging
- Note 6 For example: abdominal, vascular
- Note 7 Stress Echo Imaging
- Note 8 Aixus Edge Assisted Ejection Fraction
- Note 9 Clarify Vascular Enhancement Technology
- Note 10 SieScape Panoramic Imaging

- Note 11 syngo Arterial Health Package (AHP) (NX3 Elite only)
- Note 12 syngo Auto OB Measurements
- Note 13 syngo Auto Left Heart (Auto LH) Technology (NX3 Elite only)
- Note 14 syngo fourSight 4D imaging
- Note 15 Contrast Agent Imaging (NX3 Elite only)
- Note 16 syngo Auto follicle
- Note 17 eSie Touch (NX3 Elite only)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of Center for Devices and Radiological Health (CDRH) (Signature)

Concurrence of Center for Devices and Radiological Health (CDRH) (Signature)

510(k) _____

Diagnostic Ultrasound Indications for Use Form

510(k) Number (if known):

Device Name:

**C8-5 Transducer for use with:
 ACUSON NX3 Elite™ Diagnostic Ultrasound System**

Intended Use:

Diagnostic imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation							
Other (Track 1 Only)	Specific (Tracks 11 & 3)	B	M	PWD	CWD	Color Doppler	Power Doppler	Combined (Specify)	Other (Specify)
Ophthalmic	Ophthalmic								
Fetal Imaging & Other	Fetal								
	Abdominal	P	P	P		P	P	BMDC	Note 2,3,4,5,6,9,10,16
	Intra-operative (Note 6)								
	Intra-operative (Neuro)								
	Laparoscopic								
	Pediatric	P	P	P		P	P	BMDC	Note 2,3,4,5,6,9,10
	Small Organ (Note 1)								
	Neonatal Cephalic	P	P	P		P	P	BMDC	Note 2,3,4,5,6,9,10
	Adult Cephalic	P	P	P		P	P	BMDC	Note 2,3,4,5,6,9,10
	Trans-rectal								
	Trans-vaginal								
	Trans-urethral								
	Trans-esoph. (non-Card.)								
	Musculo-skel. (Convent.)								
	Musculo-skel. (Superfic)								
Intra-vascular									
Other (Specify)									
Cardiac	Cardiac Adult								
	Cardiac Pediatric	P	P	P		P	P	BMDC	Note 2,3,4,5,6,9,10
	Intra-vascular (Cardiac)								
	Trans-esophageal (Cardiac)								
	Intra-cardiac								
Other (Specify)									
Peripheral Vessel	Peripheral vessel								
	Other (Specify)								

N = new indication; P = previously cleared by K141846

- Note 1 For example: breast, testes, thyroid, penis, prostate, etc.
- Note 2 Dynamic TCE Technology
- Note 3 SieClear
- Note 4 Advanced SieClear
- Note 5 3-Scape 3D Imaging
- Note 6 For example: abdominal, vascular
- Note 7 Stress Echo Imaging
- Note 8 Axius Edge Assisted Ejection Fraction
- Note 9 Clarify Vascular Enhancement Technology
- Note 10 SieScape Panoramic Imaging

- Note 11 syngo Arterial Health Package (AHP) (NX3 Elite only)
- Note 12 syngo Auto OB Measurements
- Note 13 syngo Auto Left Heart (Auto LH) Technology (NX3 Elite only)
- Note 14 syngo fourSight 4D imaging
- Note 15 Contrast Agent Imaging (NX3 Elite only)
- Note 16 syngo Auto follicle
- Note 17 eSie Touch (NX3 Elite only)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

 Concurrence of Center for Devices and Radiological Health (CDRH) (Signature)

 Concurrence of Center for Devices and Radiological Health (CDRH) (Signature)

510(k) _____

Diagnostic Ultrasound Indications for Use Form

510(k) Number (if known):

Device Name:

P8-4 Transducer for use with:
ACUSON NX3 Elite™ Diagnostic Ultrasound System

Intended Use:

Diagnostic imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation							
Other (Track 1 Only)	Specific (Tracks 11 & 3)	B	M	PWD	CWD	Color Doppler	Power Doppler	Combined (Specify)	Other (Specify)
Ophthalmic	Ophthalmic								
Fetal Imaging & Other	Fetal								
	Abdominal	P	P	P		P	P	BMDC	Note 2,3,4,5,6,9,10,16
	Intra-operative (Note 6)								
	Intra-operative (Neuro)								
	Laparoscopic								
	Pediatric	P	P	P		P	P	BMDC	Note 2,3,4,5,6,9,10
	Small Organ (Note 1)								
	Neonatal Cephalic	P	P	P		P	P	BMDC	Note 2,3,4,5,6,9,10
	Adult Cephalic								
	Trans-rectal								
	Trans-vaginal								
	Trans-urethral								
	Trans-esoph. (non-Card.)								
	Musculo-skel. (Convent.)								
	Musculo-skel. (Superfic)								
	Intra-vascular								
Other (Specify)									
Cardiac	Cardiac Adult								
	Cardiac Pediatric	P	P	P		P	P	BMDC	Note 2,3,4,5,6,9,10
	Intra-vascular (Cardiac)								
	Trans-esophageal (Cardiac)								
	Intra-cardiac								
	Other (Specify)								
Peripheral Vessel	Peripheral vessel								
	Other (Specify)								

N = new indication; P = previously cleared by K141846

- Note 1 For example: breast, testes, thyroid, penis, prostate, etc.
- Note 2 Dynamic TCE Technology
- Note 3 SieClear
- Note 4 Advanced SieClear
- Note 5 3-Scape 3D Imaging
- Note 6 For example: abdominal, vascular
- Note 7 Stress Echo Imaging
- Note 8 Axius Edge Assisted Ejection Fraction
- Note 9 Clarify Vascular Enhancement Technology
- Note 10 SieScape Panoramic Imaging

- Note 11 syngo Arterial Health Package (AHP) (NX3 Elite only)
- Note 12 syngo Auto OB Measurements
- Note 13 syngo Auto Left Heart (Auto LH) Technology (NX3 Elite only)
- Note 14 syngo fourSight 4D imaging
- Note 15 Contrast Agent Imaging (NX3 Elite only)
- Note 16 syngo Auto follicle
- Note 17 eSie Touch (NX3 Elite only)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

 Concurrence of Center for Devices and Radiological Health (CDRH) (Signature)

 Concurrence of Center for Devices and Radiological Health (CDRH) (Signature)

510(k) _____

510(k) Summary
Prepared September 10, 2015

Sponsor: Siemens Medical Solutions, Inc.,
Ultrasound Division
685 East Middlefield Road
Mountain View, California 94043

Manufacturing Facility

SIEMENS LTD SEOUL
2nd -3rd floor, 143, Sunhwan-ro,
Jungwon-gu, Seongnam-si, Gyeonggi-do,
Republic of Korea

Contact Person: Shelly Pearce
Telephone: (650) 279-0134
Fax: (650) 694-5580

Device Name: ACUSON NX3 Elite™ Diagnostic Ultrasound System
ACUSON NX3™ Diagnostic Ultrasound System

Common Name: Diagnostic Ultrasound System with Accessories

Classification:

Regulatory Class: II
Review Category: Tier II
Classification Panel: Radiology

Ultrasonic Pulsed Doppler Imaging System	FR # 892.1550	Product Code 90-IYN
Ultrasonic Pulsed Echo Imaging System	FR # 892.1560	Product Code 90-IYO
Diagnostic Ultrasound Transducer	FR # 892.1570	Product Code 90-ITX

A. Legally Marketed Predicate Devices

The ACUSON NX3 Elite™ Diagnostic Ultrasound System and ACUSON NX3™ Diagnostic Ultrasound System are multi-purpose diagnostic ultrasound systems with accessories and proprietary software, and are substantially equivalent to our current product, the ACUSON X700™ Diagnostic Ultrasound System (K141846) and ACUSON X300™ Diagnostic Ultrasound System (K121699).

B. Device Description:

The Siemens ACUSON NX3 Elite™ and ACUSON NX3™ Diagnostic Ultrasound Systems are multi-purpose mobile, software controlled, diagnostic ultrasound systems with an on-screen display for thermal and mechanical indices related to potential bio-effect mechanisms. Its function is to acquire harmonic ultrasound echo data and display it in B-Mode, M-Mode, Pulsed (PW) Doppler Mode, Continuous (CW) Doppler Mode, Color Doppler Mode, Color M mode, Tissue Doppler Image, Amplitude Doppler Mode, a combination of modes, or Harmonic Imaging and 3D Imaging, or Harmonic Imaging and 4D imaging on a Flat Panel Display.

C. Intended Use

For ACUSON NX3

The ACUSON NX3 ultrasound imaging system, is intended for the following applications: Fetal, Abdominal (including liver), Pediatric, Small Parts(Small Organ), Transcranial, OB/GYN, Pelvic, Neonatal, Cardiac, Vascular(including Peripheral Vessel), Musculoskeletal, Superficial Musculoskeletal and Urology applications. The systems also provide for the measurement of anatomical structures and for analysis packages that provide information that is used for clinical diagnosis purposes.

Ultrasound is used as an imaging aid, but may have further restrictions specific to in-vitro fertilization (IVF), chorionic villus sampling (CVS), and percutaneous umbilical cord blood sampling (PUBS) procedures. Observe local laws and regulations.

For ACUSON NX3 Elite

ACUSON NX3 Elite ultrasound imaging systems are intended for the following applications: Fetal, Abdominal (including liver, intra-operative), Pediatric, Small Parts(Small Organ including intra-operative), Transcranial, OB/GYN, Pelvic, Neonatal, Cardiac, Vascular(including Peripheral Vessel, intra-operative), Musculoskeletal, Superficial Musculoskeletal and Urology applications. The systems also provide for the measurement of anatomical structures and for analysis packages that provide information that is used for clinical diagnosis purposes.

Ultrasound is used as an imaging aid, but may have further restrictions specific to in-vitro fertilization (IVF), chorionic villus sampling (CVS), and percutaneous umbilical cord blood sampling (PUBS) procedures. Observe local laws and regulations.

The Arterial Health Package (AHP) software provides the physician with the capability to measure Intima Media Thickness and the option to reference normative tables that have been validated and published in peer-reviewed studies. The information is intended to provide the physician with an easily understood tool for communicating with patients regarding state of their cardiovascular system. Note: This feature can be utilized according to the "ASE Consensus Statement; Use of Carotid Ultrasound to Identify Subclinical Vascular Disease and Evaluate Cardiovascular Disease Risk: A Consensus Statement from the American Society of Echocardiography Carotid Intima-Media Thickness Task Force. Endorsed by the Society for Vascular Medicine."

D. Substantial Equivalence

The Siemens ACUSON NX3 Elite™ and ACUSON NX3™ Diagnostic Ultrasound Systems are multi-purpose diagnostic ultrasound systems with accessories and proprietary software, and are substantially equivalent to our current products, the Siemens ACUSON X700™ Diagnostic Ultrasound System (K141846), and ACUSON X300™ Diagnostic Ultrasound System (K121699). All systems transmit ultrasonic energy into patients, and then perform post processing of received echoes to generate onscreen display of anatomic structures and fluid flow within the body. All systems allow for specialized measurements of structures and flow, and calculations.

The submission devices are substantially equivalent to the predicate with regard to both intended use and technological characteristics.

Feature / Characteristic	Predicate Device ACUSON X300™ (K121699)	Predicate Device ACUSON X700™ (K141846)	Submission Device ACUSON NX3™	Submission Device ACUSON NX3 Elite™
Indications for Use:				
▪ Fetal Echo	√	√	√	√
▪ Abdominal	√	√	√	√
▪ Intra-operative	√	√	-	√
▪ Neonatal Cephalic	√	√	-	√
▪ Small Organ(parts)	√	√	√	√
▪ Pediatric	√	√	√	√
▪ Adult Cephalic (Trans-cranial)	√	√	√	√
▪ Cardiac (Adult)	√	√	√	√
▪ Cardiac (Pediatric)	√	√	√	√
▪ Intracardiac	√	√	-	-
▪ Trans-esophageal	√	√	-	-
▪ Trans-rectal	√	√	√	√
▪ Trans-vaginal	√	√	√	√
▪ Peripheral vessel	√	√	√	√
▪ Musculo-skeletal (conventional)	√	√	√	√
▪ Musculo-skeletal (superficial)	√	√	√	√
Center Frequencies Supported:				
▪ 2.0 MHz	√	√	√	√
▪ 2.5 MHz	√	√	√	√
▪ 3.0 MHz	√	√	√	√
▪ 3.5 MHz	√	√	√	√
▪ 4.0 MHz	√	√	√	√
▪ 5.0 MHz	√	√	√	√
▪ 5.5 MHz	√	√	√	√
▪ 6.0 MHz	√	√	√	√
▪ 6.5 MHz	√	√	√	√
▪ 7.5 MHz	√	√	√	√
▪ 8.0 MHz	√	√	√	√
▪ 9.0 MHz	√	√	√	√
▪ 10.0 MHz	√	√	√	√
▪ 11.0 MHz	√	√	√	√
▪ 12.0 MHz	-	√	-	√
▪ 13.0 MHz	-	√	-	√
Modes:				
▪ B	√	√	√	√
▪ M	√	√	√	√
▪ PWD (Pulsed Wave Doppler)	√	√	√	√

Feature / Characteristic	Predicate Device ACUSON X300™ (K121699)	Predicate Device ACUSON X700™ (K141846)	Submission Device ACUSON NX3™	Submission Device ACUSON NX3 Elite™
▪ CWD (Continuous Wave Doppler)	√	√	-	√
▪ SCW (Steerable CW)	√	√	√	√
▪ CD (Color Doppler)	√	√	√	√
▪ Amplitude Doppler (Power Doppler)	√	√	√	√
▪ Directional Power Doppler	√	√	√	√
▪ Combined (BM, BC, BCM, BCD)	√	√	√	√
▪ THI (Tissue Harmonic Imaging)	√	√	√	√
▪ AMM (Anatomical M-mode)	√	√	√	√
▪ Doppler Tissue Image (Color, PW)	√	√	√	√
▪ M-THI	√	√	√	√
Features:				
▪ Multi-View Spatial Compounding (SieClear)	√	√	√	√
▪ Advanced SieClear	-	√	√	√
▪ DTCE (Dynamic Tissue Contrast Enhancement)	√	√	√	√
▪ Tissue Grayscale Optimization (TGO)	√	√	√	√
▪ Dual-Beam Processing	√	√	√	√
▪ Quad-Beam Processing	√	√	√	√
▪ Clip Capture	√	√	√	√
▪ 3D Imaging (3-Scape)	√	√	√	√
▪ 3D Measurements	√	√	√	√
▪ 4D Basic Imaging (fourSight 4D)	√	√	√	√
▪ Panoramic 2D Imaging (SieScape)	√	√	√	√
▪ Syngo Auto OB	√	√	√	√
▪ Syngo Auto Follicle	-	√	√	√

Feature / Characteristic	Predicate Device ACUSON X300™ (K121699)	Predicate Device ACUSON X700™ (K141846)	Submission Device ACUSON NX3™	Submission Device ACUSON NX3 Elite™
▪ Cardiac Imaging physiological signal display multiplane TEE fourSight TEE Imaging	√	√	-	-
▪ Stress Echo	√	√	√	√
▪ Vascular Enhancement (Clarify VE)	√	√	√	√
▪ Syngo VVI (Velocity Vector Image)	√	√	-	-
▪ Auto Left Heart	√	√	-	√
▪ Axius EF	√	√	√	√
▪ Syngo AHP	√	√	-	√
▪ Contrast Agent Image	√	√	-	√
▪ DIMAQ	√	√	√	√
▪ Multiple Frequency Imaging(MultiHertz)	√	√	√	√
▪ syngo Mitral Valve Assessments (MVA)	√	√	-	-
▪ Intracardiac Echocardiography (ICE) Imaging	√	√	-	-
▪ Digital Architecture	√	√	√	√
▪ Fully integrated DICOM	√	√	√	√
▪ CARTOSOUND™ Communication	√	√	-	-
▪ Wireless	√	√	√	√
▪ eSie Touch	-	√	-	√
▪ Monitor: FPD	(17" FPD)	√ (20" FPD)	√ (21.5" FPD)	√ (21.5" FPD)
# Channels	128	128	64	64
Output Display Standard (Track 3)	√	√	√	√
Patient Contact Materials	Tested to ISO 10993-1	Tested to ISO 10993-1	Tested to ISO 10993-1	Tested to ISO 10993-1
UL60601-1 Certified	√	√	√	√

E. A brief discussion of nonclinical tests submitted, referenced, or relied on in the 510(k) for a determination of substantial equivalence

The devices have been evaluated for acoustic output, biocompatibility, cleaning and disinfection effectiveness as well as thermal, electrical, electromagnetic and mechanical safety and have been found to conform with applicable medical device safety standards. The systems comply with the following voluntary standards:

- UL 60601-1, Safety Requirements for Medical Equipment
- IEC 60601-2-37 Diagnostic Ultrasound Safety Standards
- CSA C22.2 No. 601-1, Safety Requirements for Medical Equipment
- AIUM/NEMA UD-3, Standard for Real Time Display of Thermal and Mechanical Acoustic Output Indices on Diagnostic Ultrasound Equipment
- AIUM/NEMA UD-2, Acoustic Output Measurement Standard for Diagnostic Ultrasound
- IEC 62359, Test methods for the determination of thermal and mechanical indices
- 93/42/EEC Medical Devices Directive
- Safety and EMC Requirements for Medical Equipment
 - IEC 60601-1
 - IEC 60601-1-2
 - IEC 60601-2-18
 - IEC 60601-1-6
 - IEC 62366
- ISO 10993-1 Biocompatibility

F. A summary discussion of the clinical tests submitted, referenced, or relied on for a determination of substantial equivalence.

Since the ACUSON NX3 Elite™ and ACUSON NX3™ Diagnostic Ultrasound Systems use the same technology and principles as existing devices, clinical data is not required.

G. Summary

Intended uses and other key features are consistent with traditional clinical practice and FDA guidelines. The design and development process of the manufacturer conforms with 21 CFR 820 Quality System Regulation and ISO 13485:2003 quality system standards. The products are designed to conform with applicable medical device safety standards and compliance is verified through independent evaluation with ongoing factory surveillance. Diagnostic ultrasound system has accumulated a long history of safe and effective performance. Therefore it is the opinion of Siemens Medical that the ACUSON NX3 Elite™ and ACUSON NX3™ Diagnostic Ultrasound Systems are substantially equivalent with respect to safety and effectiveness to devices currently cleared for market.

The ACUSON NX3 Elite™ and ACUSON NX3™ Diagnostic Ultrasound Systems are verified and validated according to the company's design control process.