



Food and Drug Administration  
10903 New Hampshire Avenue  
Document Control Center – WO66-G609  
Silver Spring, MD 20993-0002

MicroPort Orthopedics, Incorporated  
Mr. Usman Rashid  
Regulatory Affairs Specialist  
5677 Airline Road  
Arlington, Tennessee 38002

March 23, 2016

Re: K152631

Trade/Device Name: MPO Total Knee Systems MR Labeling

Regulation Number: 21 CFR 888.3530

Regulation Name: Knee joint femorotibial metal/polymer semi-constrained cemented prosthesis.

Regulatory Class: Class II

Product Code: HRY, JWH, MBH

Dated: February 24, 2016

Received: February 26, 2016

Dear Mr. Rashid:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Parts 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set

forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

**Mark N. Melkerson -S**

Mark N. Melkerson  
Director  
Division of Orthopedic Devices  
Office of Device Evaluation  
Center for Devices and  
Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)

K152631

Device Name

MPO Total Knee Systems MR Labeling

Indications for Use (Describe)

MPO Total Knee Systems are indicated for use in knee arthroplasty in skeletally mature patients with the following conditions:

- 1) noninflammatory degenerative joint disease including osteoarthritis, traumatic arthritis, or avascular necrosis;
- 2) inflammatory degenerative joint disease including rheumatoid arthritis;
- 3) correction of functional deformity;
- 4) revision procedures where other treatments or devices have failed; and treatment of fractures that are unmanageable using other techniques.

ADVANCE® 913 Medial Pivot Tibial Base and Insert Components (not licensed for sale in Canada) are for use with bone cement.

Porous-Coated Total Knee Replacement Components are for use without bone cement.

The EVOLUTION® Total Knee System is for cemented use only.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

**PLEASE DO NOT WRITE BELOW THIS LINE – CONTINUE ON A SEPARATE PAGE IF NEEDED.**

### FOR FDA USE ONLY

Concurrence of Center for Devices and Radiological Health (CDRH) (Signature)

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
PRAStaff@fda.hhs.gov

*“An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number.”*

## 510(k) Summary of Safety and Effectiveness

In accordance with the Food and Drug Administration Rule to implement provisions of the Safe Medical Devices Act of 1990 and in conformance with 21 CFR 807.92, this information serves as a Summary of Safety and Effectiveness for the use of the MicroPort Orthopedics Total Knee systems in an MRI environment.

**Submitted by:** MicroPort Orthopedics Inc.  
5677 Airline Rd  
Arlington, TN 38002  
Phone: (866) 872-0211  
Fax: (855) 446-2247

**Date:** March 15, 2016

**Contact Person:** Usman Rashid  
*Regulatory Affairs Specialist*

**Proprietary Name of Modified Device:** MPO Total Knee Systems MR Labeling

**Common Name:** MPO Total Knee Systems

**Classification Name and Reference:** 21 CFR 888.3565 Knee joint  
Patellofemorotibial Metal/Polymer Porous-Coated Uncemented Prosthesis  
Class II

21 CFR 888.3560 Knee joint  
Patellofemorotibial  
Polymer/Metal/Polymer Semi-Constrained  
Cemented Prosthesis  
Class II

21 CFR 888.3530 Knee joint Femorotibial  
Metal/Polymer Semi-Constrained  
Cemented Prosthesis  
Class II

**Subject Product Code and Panel Code:** Orthopedics/87/MBH, JWH, HRY

**Predicate Devices:** EVOLUTION® Revision Stemmed Femur (K142550)

EVOLUTION® MP CS/CR Porous Femur & EVOLUTION® Adaptive CS and PS Inserts (K140735)

EVOLUTION® MP Adaptive PS Tibial  
Insert (K131679)

ADVANCE® TOTAL KNEE SYSTEM-  
PATELLA (K122218)

EVOLUTION® MP Adaptive CS Insert  
(K113325)

EVOLUTION® MP Total Knee System  
(K093552)

ADVANCE STATURE® Femoral  
Component (K063731)

ADVANCE® Spiked Porous Tibial Base  
(K063128)

ADVANCE® TOTAL KNEE SYSTEM  
(K061223)

ADVANCE® HA Coated Components  
(K043083)

ADVANCE® Double High Insert  
(K033890)

ADVANCE® Revision Product Line  
Extension (K990030)

ADVANCE® Modular Tibial Component  
(K973524)

ADVANCE® KNEE SYSTEM (K972626)

ADVANCE® TIBIAL COMPONENT  
(K960617)

ULTRACK TOTAL KNEE SYSTEM  
(K953439)

ORTHOLOC® ADVANTIM™ Tibial Base  
Component (K932858)

ORTHOLOC® ADVANTIM™ 5 Degree  
Tibial Stem (K930228)

SECOND GENERATION KNEE SYSTEM  
(K894334)

## **DEVICE INFORMATION**

### **A. Intended Use**

MicroPort Total Knee Systems are indicated for use in knee arthroplasty in skeletally mature patients with the following conditions:

- 1) noninflammatory degenerative joint disease including osteoarthritis, traumatic arthritis, or avascular necrosis;
- 2) inflammatory degenerative joint disease including rheumatoid arthritis;
- 3) correction of functional deformity;
- 4) revision procedures where other treatments or devices have failed; and treatment of fractures that are unmanageable using other techniques.

**ADVANCE® 913 Medial Pivot Tibial Base and Insert Components** (not licensed for sale in Canada) are for use with bone cement.

**Porous-Coated Total Knee Replacement Components** are for use without bone cement.

The **EVOLUTION® Total Knee System** is for cemented use only.

### **B. Device Description**

The subject devices for this submission are all the predicate devices listed above, which consists of implant components used in knee arthroplasty. The only changes to the subject devices are updates to their labeling. Specifically, the package inserts and package labels are being updated to include MR Conditional language and symbols. The subjects are identical to the predicates in all aspects except for the labeling updates. Testing is provided in this Traditional 510(k) that establishes the conditional safety and compatibility of the passive implants in a magnetic resonance (MR) environment. The basic design features are the following:

- Metal Femoral component manufactured from cobalt chrome alloy
- Femoral components available in porous and non porous versions
- Tibial inserts manufactured from UHMWPE
- Tibial bases manufactured from cobalt chrome alloy or titanium alloy
- Tibial bases available in porous and non porous versions
- All-poly patellae manufactured from UHMWPE
- Metal portion of metal backed patellae manufactured from titanium alloy

### **C. Substantial Equivalence Information**

Since this submission is only regarding device labelling change, the design features and materials of the subject devices are identical to those of the predicate devices. The indications for use are identical to the predicate devices. The safety and effectiveness of the subject devices are adequately supported by the substantial equivalence information, materials information, and analysis data provided within this Premarket Notification.

#### **D. Nonclinical Testing**

Non clinical Testing was conducted to establish the conditional safety and compatibility of the passive implants in a magnetic resonance (MR) environment according to the recommendations provided in the guidance document “Establishing Safety and Compatibility of Passive Implants in the Magnetic Resonance (MR) Environment” issued on December 11, 2014. Testing was also conducted according to the following standards:

ASTM F2052-14, “Standard test method for measurement of magnetically induced displacement force on passive implants in the magnetic resonance environment”;

ASTM F2119-7 “Standard Test Method for Evaluation of MR Image Artifacts from Passive Implants”;

ASTM F2503-13 “Standard Practice for Marking Medical Devices and Other Items for Safety in the Magnetic Resonance Environment”;

ASTM F2182-11a “Standard Test Method for Measurement of Radio Frequency Induced Heating near Passive Implants During Magnetic Resonance Imaging”

The tests determined the effects of the MRI on the implants, and the effects of the implants on the image quality. The tests evaluated the worst case components and constructs for RF Heating, field interactions, and image artifacts. The testing concluded that there are no safety issues related to magnetic field interactions under specific conditions identified in the labeling.

#### **E. Clinical Testing**

Clinical data was not provided for the subject devices.

#### **F. Conclusion**

All the information provided in this submission adequately supports the substantial equivalence of the labelling change.