



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center - WO66-G609
Silver Spring, MD 20993-0002

Zimmer, Incorporated
Sujith Kallur
Regulatory Affairs Specialist
P.O. Box 708
Warsaw, Indiana 46851-0708

December 21, 2015

Re: K152755

Trade/Device Name: *FastFrame*TM External Fixation System-Ankle Spanning
Regulation Number: 21 CFR 888.3030
Regulation Name: Single/multiple component metallic bone fixation appliances and accessories
Regulatory Class: Class II
Product Code: KTT, JDW
Dated: September 23, 2015
Received: September 24, 2015

Dear Sujith Kallur:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in

the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

Mark N. Melkerson -S

Mark N. Melkerson
Director
Division of Orthopedic Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

DEPARTMENT OF HEALTH AND HUMAN SERVICES
Food and Drug Administration

Form Approved: OMB No. 0910-0120
Expiration Date: January 31, 2017
See PRA Statement below.

Indications for Use

510(k) Number (if known)

K152755

Device Name

FastFrame External Fixation System - Ankle Spanning

Indications for Use (Describe)

The FastFrame External Fixation System - Ankle Spanning is indicated for use in treatment of distal tibia and foot fractures. Specifically, the system is intended for:

- Stabilization of open or closed fractures about the ankle, typically in the context of polytrauma or where open or alternative closed treatment is undesirable or otherwise contraindicated;
- Arthrodesis and osteotomies with associated soft tissue problems about the ankle;
- Stabilization of limbs after removal of total ankle arthroplasty for infection or other failure;
- Stabilization of non-unions about the ankle; and
- Intraoperative temporary stabilization to assist with indirect reduction.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."



510(k) Summary

Sponsor: Zimmer, Inc.
P.O. Box 708
Warsaw, IN 46581-0708

Contact Person: Sujith M Kallur
Regulatory Affairs Specialist, Trauma Regulatory Affairs
Telephone: 574-453-6350
Fax: 574-871-8760

Date: 09/23/2015

Trade Name: *FastFrame*TM External Fixation System – Ankle Spanning

Common Name: External Fixation Frame Components

Classification Names and References: Appliance, Fixation, Nail/Blade/Plate Combination, Multiple Components (KTT) per 21 CFR 888.3030, Single/multiple component metallic bone fixation appliances and accessories and Pin, Fixation, Threaded (JDW) per 21 CFR 888.3040, Smooth or threaded metallic bone fixation fastener

Classification Panel: Orthopedics/87

Predicate Device(s): *XtraFix* External Fixation System by Zimmer Inc. K113383, cleared 6/26/2012

Purpose and Device Description: The purpose of this 510(k) is to obtain marketing clearance for the new device, the *FastFrame* External Fixation System – Ankle Spanning. The system consists of fixation half- pins and transfixing full-pins attached to rigid clamps connected by adjustable telescoping tubes (bars). The system is intended for use in the treatment of fractures about the ankle that require external fixation. The system also contains a separately sterile packaged *FastFrame* External Fixation System – Ankle Spanning Optional Clamp Assembly. This Optional Clamp Assembly is included as an optional accessory that allows the surgeon to hold the foot in a fixed position post-operatively, if desired.

Intended Use/Indications for Use: The *FastFrame* External Fixation System – Ankle Spanning is indicated for use in treatment of distal tibia and foot fractures. Specifically, the system is intended for:

- Stabilization of open or closed fractures about the Ankle, typically in the context of polytrauma or where open or alternative closed treatment is undesirable or otherwise contraindicated;
- Arthrodesis and osteotomies with associated soft tissue problems about the ankle;
- Stabilization of limbs after removal of total ankle arthroplasty for infection or other failure;
- Stabilization of non-unions about the ankle; and
- Intraoperative temporary stabilization to assist with indirect reduction.

Comparison to Predicate Device: Both the subject and predicate systems can be used to reduce and fix tibial and foot anatomy. Both systems use bars, clamps, and fixation pins to achieve the clinical end result. Both systems allow for length distraction and polyaxial motion between clamp bodies.

The major difference between the subject and predicate systems is that predicate system frame needs to be built intra-operatively. The *FastFrame* External Fixation System contains a frame which comes with clamps and bars (telescoping tubes) pre-assembled, and are not disassemblable by the end user. The subject *FastFrame* External Fixation System is provided in sterile kits as compared to the predicate system that is provided non-sterile, and must be steam sterilized prior to use.

The *FastFrame* External Fixation System – Ankle Spanning Optional Clamp Assembly was proven to be able to hold a clinically relevant load over a specified period of time.

Performance Data (Nonclinical and/or Clinical):

Non-Clinical Performance and Conclusions:

- **Performance Evaluation** – The mechanical testing confirmed that the subject devices perform substantially equivalent in full construct rigidity and interconnection performance as compared to the predicate device. In addition, the *FastFrame* External

Fixation System – Ankle Spanning and FastFrame External Fixation System – Ankle Spanning when used with the FastFrame External Fixation System – Ankle Spanning Optional Clamp Assembly were found to be MRI Conditional per FDA Guidance “Establishing Safety and Compatibility of Passive Implants in the Magnetic Resonance (MR) Environment” issued on December 11, 2014.

Conclusions: The results demonstrate that the devices are safe and effective and substantially equivalent to the predicate devices.

Clinical Performance and Conclusions:

Clinical data and conclusions were not needed for these devices to show substantial equivalence.