



Food and Drug Administration 10903 New Hampshire Avenue Document Control Center -- WO66-G609 Silver Spring, MD 20993-0002

#### September 29, 2016

NeoMed, Inc. Melinda Harrison Smith, RAC, ASQ-CBA Director, Quality and Regulatory Affairs 100 Londonderry Court, Suite 112 Woodstock, GA 30188

Re:

K152857

Trade/Device Name: NeoMed NeoConnect<sup>TM</sup> Enteral Syringes with ENFit<sup>TM</sup> Connector

and compatible NeoSecure™ Tip Caps

Regulation Number: 21 CFR §876.5980

Regulation Name: Gastrointestinal Tube and accessories

Regulatory Class: II Product Code: PNR

Dated: October 16, 2015 Received: October 19, 2015

Dear Melinda Harrison Smith,

This letter corrects our substantially equivalent letter of December 17, 2015.

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the <u>Federal Register</u>.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm.

Sincerely,

For Division

Douglas Silverstein -S 2016.09:29 08:07:09 -04'00

Benjamin R. Fisher, Ph.D.
Director
Division of Reproductive, Gastro-Renal,
and Urological Devices
Office of Device Evaluation
Center for Devices and Radiological Health

Enclosure

# DEPARTMENT OF HEALTH AND HUMAN SERVICES Food and Drug Administration

#### Indications for Use

Form Approved: OMB No. 0910-0120 Expiration Date: January 31, 2017 See PRA Statement below.

510(k) Number (if known)
K152857
Device Name NeoMed NeoConnect™ Enteral Syringes with ENFit™ Connector and compatible NeoSecure™ Tip Caps
Indications for Use (Describe)
NeoConnect <sup>TM</sup> Enteral Syringe with ENFit <sup>TM</sup> Connector:
The device is indicated for use as a dispenser, a measuring device and a fluid transfer device. It is used to deliver fluids into the body enterally. It is intended to be used in clinical or home care settings by users ranging from clinicians to laypersons (under the supervision of a clinician) in all age groups.
NeoConnect <sup>TM</sup> NeoSecure <sup>TM</sup> Tip Cap:
The Comment The Cap.
A NeoConnect <sup>TM</sup> Enteral Syringe with ENFit <sup>TM</sup> connector accessory used to prevent fluid loss and contamination of syringe contents until ready for use.
Type of Use (Select one or both, as applicable)
Prescription Use (Part 21 CFR 801 Subpart D) Over-The-Counter Use (21 CFR 801 Subpart C)

## CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

#### \*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\*

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services Food and Drug Administration Office of Chief Information Officer Paperwork Reduction Act (PRA) Staff PRAStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

#### TRADITIONAL 510(K) SUMMARY (21 CFR § 807.92)

I. SUBMITTER NeoMed, Inc.

100 Londonderry Court

Suite 112

Woodstock, GA 30188 Tel: 770-516-2225 Fax: 770-516-2448

Contact: Melinda Harrison Smith, RAC, CBA

mharrison@neomedinc.com

Date Prepared: 16 December 2015

Establishment 3006520777

Registration Number:

II. DEVICE

Trade Name: NeoMed NeoConnect™ Enteral Syringes with ENFit™

Connector and compatible NeoSecure<sup>™</sup> Tip Caps

Common Name: Enteral Syringe

Classification Name: Gastrointestinal tube and accessories (21 CFR § 876.5980)

Regulatory Class:

Product Code: PIF

#### III. PREDICATE DEVICE

A. NeoMed NeoConnect™ Enteral Syringes with ENFit Connector (K143344)

B. Covidien Kangaroo™ Enteral Feeding Syringes with ENFit Connector (K142128) These predicates have not been subject to a design-related recall.

#### IV. DEVICE DESCRIPTION

The NeoMed NeoConnect™ Enteral Syringes with ENFit™ connector are standard piston style syringes consisting of a syringe barrel with an integral ENFit™ syringe tip, syringe plunger, syringe gasket, and supplied with or without a syringe tip cap. They are provided in varying sizes ranging from 0.5 mL to 100 mL nominal capacity. The integral syringe tip is a female ENFit™ connector which is compatible only with enteral access devices having ENFit™ male connectors to form a dedicated system that prevents wrong-route administration of fluids. They possess translucent barrels to provide visualization of fluid contents and volume.

The NeoConnect<sup>™</sup> NeoSecure <sup>™</sup> Tip Cap is supplied with the NeoConnect <sup>™</sup> Enteral Syringe with ENFit<sup>™</sup> connector or sold separately (sterile or non-sterile). It is an ENFit<sup>™</sup> compatible closure cap for the tip of the syringe.

#### V. INDICATIONS FOR USE

This 510(k) changes the indications for use of K143344 to extend the patient population to include all ages and extend the use environment to include home use.

NeoConnect™ Enteral Syringe with ENFit™ connector:

The device is indicated for use as a dispenser, a measuring device and a fluid transfer device. It is used to deliver fluids into the body enterally. It is intended to be used in clinical or home care settings by users ranging from clinicians to laypersons (under the supervision of a clinician) in all age groups.

NeoConnect™ NeoSecure™ Tip Cap:

A NeoConnect™ Enteral Syringe with ENFit™ connector accessory used to prevent fluid loss and contamination of syringe contents until ready for use.

# VI. COMPARISON OF TECHNOLOGICAL CHARACTERISTICS WITH THE PREDICATE

The NeoMed NeoConnect™ Enteral Syringes with ENFit™ connector and the NeoSecure™ compatible Tip Caps have the same intended use, principles of operation, materials, and device designs as the predicate device K143344. The NeoMed NeoConnect™ Enteral Syringes with ENFit™ connector have similar indications for use as the predicate device K143344.

The NeoMed NeoConnect™ Enteral Syringes with ENFit™ connector have the same intended use, principles of operation, patient population and use environment as the predicate device K142128.

### VII. PERFORMACE DATA (BENCH)

The following performance testing has been conducted:

- Finished Device
  - Risk Analysis including design, user and process FMEA (Failure Modes and Effects Analysis) in accordance with EN ISO 14971:2012
  - o Human Factors and Usability Validation
  - Biocompatibility
    - ISO 10993-5: Cytotoxicity
    - ISO 10993-10: Irritation and sensitization
    - ISO 10993-11: Acute Toxicity
  - Chemical Testing
    - Extractables and Leachables
  - Finished Device Verification Testing
    - Critical Dimension verification
    - Ink Adhesion
    - ISO 7886
      - Capacity Tolerance
      - Graduated Scale

K152857

Page **2** of **3** 

- Piston Fit in Barrel
- Air and Liquid Leakage Testing
- o Device Performance and Packaging Stability Testing (3 years)
- Syringe Tip (ISO 80369-3 (ENFit™) female connector)
  - o Enteral Connector Misconnection Assessment
  - ENFit Connector Risk Management Report (including misconnections FMEA)
  - o Human Factors Validation Study
  - o Dimensional verification
  - o Liquid Leakage Testing
  - o Stress Cracking
  - o Resistance to separation from axial load
  - Resistance to separation form unscrewing
  - Resistance to overriding
  - o Disconnection by unscrewing

#### VIII. CONCLUSIONS

The NeoMed NeoConnect™ Enteral Syringes with ENFit™ Connector and compatible tip caps are substantially equivalent to the NeoMed NeoConnect™ Enteral Syringe with ENFit Connector (K143344) and the Covidien Kangaroo™ Feeding Syringes with ENFit Connector.