



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center - WO66-G609
Silver Spring, MD 20993-0002

Intuitive Surgical Incorporated
Vishal Kanani
Senior Regulatory Affairs Specialist
1266 Kifer Road
Sunnyvale, California 94086

December 9, 2015

Re: K153126
Trade/Device Name: da Vinci Xi Hasson Cone
Regulation Number: 21 CFR 876.1500
Regulation Name: Endoscope and Accessories
Regulatory Class: Class II
Product Code: GCJ
Dated: October 26, 2015
Received: October 29, 2015

Dear Mr. Kanani:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you; however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

Joshua C. Nipper -S

For Binita S. Ashar, M.D., M.B.A., F.A.C.S.
Director
Division of Surgical Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

DEPARTMENT OF HEALTH AND HUMAN SERVICES Food and Drug Administration Indications for Use	Form Approved: OMB No. 0910-0120 Expiration Date: January 31, 2017 See PRA Statement below.
---	---

510(k) Number (if known)
K153126

Device Name
da Vinci Xi Hasson Cone

Indications for Use (Describe)
The da Vinci Xi Hasson Cone has applications in laparoscopic surgery to establish a port of entry for Intuitive Surgical da Vinci Xi EndoWrist Instruments, endoscopes, or compatible accessories.

Type of Use (Select one or both, as applicable)
 Prescription Use (Part 21 CFR 801 Subpart D) Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
 Food and Drug Administration
 Office of Chief Information Officer
 Paperwork Reduction Act (PRA) Staff
 PRAStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

da Vinci Xi Hasson Cone

Traditional 510(k)

510(k) Owner: Intuitive Surgical, Inc.
1266 Kifer Road
Sunnyvale, CA 94086

Contact: Vishal Kanani
Sr. Regulatory Affairs Specialist
Phone Number: 408-523-2035
Fax Number: 408-523-8907
Email: vishal.kanani@intusurg.com

Date Summary Prepared: October 27, 2015

Trade Name: *da Vinci Xi* Hasson Cone

Common Name: Endoscope and Accessories

Classification: Class II
21 CFR 876.1500, Endoscope and Accessories

Product Code: GCJ

Classification Advisory

Committee: General and Plastic Surgery

Predicate Device: Hybrid Trocar System (K101937)

Device Description:

The *da Vinci Xi* Hasson Cone is a cone-shaped device which is intended to be used in endoscopic surgery to anchor a port of entry (i.e., cannula) to the patient. During Hasson, or open, surgical techniques, the user places a Hasson Cone in the opening of the body wall. The cannula shaft is inserted into the cone, effectively increasing the diameter of the cannula to the contour of the incision made in the body wall.

The Hasson Cone will be offered in two sizes; 8 mm and 12 mm & Stapler.

Intended Use:

The *da Vinci Xi* Hasson Cone is intended to be used in laparoscopic surgery to anchor a port of entry (i.e., cannula) to the patient.

Indications for Use:

The *da Vinci Xi* Hasson Cone has applications in laparoscopic surgery to establish a port of entry for *Intuitive Surgical da Vinci Xi EndoWrist* Instruments, endoscopes, or compatible accessories.

Technological Characteristics:

The subject device, *da Vinci Xi* Hasson Cone, is technologically very similar to the predicate device, Aesculap 10 mm Stability Cone for Trocar Sleeve (cleared under K101937). It has the same intended use, same fundamental scientific technology, and similar technological characteristics as the predicate device.

Performance Data:

Performance data (bench and animal testing) demonstrate that the subject device is substantially equivalent to the predicate device and that the design output meets the design input requirements. The testing included dimensional measurements, mechanical and functional verification, simulated use in animal models, and human factors assessment.

Bench Testing:

The *da Vinci Xi* Hasson Cone, was subjected to a series of bench tests to evaluate performance and to demonstrate that the design outputs meet the input requirements. The design verification testing included:

- Physical Specifications
- Mechanical Requirements
- User Interface Requirements
- Functional Requirements

Animal Validation Testing:

Design Validation Testing was performed using animal and cadaveric models to evaluate the performance of the *da Vinci Xi* Hasson Cone in a simulated clinical setting. The testing summarized in this submission validated general, functional, use and misuse-related, and interaction (compatibility) requirements for the subject device.

Human Factors Evaluation:

As part of the Usability Engineering Process for the *da Vinci Xi* Hasson Cone, a Usability Risk Analysis was created to identify any new usability characteristics related to safety, as well as foreseeable hazards and hazardous situations. All risks were assessed as having “Tolerable” or “Broadly Acceptable” pre-mitigated risk profiles. As a result, no further usability testing was conducted to evaluate the safety and usability of the subject device.

Summary:

Based on the intended use, indications for use, technological characteristics, and performance data, the *da Vinci Xi* Hasson Cone is substantially equivalent to its predicate device.